


Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA_M\DATA\BM013018\
 Data File : BM013999.D
 Acq On : 30 Jan 2018 16:11
 Operator : SJ/JU
 Sample : DFTPP87
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP87

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM011018.M
 Title : SVOA CALIBRATION
 Last Update : Wed Jan 31 07:06:39 2018



AutoFind: Scans 2453, 2454, 2455; Background Corrected with Scan 2448

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	46.0	62440	PASS
68	69	0.00	2	2.0	1198	PASS
69	198	0.00	100	43.9	59605	PASS
70	69	0.00	2	0.4	221	PASS
127	198	10	80	44.3	60213	PASS
197	198	0.00	2	0.6	779	PASS
198	198	100	100	100.0	135792	PASS
199	198	5	9	6.1	8304	PASS
275	198	10	60	29.2	39664	PASS
365	198	1	100	4.8	6498	PASS
441	443	0.01	100	72.0	15072	PASS
442	198	50	100	75.9	103133	PASS
443	442	15	24	20.3	20920	PASS

m/z	Abundance
38.00	1437.0
39.00	4070.0
40.00	458.0
44.10	546.0
49.10	387.0
50.00	17056.0
51.10	56400.0
52.00	3025.0
53.20	498.0
55.00	809.0
56.00	1422.0
57.00	2091.0
60.90	1256.0
62.20	467.0
62.90	2389.0
64.10	316.0
65.10	535.0
68.10	1396.0
69.00	50200.0
72.80	789.0
73.10	816.0
74.00	6025.0
75.00	9895.0
76.10	3128.0
77.00	54736.0
78.00	4291.0
79.00	4580.0
79.90	3589.0
80.90	4859.0
82.00	827.0
83.00	391.0
85.10	863.0
85.90	1549.0
87.20	557.0
90.90	891.0
91.90	726.0
93.00	9813.0
93.80	710.0
98.00	6793.0
99.00	5096.0
100.10	606.0
100.90	1721.0
103.00	771.0
103.90	1825.0
105.10	1604.0
106.00	464.0
107.00	18088.0
108.00	2075.0
109.00	636.0
110.00	25464.0
111.00	5581.0
111.90	714.0
113.10	356.0
116.00	1148.0
117.00	26376.0
118.00	1914.0
122.10	1959.0
123.00	1975.0
124.10	907.0
125.00	765.0
127.00	46928.0
128.10	3136.0
129.00	22448.0
129.90	1967.0
131.10	344.0
134.00	311.0
135.10	2072.0
136.00	1691.0
136.90	1560.0
137.80	332.0
140.00	510.0
140.90	2954.0
142.00	845.0
142.70	331.0
143.20	364.0
146.00	911.0
147.10	1846.0
148.10	5087.0
148.90	1164.0
150.00	300.0
151.10	535.0
151.90	458.0
152.90	859.0
154.00	504.0
155.00	2250.0
156.10	2427.0
158.00	828.0
158.90	413.0
160.00	1787.0
160.80	1590.0
161.90	780.0

Instrument :
BNA_M
ClientSampleId :
DFTPP87

165.10	1493.0
165.90	916.0
167.00	8215.0
168.00	4813.0
168.90	627.0
171.80	617.0
173.00	1136.0
173.90	1462.0
175.10	2629.0
176.20	488.0
176.90	1544.0
178.90	5442.0
180.00	3069.0
181.00	913.0
181.70	396.0
182.40	312.0
185.00	2176.0
186.00	16552.0
187.00	4708.0
188.00	596.0
188.90	1285.0
190.00	571.0
190.80	527.0
191.30	587.0
192.00	1881.0
192.90	2081.0
195.90	3301.0
196.80	607.0
198.00	99728.0
198.90	5713.0
200.10	1058.0
201.10	661.0
202.90	879.0
203.90	5125.0
205.00	7973.0
206.00	28952.0
207.00	5334.0
208.00	1665.0
209.00	715.0
209.90	764.0
211.00	1435.0
214.90	904.0
215.80	538.0
217.00	10614.0
218.10	1164.0
220.90	3869.0
222.10	1083.0
223.00	2237.0
224.00	16116.0
225.00	4841.0
225.80	334.0
226.90	9204.0
227.90	1608.0
229.10	1703.0
231.10	631.0
232.90	470.0
234.00	634.0
236.10	772.0
237.00	1061.0
239.90	525.0
240.90	614.0
241.90	751.0
243.20	593.0
244.10	13387.0
244.90	1184.0
245.90	2699.0
246.90	767.0
249.00	474.0
253.00	437.0
254.20	614.0
255.00	64664.0
256.00	10937.0
257.00	699.0
258.00	4017.0
259.10	1092.0
265.00	1964.0
266.90	397.0
271.10	530.0
272.90	1448.0
274.00	5799.0
275.00	23568.0
276.00	3712.0
276.90	3027.0
277.90	876.0
279.00	323.0
281.10	429.0
282.70	393.0
285.00	316.0
292.00	470.0
293.10	779.0
296.00	11572.0
297.10	1423.0
303.10	514.0
314.00	478.0
315.00	568.0

Instrument :
BNA_M
ClientSampleId :
DFTPP87

315.80	516.0
321.00	488.0
322.00	354.0
323.10	1757.0
323.90	598.0
327.00	406.0
328.10	466.0
332.00	347.0
332.80	334.0
334.10	1807.0
334.90	464.0
345.80	462.0
352.20	473.0
354.10	811.0
355.10	456.0
365.00	3970.0
366.00	643.0
372.00	1111.0
402.00	721.0
402.90	671.0
421.10	357.0
422.00	300.0
422.90	2859.0
424.00	734.0
441.10	8198.0
442.00	54464.0
443.00	12626.0
443.80	1198.0

Instrument :
BNA_M
ClientSampleId :
DFTPP87

m/z	Abundance
37.00	504.0
38.00	1052.0
39.00	5626.0
40.10	623.0
41.10	317.0
44.00	698.0
48.80	426.0
50.00	21288.0
51.00	75592.0
51.90	2447.0
53.20	331.0
54.90	382.0
56.10	2071.0
57.00	5243.0
60.90	1350.0
62.00	1147.0
63.00	2883.0
63.80	317.0
65.10	1126.0
67.90	1018.0
69.00	70008.0
70.00	665.0
73.10	646.0
74.00	9282.0
75.00	12396.0
76.10	4785.0
77.00	76288.0
78.10	5076.0
79.00	7595.0
80.00	6457.0
81.00	5946.0
81.90	2061.0
83.00	970.0
85.10	904.0
86.00	2976.0
87.00	1010.0
88.10	681.0
90.90	1774.0
91.90	1325.0
93.00	14136.0
94.00	868.0
95.90	478.0
98.00	12991.0
99.00	7189.0
99.80	794.0
101.00	3461.0
103.00	1165.0
104.00	2791.0
104.90	2067.0
106.00	1097.0
107.00	27688.0
108.00	3640.0
110.00	40240.0
111.00	5847.0
111.90	1760.0
113.10	371.0
116.10	1301.0
117.00	36488.0
118.10	2733.0
119.00	362.0
122.00	2583.0
122.90	4218.0
124.10	1390.0
125.20	1023.0
126.10	361.0
127.00	74376.0
128.10	5771.0
129.00	38296.0
130.00	3275.0
130.80	758.0
131.80	345.0
132.80	619.0
134.10	1271.0
134.90	2224.0
135.90	1344.0
137.10	1808.0
137.80	369.0
139.70	344.0
140.00	371.0
141.00	4783.0
142.10	983.0
143.00	1042.0
146.00	1212.0
147.00	2826.0
148.00	6875.0
149.00	1223.0
150.10	488.0
151.00	1158.0
153.00	2065.0
154.00	1776.0
154.90	2984.0

Instrument :
BNA_M
ClientSampleId :
DFTPP87

156.00	4001.0
156.80	604.0
157.90	1693.0
158.90	837.0
159.90	1651.0
161.00	2931.0
162.00	570.0
162.90	334.0
165.00	2020.0
166.10	1531.0
167.00	12511.0
168.00	6117.0
169.00	839.0
171.80	906.0
173.00	1202.0
173.90	2948.0
175.10	2724.0
176.10	1345.0
177.10	1446.0
177.80	787.0
179.00	7988.0
180.00	4836.0
181.00	1828.0
182.80	410.0
183.70	542.0
185.00	3749.0
186.00	26144.0
187.10	8626.0
187.90	811.0
188.90	2736.0
189.90	533.0
190.90	1347.0
192.00	3195.0
193.00	2457.0
194.10	552.0
194.90	564.0
195.90	3146.0
196.70	924.0
198.00	160320.0
198.90	10675.0
199.90	930.0
202.20	593.0
203.00	1321.0
204.00	7268.0
205.00	14809.0
206.10	44992.0
207.00	8428.0
207.90	1776.0
208.90	752.0
209.90	1159.0
211.00	2144.0
214.90	1167.0
215.90	951.0
217.00	13790.0
218.00	1307.0
221.00	5378.0
222.00	2291.0
223.10	2547.0
224.00	28520.0
225.00	5988.0
225.80	628.0
227.00	16920.0
227.90	1889.0
228.90	2887.0
229.80	328.0
231.10	1250.0
233.90	891.0
234.90	958.0
236.00	654.0
237.00	1033.0
239.00	1384.0
240.00	708.0
240.90	473.0
241.90	1839.0
243.00	2090.0
244.00	22888.0
245.10	2674.0
246.10	4748.0
247.00	1117.0
247.90	414.0
249.10	752.0
251.80	340.0
253.00	403.0
254.10	1166.0
255.00	113408.0
256.00	15310.0
257.00	1621.0
258.00	9845.0
258.90	1864.0
259.90	424.0
265.00	4363.0
265.90	606.0
270.90	451.0
271.80	452.0
272.90	3168.0

Instrument :
BNA_M
ClientSampleId :
DFTPP87

274.00	10599.0
275.00	48352.0
275.90	6292.0
276.90	5824.0
277.80	800.0
279.00	440.0
283.00	736.0
284.00	589.0
285.00	829.0
293.10	643.0
294.00	407.0
295.00	458.0
296.00	18184.0
297.00	3146.0
303.10	1983.0
304.00	889.0
307.90	421.0
314.00	700.0
315.10	2031.0
316.10	1226.0
321.80	391.0
323.10	3241.0
324.10	862.0
327.00	1118.0
328.10	605.0
331.80	320.0
332.90	308.0
334.00	3660.0
335.00	868.0
341.20	953.0
345.90	1066.0
352.10	1024.0
352.90	830.0
354.30	1161.0
355.00	617.0
365.00	7448.0
366.00	811.0
372.00	1888.0
373.20	442.0
382.80	406.0
383.10	434.0
390.00	456.0
401.90	801.0
403.10	1235.0
405.00	331.0
420.60	535.0
421.10	625.0
422.00	853.0
423.10	5240.0
424.10	1660.0
441.00	14890.0
442.10	114392.0
443.10	22784.0
444.00	2306.0

Instrument :
BNA_M
ClientSampleId :
DFTPP87

m/z	Abundance
37.00	686.0
37.90	846.0
39.00	5027.0
40.10	382.0
40.90	548.0
43.90	1239.0
44.90	502.0
49.10	772.0
50.10	17928.0
51.00	55328.0
52.10	2473.0
52.90	330.0
55.10	988.0
56.00	2202.0
56.90	2734.0
61.00	449.0
62.00	730.0
63.00	1358.0
63.90	862.0
65.20	1047.0
68.10	1181.0
69.00	58608.0
73.00	1036.0
74.00	7127.0
75.00	11470.0
76.00	3800.0
77.00	62384.0
78.10	3639.0
79.00	4673.0
80.00	3864.0
81.10	5154.0
81.90	1216.0
82.90	1576.0
85.10	700.0
86.10	2416.0
86.80	843.0
87.90	491.0
91.10	1420.0
92.00	1585.0
93.00	11595.0
93.90	535.0
95.90	491.0
97.00	619.0
98.00	9354.0
99.00	5731.0
101.00	2044.0
102.90	1191.0
104.00	2029.0
104.90	1955.0
105.90	652.0
107.00	23560.0
108.00	2799.0
110.00	28912.0
111.00	5686.0
111.90	1103.0
112.90	462.0
116.10	1237.0
117.00	32600.0
118.00	2095.0
122.00	2711.0
123.00	2448.0
124.00	1496.0
125.00	1047.0
126.20	772.0
127.00	59336.0
128.00	5242.0
129.00	30088.0
129.90	2535.0
130.80	334.0
133.10	418.0
134.10	799.0
135.00	3051.0
135.90	624.0
137.00	1218.0
138.90	397.0
139.90	794.0
141.00	4052.0
142.10	1561.0
143.10	754.0
146.00	1289.0
146.90	2182.0
148.00	7284.0
149.10	862.0
150.00	491.0
150.80	589.0
153.10	1277.0
154.00	1712.0
155.00	2945.0
156.00	3241.0
157.00	695.0
157.80	1156.0

Instrument :
BNA_M
ClientSampleId :
DFTPP87

159.00	672.0
159.90	1743.0
161.00	1678.0
161.80	600.0
163.90	432.0
165.00	1548.0
166.00	1739.0
166.90	10755.0
168.00	5366.0
168.90	933.0
169.80	389.0
171.10	465.0
171.90	381.0
173.10	1171.0
174.00	1775.0
175.00	3296.0
176.00	811.0
177.00	2135.0
178.00	523.0
179.00	6636.0
179.90	5351.0
181.00	3050.0
182.00	349.0
183.80	895.0
185.00	3569.0
186.10	23424.0
187.00	6571.0
188.20	300.0
188.90	1634.0
190.00	550.0
191.90	2366.0
193.00	2352.0
194.10	801.0
196.00	4773.0
196.90	807.0
197.90	147328.0
198.90	8524.0
200.00	945.0
201.20	472.0
202.90	1972.0
204.00	7165.0
205.00	10461.0
206.00	41448.0
207.00	6003.0
207.90	1764.0
209.00	873.0
210.00	1221.0
210.90	2385.0
211.80	602.0
213.20	326.0
215.00	970.0
215.90	1475.0
217.00	13598.0
217.90	1843.0
220.20	369.0
220.90	6427.0
221.70	1648.0
223.10	2759.0
224.00	25880.0
225.00	6387.0
225.90	1183.0
226.90	13843.0
228.00	2925.0
228.90	2097.0
230.90	922.0
231.90	553.0
234.10	982.0
235.10	1399.0
236.00	727.0
236.90	1386.0
238.90	958.0
239.90	457.0
240.90	509.0
241.90	1610.0
242.90	1025.0
244.00	21400.0
244.90	2469.0
246.00	4970.0
247.00	764.0
247.90	411.0
249.10	426.0
250.80	337.0
253.00	601.0
254.20	653.0
255.00	109096.0
256.00	17016.0
256.90	904.0
258.00	8397.0
258.90	1401.0
265.00	3094.0
266.10	590.0
271.80	714.0
273.10	3805.0
274.00	9468.0
275.00	47072.0

Instrument :
BNA_M
ClientSampleId :
DFTPP87

276.00	6354.0
277.00	4904.0
278.00	1016.0
281.00	911.0
283.00	661.0
284.10	453.0
285.10	781.0
293.00	1950.0
295.20	519.0
296.00	18256.0
297.10	2534.0
302.30	300.0
303.10	1414.0
303.90	474.0
307.90	429.0
312.70	356.0
314.00	691.0
314.80	2680.0
315.90	751.0
321.10	688.0
323.00	5178.0
324.00	1294.0
327.00	1758.0
328.00	857.0
332.90	1101.0
334.10	3449.0
335.00	604.0
341.20	500.0
342.30	359.0
345.90	1048.0
351.80	884.0
353.00	1179.0
354.10	1597.0
365.00	8078.0
366.10	695.0
371.10	522.0
372.20	2336.0
373.00	945.0
383.10	517.0
401.80	949.0
402.90	1839.0
403.90	387.0
421.10	1618.0
421.80	567.0
423.00	6425.0
423.90	1636.0
441.10	22128.0
442.10	140544.0
443.10	27352.0
444.00	2479.0

Instrument :
BNA_M
ClientSampleId :
DFTPP87

Scan 2448 (17.386 min): BM013999.D
DFTPP87

m/z	Abundance
44.00	854.0
73.10	535.0
147.10	631.0
206.90	713.0
281.00	416.0

Instrument :
BNA_M
ClientSampleId :
DFTPP87