

Data Path : Z:\HPCHEM1\BNA M\DATA\BM013118\
 Data File : BM014035.D
 Acq On : 31 Jan 2018 15:41
 Operator : SJ/JU
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02020

Quant Time: Feb 01 00:56:06 2018
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM013118.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 31 16:01:40 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
2	1,4-Dioxane	0.479	0.429	10.4	84	0.00
3 S	1,4-Dioxane-d8	0.436	0.404	7.3	89	0.00
4	Benzaldehyde	0.721	0.697	3.3	88	0.00
5 S	Phenol-d5	1.503	1.350	10.2	88	0.00
6	Phenol	1.504	1.342	10.8	86	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.907	0.829	8.6	85	0.00
8	Bis(2-Chloroethyl)ether	1.169	1.054	9.8	85	0.00
9 S	2-Chlorophenol-d4	1.190	1.195	-0.4	93	0.00
10	2-Chlorophenol	1.201	1.123	6.5	86	0.00
11	2-Methylphenol	1.158	1.090	5.9	87	0.00
12	2,2'-oxybis(1-Chloropropane	1.174	1.067	9.1	83	0.00
13 S	4-Methylphenol-d8	1.259	1.131	10.2	87	0.00
14	Acetophenone	2.125	2.015	5.2	93	0.00
15 P	N-Nitroso-di-n-propylamine	1.087	1.070	1.6	90	0.00
16	4-Methylphenol	1.237	1.158	6.4	87	0.00
17	Hexachloroethane	0.627	0.627	0.0	93	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	89	0.00
19 S	Nitrobenzene-d5	0.144	0.154	-6.9	99	0.00
20	Nitrobenzene	0.406	0.414	-2.0	94	0.00
21	Isophorone	0.683	0.671	1.8	87	0.00
22 S	2-Nitrophenol-d4	0.160	0.167	-4.4	93	0.00
23 C	2-Nitrophenol	0.164	0.167	-1.8	90	0.00
24	2,4-Dimethylphenol	0.370	0.381	-3.0	90	0.00
25	Bis(2-Chloroethoxy)methane	0.376	0.381	-1.3	91	0.00
26 S	2,4-Dichlorophenol-d3	0.336	0.334	0.6	86	0.00
27 C	2,4-Dichlorophenol	0.312	0.302	3.2	82	0.00
28	Naphthalene	1.007	0.994	1.3	89	0.00
29 S	4-Chloroaniline-d4	0.291	0.265	8.9	93	0.00
30	4-Chloroaniline	0.292	0.270	7.5	92	0.00
31 C	Hexachlorobutadiene	0.282	0.283	-0.4	93	0.00
32	Caprolactam	0.086	0.084	2.3	94	0.01
33 C	4-Chloro-3-methylphenol	0.327	0.329	-0.6	90	0.00
34	2-Methylnaphthalene	0.760	0.752	1.1	89	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
36	1,2,4,5-Tetrachlorobenzene	0.730	0.722	1.1	90	0.00
37	Hexachlorocyclopentadiene	0.398	0.336	15.6	92	0.00
38 C	2,4,6-Trichlorophenol	0.425	0.419	1.4	87	0.00
39	2,4,5-Trichlorophenol	0.437	0.411	5.9	89	0.00
40	1,1'-Biphenyl	1.554	1.514	2.6	89	0.00
41	2-Chloronaphthalene	1.255	1.225	2.4	88	0.00
42	2-Nitroaniline	0.357	0.335	6.2	86	0.00
43 S	Dimethylphthalate-d6	1.660	1.622	2.3	90	0.00
44	Dimethylphthalate	1.600	1.592	0.5	93	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM013118\
 Data File : BM014035.D
 Acq On : 31 Jan 2018 15:41
 Operator : SJ/JU
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02020

Quant Time: Feb 01 00:56:06 2018
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM013118.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 31 16:01:40 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.311	0.299	3.9	90	0.00
46 S	Acenaphthylene-d8	2.014	1.991	1.1	91	0.00
47	Acenaphthylene	1.870	1.844	1.4	90	0.00
48	3-Nitroaniline	0.234	0.204	12.8	91	0.00
49 C	Acenaphthene	1.311	1.293	1.4	91	0.00
50	2,4-Dinitrophenol	0.167	0.133	20.4#	95	0.00
51 S	4-Nitrophenol-d4	0.229	0.191	16.6	84	0.00
52	4-Nitrophenol	0.276	0.252	8.7	90	0.00
53	Dibenzofuran	1.996	1.968	1.4	90	0.00
54	2,4-Dinitrotoluene	0.470	0.475	-1.1	93	0.00
55	2,3,4,6-Tetrachlorophenol	0.440	0.432	1.8	89	0.00
56	Diethylphthalate	1.679	1.686	-0.4	93	0.00
57 S	Fluorene-d10	1.486	1.465	1.4	91	0.00
58	Fluorene	1.665	1.637	1.7	92	0.00
59	4-Chlorophenyl-phenylether	0.953	0.951	0.2	93	0.00
60	4-Nitroaniline	0.262	0.217	17.2	84	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.123	0.109	11.4	88	0.00
63	4,6-Dinitro-2-methylphenol	0.124	0.111	10.5	91	0.00
64	N-Nitrosodiphenylamine	0.577	0.547	5.2	89	0.00
65	4-Bromophenyl-phenylether	0.250	0.239	4.4	91	0.00
66	Hexachlorobenzene	0.266	0.260	2.3	93	0.00
67	Atrazine	0.232	0.223	3.9	94	0.00
68 C	Pentachlorophenol	0.131	0.111	15.3	88	0.00
69	Phenanthrene	1.125	1.087	3.4	91	0.00
70 S	Anthracene-d10	0.971	0.957	1.4	94	0.00
71	Anthracene	1.139	1.107	2.8	93	0.00
72	Carbazole	0.929	0.876	5.7	92	0.00
73	Di-n-butylphthalate	1.166	1.138	2.4	92	0.00
74 C	Fluoranthene	1.399	1.323	5.4	93	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
76 S	Pyrene-d10	0.963	0.938	2.6	96	0.00
77	Pyrene	1.193	1.178	1.3	96	0.00
78	Butylbenzylphthalate	0.470	0.452	3.8	94	0.00
79	3,3'-Dichlorobenzidine	0.312	0.301	3.5	101	0.00
80	Benzo(a)anthracene	1.246	1.209	3.0	95	0.00
81	Bis(2-ethylhexyl)phthalate	0.686	0.643	6.3	91	0.00
82	Chrysene	1.171	1.130	3.5	96	0.00
83 I	Perylene-d12	1.000	1.000	0.0	99	0.00
84	Di-n-octyl phthalate	1.273	1.117	12.3	93	0.00
85	Benzo(b)fluoranthene	1.262	1.237	2.0	98	0.00
86	Benzo(k)fluoranthene	1.241	1.201	3.2	97	0.00
87 S	Benzo(a)pyrene-d12	0.982	0.953	3.0	99	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM013118\
Data File : BM014035.D
Acq On : 31 Jan 2018 15:41
Operator : SJ/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02020

Quant Time: Feb 01 00:56:06 2018
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM013118.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jan 31 16:01:40 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.189	1.134	4.6	95	0.00
89	Indeno(1,2,3-cd)pyrene	1.303	1.279	1.8	97	0.01
90	Dibenzo(a,h)anthracene	1.105	1.083	2.0	99	0.01
91	Benzo(a,h,i)perylene	1.078	1.059	1.8	100	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0