

Data Path : Z:\HPCHEM1\BNA M\DATA\BM013118\
 Data File : BM014043.D
 Acq On : 31 Jan 2018 20:30
 Operator : SJ/JU
 Sample : J1314-18
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 F6C55

Manual Integrations
 APPROVED

Sohil
 2/1/2018 6:04:43 PM

Quant Time: Feb 01 01:45:14 2018
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM013118.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 01 01:00:18 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.56	152	95295	20.00	ng/ul	0.00
18) Naphthalene-d8	10.33	136	392661	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.22	164	269033	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.97	188	652560	20.00	ng/ul	0.00
75) Chrysene-d12	21.18	240	689193	20.00	ng/ul	0.00
83) Perylene-d12	23.36	264	654861	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.12	96	10820	5.21	ng/uL	0.00
5) Phenol-d5	6.75	99	202454	28.27	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.91	67	136522	31.60	ng/ul	0.00
9) 2-Chlorophenol-d4	7.10	132	166956	29.44	ng/ul	0.00
13) 4-Methylphenol-d8	8.27	113	188748	31.47	ng/ul	0.00
19) Nitrobenzene-d5	8.72	128	97210	34.30	ng/ul	0.00
22) 2-Nitrophenol-d4	9.43	143	105285	33.56	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.97	165	198945	30.15	ng/ul	0.00
29) 4-Chloroaniline-d4	10.49	131	232559	40.65	ng/ul	0.00
43) Dimethylphthalate-d6	13.64	166	763834	34.21	ng/ul	0.00
46) Acenaphthylene-d8	13.91	160	914797	33.77	ng/ul	0.00
51) 4-Nitrophenol-d4	14.46	143	76767m	24.95	ng/ul	0.00
57) Fluorene-d10	15.22	176	669127	33.49	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.36	200	98137	24.49	ng/ul	0.00
70) Anthracene-d10	17.07	188	1104661	34.87	ng/ul	0.00
76) Pyrene-d10	19.37	212	1218694	36.72	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.22	264	1088167	33.86	ng/ul	0.00

Target Compounds				Ovalue
44) Dimethylphthalate	13.69	163	135862	6.312 ng/ul# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA M\DATA\BM013118\
Data File : BM014043.D
Acq On : 31 Jan 2018 20:30
Operator : SJ/JU
Sample : J1314-18
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
F6C55

Manual Integrations
APPROVED

Sohil
2/1/2018 6:04:43 PM

Quant Time: Feb 01 01:45:14 2018
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM013118.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 01 01:00:18 2018
Response via : Initial Calibration





#44

Dimethylphthalate

Concen: 6.312 ng/ul

RT: 13.69 min Scan# 1819

Delta R.T. 0.00 min

Lab File: BM014043.D

Acq: 31 Jan 2018 20:30

Instrument :

BNA_M

ClientSampleId :

F6C55

Tot Ion	Ratio	Resp Lower	Upper
163	100		
194	5.7	3.5	5.3#
164	11.3	8.2	12.4

Manual Integrations
APPROVED

Sohil
2/1/2018 6:04:43 PM

