

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM013122\
 Data File : BM033924.D
 Acq On : 31 Jan 2022 13:46
 Operator : CG/JU
 Sample : SSTD00525
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD005025

Quant Time: Jan 31 16:45:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM013122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 31 16:44:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.931	152	113682	20.000	ng/ul	0.00
20) Naphthalene-d8	10.742	136	472610	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.566	164	312359	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.301	188	656095	20.000	ng/ul	0.00
79) Chrysene-d12	21.465	240	597713	20.000	ng/ul	0.00
88) Perylene-d12	23.859	264	534446	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.290	96	5613	1.855	ng/uL	0.00
4) Pyridine-d5	3.725	84	33348	3.969	ng/ul	0.00
7) Phenol-d5	7.084	99	42482	4.161	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.248	67	28447	4.575	ng/ul	0.00
11) 2-Chlorophenol-d4	7.460	132	34189	4.380	ng/ul	0.00
15) 4-Methylphenol-d8	8.637	113	33560	3.971	ng/ul	0.00
21) Nitrobenzene-d5	9.089	128	16392	4.392	ng/ul	0.00
24) 2-Nitrophenol-d4	9.819	143	17116	4.327	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.360	165	32370	4.269	ng/ul	0.00
31) 4-Chloroaniline-d4	10.878	131	46095	4.089	ng/ul	0.00
46) Dimethylphthalate-d6	13.977	166	111800	4.543	ng/ul	0.00
49) Acenaphthylene-d8	14.260	160	126013	4.340	ng/ul	0.00
54) 4-Nitrophenol-d4	14.754	143	16251	3.441	ng/ul	0.00
60) Fluorene-d10	15.554	176	93116	4.561	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.666	200	13882	3.557	ng/ul	0.00
73) Anthracene-d10	17.401	188	144645	4.543	ng/ul	0.00
81) Pyrene-d10	19.683	212	164032	5.082	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.700	264	126754	4.558	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.325	88	6044	1.914	ng/uL#	85
5) Pyridine	3.743	79	36917	4.284	ng/ul	90
6) Benzaldehyde	7.060	77	29271	4.968	ng/ul#	82
8) Phenol	7.113	94	43989	4.194	ng/ul#	85
10) Bis(2-Chloroethyl)ether	7.343	93	36252	4.437	ng/ul	88
12) 2-Chlorophenol	7.490	128	36612	4.465	ng/ul#	86
13) 2-Methylphenol	8.372	108	32961	4.087	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.460	45	52101	4.533	ng/ul#	94
16) Acetophenone	8.754	105	57486	4.323	ng/ul#	88
17) N-Nitroso-di-n-propyla...	8.737	70	28919	4.199	ng/ul	91
18) 4-Methylphenol	8.701	108	35331	4.008	ng/ul	92
19) Hexachloroethane	9.019	117	15585	4.482	ng/ul#	90
22) Nitrobenzene	9.131	77	42480	4.489	ng/ul	92
23) Isophorone	9.666	82	77939	4.309	ng/ul	95
25) 2-Nitrophenol	9.848	139	18692	4.380	ng/ul#	85
26) 2,4-Dimethylphenol	9.913	107	44353	4.493	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.148	93	49734	4.587	ng/ul	98
29) 2,4-Dichlorophenol	10.383	162	33136	4.343	ng/ul	98
30) Naphthalene	10.789	128	123128	4.704	ng/ul	99
32) 4-Chloroaniline	10.895	127	47622	4.180	ng/ul	98
33) Hexachlorobutadiene	11.083	225	25225	4.797	ng/ul	97
34) Caprolactam	11.654	113	9342	3.580	ng/ul#	73
35) 4-Chloro-3-methylphenol	12.025	107	39815	4.275	ng/ul	91
36) 2-Methylnaphthalene	12.401	142	83170	4.613	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM013122\
 Data File : BM033924.D
 Acq On : 31 Jan 2022 13:46
 Operator : CG/JU
 Sample : SST00525
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SST005025


Quant Time: Jan 31 16:45:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM013122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 31 16:44:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.619	142	86006	4.631	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.772	216	44153	4.730	ng/ul	98
40) Hexachlorocyclopentadiene	12.754	237	26214	4.277	ng/ul	98
41) 2,4,6-Trichlorophenol	13.007	196	26271	4.380	ng/ul	98
42) 2,4,5-Trichlorophenol	13.077	196	27946	4.211	ng/ul	91
43) 1,1'-Biphenyl	13.407	154	113795	4.646	ng/ul	98
44) 2-Chloronaphthalene	13.448	162	84221	4.576	ng/ul	98
45) 2-Nitroaniline	13.648	65	22014	3.676	ng/ul#	83
47) Dimethylphthalate	14.019	163	113729	4.622	ng/ul	99
48) 2,6-Dinitrotoluene	14.136	165	18471	3.917	ng/ul	96
50) Acenaphthylene	14.289	152	136598	4.430	ng/ul	98
51) 3-Nitroaniline	14.466	138	17811	3.675	ng/ul	86
52) Acenaphthene	14.630	153	92813	4.573	ng/ul	100
53) 2,4-Dinitrophenol	14.671	184	8828	3.110	ng/ul	89
55) 4-Nitrophenol	14.771	109	14830	3.241	ng/ul	86
56) Dibenzofuran	14.966	168	132434	4.549	ng/ul	98
57) 2,4-Dinitrotoluene	14.918	165	27233	3.857	ng/ul	87
58) 2,3,4,6-Tetrachlorophenol	15.189	232	24755	4.401	ng/ul	97
59) Diethylphthalate	15.377	149	117346	4.514	ng/ul	99
61) Fluorene	15.607	166	107847	4.561	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.601	204	55658	4.690	ng/ul	97
63) 4-Nitroaniline	15.618	138	17599	3.769	ng/ul#	85
66) 4,6-Dinitro-2-methylph...	15.683	198	14120	3.600	ng/ul#	95
67) N-Nitrosodiphenylamine	15.813	169	89512	4.494	ng/ul	99
68) 4-Bromophenyl-phenylether	16.495	248	31835	4.883	ng/ul	96
69) Hexachlorobenzene	16.613	284	37292	4.697	ng/ul	98
70) Atrazine	16.760	200	34263	4.322	ng/ul	99
71) Pentachlorophenol	16.954	266	19784	3.809	ng/ul	98
72) Phenanthrene	17.348	178	170998	4.570	ng/ul	98
74) Anthracene	17.436	178	168406	4.431	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.372	216	48238	4.887	ng/ul	97
76) Pentachlorobenzene	14.889	250	44703	4.822	ng/ul	97
77) Carbazole	17.701	167	149847	4.395	ng/ul	98
78) Di-n-butylphthalate	18.271	149	191862	4.469	ng/ul	99
80) Fluoranthene	19.354	202	197852	5.157	ng/ul	100
82) Pyrene	19.712	202	203309	5.125	ng/ul	99
83) Butylbenzylphthalate	20.606	149	79464	4.631	ng/ul	88
84) 3,3'-Dichlorobenzidine	21.383	252	56103	4.219	ng/ul	97
85) Benzo(a)anthracene	21.453	228	181898	4.648	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.383	149	121556	4.517	ng/ul	100
87) Chrysene	21.506	228	176471	4.645	ng/ul	97
89) Di-n-octyl phthalate	22.300	149	186594	4.257	ng/ul	100
90) Benzo(b)fluoranthene	23.130	252	170158	4.688	ng/ul	98
91) Benzo(k)fluoranthene	23.177	252	155314	4.597	ng/ul	99
93) Benzo(a)pyrene	23.753	252	155308	4.575	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.353	276	166628	4.877	ng/ul	97
95) Dibenzo(a,h)anthracene	26.371	278	141503	4.807	ng/ul#	95
96) Benzo(g,h,i)perylene	27.118	276	138339	4.962	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 31 16:45:49 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM013122.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jan 31 16:44:50 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 1.914 ng/uL
 RT: 3.325 min Scan# 15
 Delta R.T. -0.000 min
 Lab File: BM033924.D
 Acq: 31 Jan 2022 13:46




Instrument :
 BNA_M
 ClientSampleId :
 SST005025


Tgt Ion: 88 Resp: 6044
 Ion Ratio Lower Upper
 88 100
 43 40.0 36.2 54.4
 58 68.1 68.5 102.7#



#5
 Pyridine
 Concen: 4.284 ng/uL
 RT: 3.743 min Scan# 86
 Delta R.T. 0.006 min
 Lab File: BM033924.D
 Acq: 31 Jan 2022 13:46

Tgt Ion: 79 Resp: 36917
 Ion Ratio Lower Upper
 79 100
 52 83.3 78.1 117.1
 51 44.9 36.0 54.0









#6
Benzaldehyde
Concen: 4.968 ng/ul
RT: 7.060 min Scan# 610
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion: 77 Resp: 29271
Ion Ratio Lower Upper
77 100
105 97.3 65.6 98.4
106 92.2 60.6 90.8#



#8
Phenol
Concen: 4.194 ng/ul
RT: 7.113 min Scan# 659
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion: 94 Resp: 43989
Ion Ratio Lower Upper
94 100
65 31.2 31.7 47.5#
66 44.4 45.4 68.0#









#10
Bis(2-Chloroethyl)ether
Concen: 4.437 ng/ul
RT: 7.343 min Scan# 699
Delta R.T. -0.006 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion: 93 Resp: 36252
Ion Ratio Lower Upper
93 100
63 75.7 71.7 107.5
95 30.4 26.3 39.5



#12
2-Chlorophenol
Concen: 4.465 ng/ul
RT: 7.490 min Scan# 723
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion: 128 Resp: 36612
Ion Ratio Lower Upper
128 100
64 46.8 49.4 74.2#
130 31.7 26.8 40.2









#13
2-Methylphenol
Concen: 4.087 ng/ul
RT: 8.372 min Scan# 873
Delta R.T. -0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:108 Resp: 32961
Ion Ratio Lower Upper
108 100
107 88.0 70.2 105.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 4.533 ng/ul
RT: 8.460 min Scan# 888
Delta R.T. -0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion: 45 Resp: 52101
Ion Ratio Lower Upper
45 100
77 15.3 9.9 14.9#
79 12.4 8.3 12.5









#16
 Acetophenone
 Concen: 4.323 ng/ul
 RT: 8.754 min Scan# 91
 Delta R.T. 0.000 min
 Lab File: BM033924.D
 Acq: 31 Jan 2022 13:46




Instrument :
 BNA_M
 ClientSampleId :
 SST005025


Tgt Ion:105 Resp: 57486
 Ion Ratio Lower Upper
 105 100
 77 80.7 73.1 109.7
 51 29.8 30.5 45.7#



#17
 N-Nitroso-di-n-propylamine
 Concen: 4.199 ng/ul
 RT: 8.737 min Scan# 935
 Delta R.T. -0.006 min
 Lab File: BM033924.D
 Acq: 31 Jan 2022 13:46

Tgt Ion: 70 Resp: 28919
 Ion Ratio Lower Upper
 70 100
 42 62.4 56.9 85.3
 101 9.1 6.9 10.3
 130 19.7 13.8 20.8









#18
4-Methylphenol
Concen: 4.008 ng/ul
RT: 8.701 min Scan# 91
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:108 Resp: 35331
Ion Ratio Lower Upper
108 100
107 117.2 87.4 131.2



#19
Hexachloroethane
Concen: 4.482 ng/ul
RT: 9.019 min Scan# 983
Delta R.T. -0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion:117 Resp: 15585
Ion Ratio Lower Upper
117 100
201 93.4 69.6 104.4
199 61.0 40.6 60.8#









#22
Nitrobenzene
Concen: 4.489 ng/ul
RT: 9.131 min Scan# 1002
Delta R.T. -0.006 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion: 77 Resp: 42480
Ion Ratio Lower Upper
77 100
123 42.7 29.0 43.6
65 13.8 11.4 17.2



#23
Isophorone
Concen: 4.309 ng/ul
RT: 9.666 min Scan# 1093
Delta R.T. -0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion: 82 Resp: 77939
Ion Ratio Lower Upper
82 100
95 9.0 6.5 9.7
138 17.7 11.8 17.8









#25
2-Nitrophenol
Concen: 4.380 ng/ul
RT: 9.848 min Scan# 1124
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:139 Resp: 18692
Ion Ratio Lower Upper
139 100
65 60.4 59.0 88.6
109 36.9 37.2 55.8#



#26
2,4-Dimethylphenol
Concen: 4.493 ng/ul
RT: 9.913 min Scan# 1135
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion:107 Resp: 44353
Ion Ratio Lower Upper
107 100
121 45.0 34.8 52.2
122 76.2 60.0 90.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 4.587 ng/ul
 RT: 10.148 min Scan# 1175
 Delta R.T. 0.000 min
 Lab File: BM033924.D
 Acq: 31 Jan 2022 13:46




Instrument :
 BNA_M
 ClientSampleId :
 SSTD005025


Tgt Ion: 93 Resp: 49734
 Ion Ratio Lower Upper
 93 100
 95 33.0 25.3 37.9
 123 9.7 7.9 11.9



#29
 2,4-Dichlorophenol
 Concen: 4.343 ng/ul
 RT: 10.383 min Scan# 1215
 Delta R.T. -0.006 min
 Lab File: BM033924.D
 Acq: 31 Jan 2022 13:46

Tgt Ion: 162 Resp: 33136
 Ion Ratio Lower Upper
 162 100
 164 64.4 50.0 75.0
 98 38.1 30.6 45.8









#30
Naphthalene
Concen: 4.704 ng/ul
RT: 10.789 min Scan# 1284
Delta R.T. -0.006 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:128 Resp: 123128
Ion Ratio Lower Upper
128 100
129 11.0 9.1 13.7
127 12.2 10.2 15.4



#32
4-Chloroaniline
Concen: 4.180 ng/ul
RT: 10.895 min Scan# 1302
Delta R.T. -0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion:127 Resp: 47622
Ion Ratio Lower Upper
127 100
129 32.8 25.4 38.0









#33
Hexachlorobutadiene
Concen: 4.797 ng/ul
RT: 11.083 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:225 Resp: 25225
Ion Ratio Lower Upper
225 100
223 62.0 52.2 78.2
227 64.1 52.8 79.2



#34
Caprolactam
Concen: 3.580 ng/ul
RT: 11.654 min Scan# 1431
Delta R.T. -0.012 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion:113 Resp: 9342
Ion Ratio Lower Upper
113 100
55 156.3 157.9 236.9#
56 130.2 131.8 197.6#









#35
4-Chloro-3-methylphenol
Concen: 4.275 ng/ul
RT: 12.025 min Scan# 1494
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:107 Resp: 39815
Ion Ratio Lower Upper
107 100
144 23.6 17.1 25.7
142 69.7 49.4 74.2



#36
2-Methylnaphthalene
Concen: 4.613 ng/ul
RT: 12.401 min Scan# 1558
Delta R.T. -0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion:142 Resp: 83170
Ion Ratio Lower Upper
142 100
141 88.3 71.1 106.7





#37

1-Methylnaphthalene

Concen: 4.631 ng/ul

RT: 12.619 min Scan# 1595

Delta R.T. -0.000 min

Lab File: BM033924.D

Acq: 31 Jan 2022 13:46

Instrument :

BNA_M

ClientSampleId :

SSTD005025





Tgt Ion:142 Resp: 86006

Ion Ratio Lower Upper

142 100

141 94.0 75.5 113.3

116 4.4 4.0 6.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 4.730 ng/ul

RT: 12.772 min Scan# 1621

Delta R.T. -0.000 min

Lab File: BM033924.D

Acq: 31 Jan 2022 13:46

Tgt Ion:216 Resp: 44153




Ion Ratio Lower Upper


216 100

214 79.3 61.3 91.9

179 23.1 18.0 27.0

108 20.3 16.6 25.0









#40
Hexachlorocyclopentadiene
Concen: 4.277 ng/ul
RT: 12.754 min Scan# 1618
Delta R.T. -0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:237 Resp: 26214
Ion Ratio Lower Upper
237 100
235 64.9 50.6 76.0
272 11.1 9.6 14.4



#41
2,4,6-Trichlorophenol
Concen: 4.380 ng/ul
RT: 13.007 min Scan# 1661
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion:196 Resp: 26271
Ion Ratio Lower Upper
196 100
198 96.3 78.9 118.3
200 30.0 23.7 35.5





#42

2,4,5-Trichlorophenol

Concen: 4.211 ng/ul

RT: 13.077 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM033924.D

Acq: 31 Jan 2022 13:46

Instrument :

BNA_M

ClientSampleId :

SSTD005025





Tgt Ion:196 Resp: 27946

Ion Ratio Lower Upper

196 100

198 94.2 82.2 123.2

200 28.9 27.9 41.9



#43

1,1'-Biphenyl

Concen: 4.646 ng/ul

RT: 13.407 min Scan# 1729

Delta R.T. -0.000 min

Lab File: BM033924.D

Acq: 31 Jan 2022 13:46




Tgt Ion:154 Resp: 113795


Ion Ratio Lower Upper

154 100

153 40.1 33.1 49.7

76 13.0 10.7 16.1









#44
2-Chloronaphthalene
Concen: 4.576 ng/ul
RT: 13.448 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:162 Resp: 84221
Ion Ratio Lower Upper
162 100
164 32.5 26.2 39.4
127 44.2 33.7 50.5



#45
2-Nitroaniline
Concen: 3.676 ng/ul
RT: 13.648 min Scan# 1770
Delta R.T. 0.006 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion: 65 Resp: 22014
Ion Ratio Lower Upper
65 100
92 66.3 47.8 71.6
138 96.4 61.0 91.6#





#47

Dimethylphthalate

Concen: 4.622 ng/ul

RT: 14.019 min Scan# 1833

Delta R.T. -0.006 min

Lab File: BM033924.D

Acq: 31 Jan 2022 13:46

Instrument :

BNA_M

ClientSampleId :

SSTD005025





Tgt Ion:163 Resp: 113729

Ion Ratio Lower Upper

163 100

194 4.0 3.3 4.9

164 10.8 8.2 12.2



#48

2,6-Dinitrotoluene

Concen: 3.917 ng/ul

RT: 14.136 min Scan# 1853

Delta R.T. -0.006 min

Lab File: BM033924.D

Acq: 31 Jan 2022 13:46




Tgt Ion:165 Resp: 18471


Ion Ratio Lower Upper

165 100

89 65.5 55.7 83.5

121 26.3 20.0 30.0









#50
Acenaphthylene
Concen: 4.430 ng/ul
RT: 14.289 min Scan# 1879
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:152 Resp: 136598
Ion Ratio Lower Upper
152 100
151 20.5 15.8 23.8
153 13.3 10.1 15.1



#51
3-Nitroaniline
Concen: 3.675 ng/ul
RT: 14.466 min Scan# 1909
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion:138 Resp: 17811
Ion Ratio Lower Upper
138 100
108 13.6 10.8 16.2
92 120.1 110.6 166.0









#52
Acenaphthene
Concen: 4.573 ng/ul
RT: 14.630 min Scan# 1937
Delta R.T. -0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:153 Resp: 92813
Ion Ratio Lower Upper
153 100
152 49.2 39.4 59.2
154 90.5 72.7 109.1



#53
2,4-Dinitrophenol
Concen: 3.110 ng/ul
RT: 14.671 min Scan# 1944
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion:184 Resp: 8828
Ion Ratio Lower Upper
184 100
63 75.4 73.8 110.8
154 81.6 63.5 95.3









#55
4-Nitrophenol
Concen: 3.241 ng/ul
RT: 14.771 min Scan# 1961
Delta R.T. 0.006 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:109 Resp: 14830
Ion Ratio Lower Upper
109 100
139 103.2 77.6 116.4
65 116.4 112.2 168.4



#56
Dibenzofuran
Concen: 4.549 ng/ul
RT: 14.966 min Scan# 1994
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion:168 Resp: 132434
Ion Ratio Lower Upper
168 100
139 43.7 34.0 51.0









#57
2,4-Dinitrotoluene
Concen: 3.857 ng/ul
RT: 14.918 min Scan# 1986
Delta R.T. -0.006 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:165 Resp: 27233
Ion Ratio Lower Upper
165 100
63 54.2 51.8 77.8
182 2.8 2.6 4.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 4.401 ng/ul
RT: 15.189 min Scan# 2032
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion:232 Resp: 24755
Ion Ratio Lower Upper
232 100
131 54.0 41.4 62.2
130 2.1 2.0 3.0
166 33.3 27.4 41.0









#59
Diethylphthalate
Concen: 4.514 ng/ul
RT: 15.377 min Scan# 2064
Delta R.T. -0.006 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:149 Resp: 117346
Ion Ratio Lower Upper
149 100
177 20.8 17.1 25.7
150 12.3 9.9 14.9



#61
Fluorene
Concen: 4.561 ng/ul
RT: 15.607 min Scan# 2103
Delta R.T. -0.006 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion:166 Resp: 107847
Ion Ratio Lower Upper
166 100
165 97.1 77.9 116.9
167 14.2 10.3 15.5





#62

4-Chlorophenyl-phenylether

Concen: 4.690 ng/ul

RT: 15.601 min Scan# 2103

Delta R.T. -0.006 min

Lab File: BM033924.D

Acq: 31 Jan 2022 13:46

Instrument :

BNA_M

ClientSampleId :

SSTD005025





Tgt Ion:204 Resp: 55658

Ion Ratio Lower Upper

204 100

206 32.1 27.8 41.8

141 70.8 54.9 82.3



#63

4-Nitroaniline

Concen: 3.769 ng/ul

RT: 15.618 min Scan# 2105

Delta R.T. 0.000 min

Lab File: BM033924.D

Acq: 31 Jan 2022 13:46




Tgt Ion:138 Resp: 17599


Ion Ratio Lower Upper

138 100

92 57.8 48.0 72.0

108 88.2 89.5 134.3#









#66
 4,6-Dinitro-2-methylphenol
 Concen: 3.600 ng/ul
 RT: 15.683 min Scan# 2116
 Delta R.T. 0.000 min
 Lab File: BM033924.D
 Acq: 31 Jan 2022 13:46




Instrument :
 BNA_M
 ClientSampleId :
 SST005025


Tgt Ion:198 Resp: 14120
 Ion Ratio Lower Upper
 198 100
 182 7.3 4.2 6.2#
 77 30.3 26.3 39.5



#67
 N-Nitrosodiphenylamine
 Concen: 4.494 ng/ul
 RT: 15.813 min Scan# 2138
 Delta R.T. 0.000 min
 Lab File: BM033924.D
 Acq: 31 Jan 2022 13:46

Tgt Ion:169 Resp: 89512
 Ion Ratio Lower Upper
 169 100
 168 67.4 52.9 79.3
 167 37.3 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 4.883 ng/ul
RT: 16.495 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:248 Resp: 31835
Ion Ratio Lower Upper
248 100
250 97.3 79.0 118.4
141 94.4 69.7 104.5



#69
Hexachlorobenzene
Concen: 4.697 ng/ul
RT: 16.613 min Scan# 2274
Delta R.T. -0.006 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion:284 Resp: 37292
Ion Ratio Lower Upper
284 100
142 42.9 34.3 51.5
249 35.2 26.1 39.1









#70
Atrazine
Concen: 4.322 ng/ul
RT: 16.760 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:200 Resp: 34263
Ion Ratio Lower Upper
200 100
173 26.3 20.6 30.8
215 45.6 36.9 55.3



#71
Pentachlorophenol
Concen: 3.809 ng/ul
RT: 16.954 min Scan# 2332
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion:266 Resp: 19784
Ion Ratio Lower Upper
266 100
264 62.5 49.1 73.7
268 61.7 47.8 71.8









#72
Phenanthrene
Concen: 4.570 ng/ul
RT: 17.348 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:178 Resp: 170998
Ion Ratio Lower Upper
178 100
179 16.0 11.7 17.5
176 19.3 15.1 22.7



#74
Anthracene
Concen: 4.431 ng/ul
RT: 17.436 min Scan# 2414
Delta R.T. -0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion:178 Resp: 168406
Ion Ratio Lower Upper
178 100
179 16.0 12.3 18.5
176 18.4 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.887 ng/uL
RT: 13.372 min Scan# 1723
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:216 Resp: 48238
Ion Ratio Lower Upper
216 100
214 78.5 65.4 98.0
218 49.3 38.8 58.2



#76
Pentachlorobenzene
Concen: 4.822 ng/uL
RT: 14.889 min Scan# 1981
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion:250 Resp: 44703
Ion Ratio Lower Upper
250 100
248 61.3 51.6 77.4
252 64.1 52.0 78.0









#77
 Carbazole
 Concen: 4.395 ng/ul
 RT: 17.701 min Scan# 2459
 Delta R.T. 0.000 min
 Lab File: BM033924.D
 Acq: 31 Jan 2022 13:46




Instrument :
 BNA_M
 ClientSampleId :
 SSTD005025


Tgt Ion:167 Resp: 149847
 Ion Ratio Lower Upper
 167 100
 166 22.1 16.9 25.3
 139 13.9 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 4.469 ng/ul
 RT: 18.271 min Scan# 2556
 Delta R.T. 0.000 min
 Lab File: BM033924.D
 Acq: 31 Jan 2022 13:46

Tgt Ion:149 Resp: 191862
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.0 10.6
 104 6.0 5.1 7.7









#80
Fluoranthene
Concen: 5.157 ng/u1
RT: 19.354 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:202 Resp: 197852
Ion Ratio Lower Upper
202 100
101 10.1 8.2 12.4
100 8.6 6.9 10.3



#82
Pyrene
Concen: 5.125 ng/u1
RT: 19.712 min Scan# 2801
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion:202 Resp: 203309
Ion Ratio Lower Upper
202 100
101 12.7 9.8 14.6
100 11.0 8.6 13.0









#83
Butylbenzylphthalate
Concen: 4.631 ng/ul
RT: 20.606 min Scan# 2953
Delta R.T. -0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:149 Resp: 79464
Ion Ratio Lower Upper
149 100
91 73.4 69.3 103.9
206 21.6 16.1 24.1



#84
3,3'-Dichlorobenzidine
Concen: 4.219 ng/ul
RT: 21.383 min Scan# 3085
Delta R.T. -0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion:252 Resp: 56103
Ion Ratio Lower Upper
252 100
254 63.8 49.4 74.2
126 11.7 8.2 12.4









#85
Benzo(a)anthracene
Concen: 4.648 ng/ul
RT: 21.453 min Scan# 3097
Delta R.T. -0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:228 Resp: 181898
Ion Ratio Lower Upper
228 100
229 18.4 15.0 22.4
226 26.2 21.4 32.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 4.517 ng/ul
RT: 21.383 min Scan# 3085
Delta R.T. -0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion:149 Resp: 121556
Ion Ratio Lower Upper
149 100
167 27.5 21.8 32.8
279 4.3 3.2 4.8









#87
Chrysene
Concen: 4.645 ng/ul
RT: 21.506 min Scan# 3106
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:228 Resp: 176471
Ion Ratio Lower Upper
228 100
226 28.7 24.5 36.7
229 19.2 15.9 23.9



#89
Di-n-octyl phthalate
Concen: 4.257 ng/ul
RT: 22.300 min Scan# 3241
Delta R.T. 0.000 min
Lab File: BM033924.D
Acq: 31 Jan 2022 13:46

Tgt Ion:149 Resp: 186594





#90

Benzo(b)fluoranthene

Concen: 4.688 ng/ul

RT: 23.130 min Scan# 3382

Delta R.T. -0.000 min

Lab File: BM033924.D

Acq: 31 Jan 2022 13:46

Instrument :

BNA_M

ClientSampleId :

SSTD005025





Tgt Ion:252 Resp: 170158

Ion Ratio Lower Upper

252 100

253 23.0 17.7 26.5

125 10.9 7.8 11.6



#91

Benzo(k)fluoranthene

Concen: 4.597 ng/ul

RT: 23.177 min Scan# 3390

Delta R.T. 0.000 min

Lab File: BM033924.D

Acq: 31 Jan 2022 13:46




Tgt Ion:252 Resp: 155314


Ion Ratio Lower Upper

252 100

253 22.3 17.6 26.4

125 9.7 7.4 11.2





#93

Benzo(a)pyrene

Concen: 4.575 ng/ul

RT: 23.753 min Scan# 34

Delta R.T. 0.000 min

Lab File: BM033924.D

Acq: 31 Jan 2022 13:46

Instrument :

BNA_M

ClientSampleId :

SSTD005025





Tgt Ion:252 Resp: 155308

Ion Ratio Lower Upper

252 100

253 19.9 17.4 26.0

125 12.0 7.8 11.6#



#94

Indeno(1,2,3-cd)pyrene

Concen: 4.877 ng/ul

RT: 26.353 min Scan# 3930

Delta R.T. -0.006 min

Lab File: BM033924.D

Acq: 31 Jan 2022 13:46




Tgt Ion:276 Resp: 166628


Ion Ratio Lower Upper

276 100

138 22.5 16.5 24.7

277 24.6 20.6 31.0









#95
 Dibenzo(a,h)anthracene
 Concen: 4.807 ng/ul
 RT: 26.371 min Scan# 3933
 Delta R.T. -0.000 min
 Lab File: BM033924.D
 Acq: 31 Jan 2022 13:46

Instrument :
 BNA_M
 ClientSampleId :
 SSTD005025

Tgt Ion:278 Resp: 141503
 Ion Ratio Lower Upper
 278 100
 139 16.7 11.0 16.4#
 279 24.5 18.1 27.1



#96
 Benzo(g,h,i)perylene
 Concen: 4.962 ng/ul
 RT: 27.118 min Scan# 4060
 Delta R.T. -0.006 min
 Lab File: BM033924.D
 Acq: 31 Jan 2022 13:46

Tgt Ion:276 Resp: 138339
 Ion Ratio Lower Upper
 276 100
 138 22.7 15.6 23.4
 277 24.5 18.7 28.1

