

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM013122\
 Data File : BM033927.D
 Acq On : 31 Jan 2022 15:35
 Operator : CG/JU
 Sample : SSTD04028
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD040028

Quant Time: Jan 31 16:47:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM013122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 31 16:44:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.931	152	116040	20.000	ng/ul	0.00
20) Naphthalene-d8	10.742	136	458117	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.566	164	270603	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.307	188	517922	20.000	ng/ul	0.00
79) Chrysene-d12	21.471	240	452829	20.000	ng/ul	0.00
88) Perylene-d12	23.865	264	452415	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.290	96	45032	14.580	ng/uL	0.00
4) Pyridine-d5	3.719	84	316094	36.854	ng/ul	0.00
7) Phenol-d5	7.090	99	370202	35.520	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.254	67	228352	35.979	ng/ul	0.00
11) 2-Chlorophenol-d4	7.460	132	306414	38.454	ng/ul	0.00
15) 4-Methylphenol-d8	8.642	113	295627	34.266	ng/ul	0.00
21) Nitrobenzene-d5	9.095	128	147144	40.671	ng/ul	0.00
24) 2-Nitrophenol-d4	9.819	143	163407	42.617	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.360	165	287712	39.141	ng/ul	0.00
31) 4-Chloroaniline-d4	10.878	131	385170	35.250	ng/ul	0.00
46) Dimethylphthalate-d6	13.977	166	791555	37.127	ng/ul	0.00
49) Acenaphthylene-d8	14.266	160	995001	39.559	ng/ul	0.00
54) 4-Nitrophenol-d4	14.760	143	138099	33.750	ng/ul	0.00
60) Fluorene-d10	15.560	176	661103	37.376	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.671	200	127211	41.292	ng/ul	0.00
73) Anthracene-d10	17.401	188	981623	39.059	ng/ul	0.00
81) Pyrene-d10	19.689	212	1047011	42.815	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.712	264	938934	39.887	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.325	88	48745	15.119	ng/uL#	87
5) Pyridine	3.737	79	328034	37.295	ng/ul	85
6) Benzaldehyde	7.060	77	231218	38.446	ng/ul#	82
8) Phenol	7.113	94	383821	35.848	ng/ul#	85
10) Bis(2-Chloroethyl)ether	7.348	93	302250	36.242	ng/ul	90
12) 2-Chlorophenol	7.489	128	316347	37.795	ng/ul	88
13) 2-Methylphenol	8.378	108	289223	35.130	ng/ul	97
14) 2,2'-oxybis(1-Chloropr...	8.472	45	403310	34.375	ng/ul#	95
16) Acetophenone	8.760	105	467285	34.426	ng/ul#	87
17) N-Nitroso-di-n-propyla...	8.748	70	237347	33.766	ng/ul#	89
18) 4-Methylphenol	8.707	108	307399	34.159	ng/ul	94
19) Hexachloroethane	9.019	117	138459	39.008	ng/ul	89
22) Nitrobenzene	9.136	77	360850	39.335	ng/ul#	91
23) Isophorone	9.672	82	640547	36.534	ng/ul	95
25) 2-Nitrophenol	9.854	139	172359	41.664	ng/ul#	84
26) 2,4-Dimethylphenol	9.913	107	364551	38.099	ng/ul	93
27) Bis(2-Chloroethoxy)met...	10.148	93	396331	37.714	ng/ul	99
29) 2,4-Dichlorophenol	10.389	162	284550	38.475	ng/ul	96
30) Naphthalene	10.795	128	992669	39.120	ng/ul	99
32) 4-Chloroaniline	10.901	127	390050	35.317	ng/ul	99
33) Hexachlorobutadiene	11.089	225	207996	40.808	ng/ul	94
34) Caprolactam	11.672	113	60569	23.948	ng/ul	84
35) 4-Chloro-3-methylphenol	12.025	107	323563	35.843	ng/ul	88
36) 2-Methylnaphthalene	12.401	142	648736	37.122	ng/ul	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM013122\
 Data File : BM033927.D
 Acq On : 31 Jan 2022 15:35
 Operator : CG/JU
 Sample : SST04028
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST040028

Quant Time: Jan 31 16:47:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM013122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 31 16:44:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.619	142	652334	36.236	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.772	216	340223	42.075	ng/ul	99
40) Hexachlorocyclopentadiene	12.754	237	245266	46.192	ng/ul	100
41) 2,4,6-Trichlorophenol	13.007	196	217807	41.915	ng/ul	97
42) 2,4,5-Trichlorophenol	13.077	196	230709	40.130	ng/ul	93
43) 1,1'-Biphenyl	13.407	154	855502	40.314	ng/ul	100
44) 2-Chloronaphthalene	13.448	162	647172	40.586	ng/ul	98
45) 2-Nitroaniline	13.724	65	619	0.119	ng/ul	92
47) Dimethylphthalate	14.024	163	789183	37.024	ng/ul	100
48) 2,6-Dinitrotoluene	14.142	165	160669	39.324	ng/ul	94
50) Acenaphthylene	14.289	152	1046589	39.179	ng/ul	99
51) 3-Nitroaniline	14.466	138	160067	38.123	ng/ul	91
52) Acenaphthene	14.630	153	679731	38.656	ng/ul	96
53) 2,4-Dinitrophenol	14.671	184	92908	37.784	ng/ul	87
55) 4-Nitrophenol	14.771	109	131261	33.112	ng/ul	86
56) Dibenzofuran	14.965	168	955906	37.903	ng/ul	99
57) 2,4-Dinitrotoluene	14.924	165	227640	37.212	ng/ul#	78
58) 2,3,4,6-Tetrachlorophenol	15.189	232	185965	38.165	ng/ul#	99
59) Diethylphthalate	15.383	149	802147	35.619	ng/ul	100
61) Fluorene	15.613	166	757985	37.005	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.607	204	383434	37.298	ng/ul	99
63) 4-Nitroaniline	15.624	138	159912	39.533	ng/ul	85
66) 4,6-Dinitro-2-methylph...	15.689	198	127566	41.202	ng/ul	90
67) N-Nitrosodiphenylamine	15.818	169	645645	41.059	ng/ul	99
68) 4-Bromophenyl-phenylether	16.495	248	222840	43.301	ng/ul	97
69) Hexachlorobenzene	16.618	284	252614	40.307	ng/ul	96
70) Atrazine	16.765	200	236716	37.822	ng/ul	99
71) Pentachlorophenol	16.954	266	158982	38.778	ng/ul	97
72) Phenanthrene	17.348	178	1146342	38.806	ng/ul	99
74) Anthracene	17.436	178	1166972	38.899	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.377	216	362062	46.470	ng/uL	98
76) Pentachlorobenzene	14.889	250	318778	43.559	ng/uL	98
77) Carbazole	17.701	167	1004049	37.309	ng/ul	99
78) Di-n-butylphthalate	18.271	149	1279780	37.763	ng/ul	98
80) Fluoranthene	19.353	202	1260045	43.347	ng/ul	99
82) Pyrene	19.718	202	1272262	42.336	ng/ul	99
83) Butylbenzylphthalate	20.606	149	533903	41.067	ng/ul	88
84) 3,3'-Dichlorobenzidine	21.383	252	377324	37.453	ng/ul	96
85) Benzo(a)anthracene	21.453	228	1150056	38.792	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.383	149	763530	37.447	ng/ul	100
87) Chrysene	21.506	228	1121188	38.953	ng/ul	99
89) Di-n-octyl phthalate	22.300	149	1271913	34.282	ng/ul	100
90) Benzo(b)fluoranthene	23.130	252	1209932	39.376	ng/ul	99
91) Benzo(k)fluoranthene	23.183	252	1074345	37.564	ng/ul	99
93) Benzo(a)pyrene	23.759	252	1148727	39.974	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.365	276	1356483	46.905	ng/ul	96
95) Dibenzo(a,h)anthracene	26.377	278	1179880	47.345	ng/ul	97
96) Benzo(g,h,i)perylene	27.129	276	1147445	48.621	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 15.119 ng/uL
RT: 3.325 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion: 88 Resp: 48745
Ion Ratio Lower Upper
88 100
43 31.2 36.2 54.4#
58 78.4 68.5 102.7



#5
Pyridine
Concen: 37.295 ng/uL
RT: 3.737 min Scan# 85
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion: 79 Resp: 328034
Ion Ratio Lower Upper
79 100
52 80.2 78.1 117.1
51 39.9 36.0 54.0









#6
Benzaldehyde
Concen: 38.446 ng/ul
RT: 7.060 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion: 77 Resp: 231218
Ion Ratio Lower Upper
77 100
105 97.0 65.6 98.4
106 92.8 60.6 90.8#



#8
Phenol
Concen: 35.848 ng/ul
RT: 7.113 min Scan# 659
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion: 94 Resp: 383821
Ion Ratio Lower Upper
94 100
65 31.1 31.7 47.5#
66 44.7 45.4 68.0#









#10
Bis(2-Chloroethyl)ether
Concen: 36.242 ng/ul
RT: 7.348 min Scan# 699
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion: 93 Resp: 302250
Ion Ratio Lower Upper
93 100
63 77.5 71.7 107.5
95 31.9 26.3 39.5



#12
2-Chlorophenol
Concen: 37.795 ng/ul
RT: 7.489 min Scan# 723
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion: 128 Resp: 316347
Ion Ratio Lower Upper
128 100
64 49.7 49.4 74.2
130 31.1 26.8 40.2





#13

2-Methylphenol

Concen: 35.130 ng/ul

RT: 8.378 min Scan# 81

Delta R.T. 0.006 min

Lab File: BM033927.D

Acq: 31 Jan 2022 15:35

Instrument :

BNA_M

ClientSampleId :





SSTD040028

Tgt Ion:108 Resp: 289223

Ion Ratio Lower Upper

108 100

107 90.4 70.2 105.4



#14

2,2'-oxybis(1-Chloropropane)

Concen: 34.375 ng/ul

RT: 8.472 min Scan# 890

Delta R.T. 0.012 min

Lab File: BM033927.D

Acq: 31 Jan 2022 15:35




Tgt Ion: 45 Resp: 403310


Ion Ratio Lower Upper

45 100

77 15.6 9.9 14.9#

79 11.2 8.3 12.5









#16
Acetophenone
Concen: 34.426 ng/ul
RT: 8.760 min Scan# 91
Delta R.T. 0.006 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:105 Resp: 467285
Ion Ratio Lower Upper
105 100
77 78.9 73.1 109.7
51 29.7 30.5 45.7#



#17
N-Nitroso-di-n-propylamine
Concen: 33.766 ng/ul
RT: 8.748 min Scan# 937
Delta R.T. 0.006 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion: 70 Resp: 237347
Ion Ratio Lower Upper
70 100
42 60.7 56.9 85.3
101 10.0 6.9 10.3
130 20.9 13.8 20.8#









#18
4-Methylphenol
Concen: 34.159 ng/ul
RT: 8.707 min Scan# 911
Delta R.T. 0.006 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:108 Resp: 307399
Ion Ratio Lower Upper
108 100
107 115.7 87.4 131.2



#19
Hexachloroethane
Concen: 39.008 ng/ul
RT: 9.019 min Scan# 983
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion:117 Resp: 138459
Ion Ratio Lower Upper
117 100
201 97.4 69.6 104.4
199 58.9 40.6 60.8









#22
Nitrobenzene
Concen: 39.335 ng/ul
RT: 9.136 min Scan# 1003
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion: 77 Resp: 360850
Ion Ratio Lower Upper
77 100
123 44.0 29.0 43.6#
65 14.3 11.4 17.2



#23
Isophorone
Concen: 36.534 ng/ul
RT: 9.672 min Scan# 1094
Delta R.T. 0.006 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion: 82 Resp: 640547
Ion Ratio Lower Upper
82 100
95 7.8 6.5 9.7
138 17.5 11.8 17.8









#25
2-Nitrophenol
Concen: 41.664 ng/ul
RT: 9.854 min Scan# 1125
Delta R.T. 0.006 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:139 Resp: 172359
Ion Ratio Lower Upper
139 100
65 58.2 59.0 88.6#
109 39.0 37.2 55.8



#26
2,4-Dimethylphenol
Concen: 38.099 ng/ul
RT: 9.913 min Scan# 1135
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion:107 Resp: 364551
Ion Ratio Lower Upper
107 100
121 49.1 34.8 52.2
122 80.6 60.0 90.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 37.714 ng/ul
 RT: 10.148 min Scan# 1175
 Delta R.T. -0.000 min
 Lab File: BM033927.D
 Acq: 31 Jan 2022 15:35




Instrument :
 BNA_M
 ClientSampleId :
 SSTD040028


Tgt Ion: 93 Resp: 396331
 Ion Ratio Lower Upper
 93 100
 95 31.8 25.3 37.9
 123 11.0 7.9 11.9



#29
 2,4-Dichlorophenol
 Concen: 38.475 ng/ul
 RT: 10.389 min Scan# 1216
 Delta R.T. -0.000 min
 Lab File: BM033927.D
 Acq: 31 Jan 2022 15:35

Tgt Ion: 162 Resp: 284550
 Ion Ratio Lower Upper
 162 100
 164 65.3 50.0 75.0
 98 40.5 30.6 45.8









#30
Naphthalene
Concen: 39.120 ng/ul
RT: 10.795 min Scan# 1285
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:128 Resp: 992669
Ion Ratio Lower Upper
128 100
129 10.8 9.1 13.7
127 12.7 10.2 15.4



#32
4-Chloroaniline
Concen: 35.317 ng/ul
RT: 10.901 min Scan# 1303
Delta R.T. 0.006 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion:127 Resp: 390050
Ion Ratio Lower Upper
127 100
129 32.3 25.4 38.0









#33
Hexachlorobutadiene
Concen: 40.808 ng/ul
RT: 11.089 min Scan# 1335
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:225 Resp: 207996
Ion Ratio Lower Upper
225 100
223 60.4 52.2 78.2
227 61.9 52.8 79.2



#34
Caprolactam
Concen: 23.948 ng/ul
RT: 11.672 min Scan# 1434
Delta R.T. 0.006 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion:113 Resp: 60569
Ion Ratio Lower Upper
113 100
55 177.1 157.9 236.9
56 139.9 131.8 197.6









#35
4-Chloro-3-methylphenol
Concen: 35.843 ng/ul
RT: 12.025 min Scan# 1494
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:107 Resp: 323563
Ion Ratio Lower Upper
107 100
144 23.6 17.1 25.7
142 72.7 49.4 74.2



#36
2-Methylnaphthalene
Concen: 37.122 ng/ul
RT: 12.401 min Scan# 1558
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion:142 Resp: 648736
Ion Ratio Lower Upper
142 100
141 89.8 71.1 106.7









#37
1-Methylnaphthalene
Concen: 36.236 ng/ul
RT: 12.619 min Scan# 1595
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:142 Resp: 652334
Ion Ratio Lower Upper
142 100
141 94.3 75.5 113.3
116 4.4 4.0 6.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 42.075 ng/ul
RT: 12.772 min Scan# 1621
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion:216 Resp: 340223
Ion Ratio Lower Upper
216 100
214 77.2 61.3 91.9
179 22.4 18.0 27.0
108 18.7 16.6 25.0









#40
Hexachlorocyclopentadiene
Concen: 46.192 ng/ul
RT: 12.754 min Scan# 1618
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:237 Resp: 245266
Ion Ratio Lower Upper
237 100
235 63.3 50.6 76.0
272 12.0 9.6 14.4



#41
2,4,6-Trichlorophenol
Concen: 41.915 ng/ul
RT: 13.007 min Scan# 1661
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion:196 Resp: 217807
Ion Ratio Lower Upper
196 100
198 95.1 78.9 118.3
200 31.0 23.7 35.5









#42
2,4,5-Trichlorophenol
Concen: 40.130 ng/ul
RT: 13.077 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:196 Resp: 230709
Ion Ratio Lower Upper
196 100
198 95.8 82.2 123.2
200 31.6 27.9 41.9



#43
1,1'-Biphenyl
Concen: 40.314 ng/ul
RT: 13.407 min Scan# 1729
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion:154 Resp: 855502
Ion Ratio Lower Upper
154 100
153 41.3 33.1 49.7
76 13.0 10.7 16.1









#44
2-Chloronaphthalene
Concen: 40.586 ng/ul
RT: 13.448 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:162 Resp: 647172
Ion Ratio Lower Upper
162 100
164 32.5 26.2 39.4
127 44.6 33.7 50.5



#45
2-Nitroaniline
Concen: 0.119 ng/ul
RT: 13.724 min Scan# 1783
Delta R.T. 0.082 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion: 65 Resp: 619
Ion Ratio Lower Upper
65 100
92 66.6 47.8 71.6
138 70.6 61.0 91.6





#47

Dimethylphthalate

Concen: 37.024 ng/ul

RT: 14.024 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM033927.D

Acq: 31 Jan 2022 15:35

Instrument :

BNA_M

ClientSampleId :

SSTD040028





Tgt Ion:163 Resp: 789183

Ion Ratio Lower Upper

163 100

194 3.9 3.3 4.9

164 10.2 8.2 12.2



#48

2,6-Dinitrotoluene

Concen: 39.324 ng/ul

RT: 14.142 min Scan# 1854

Delta R.T. -0.000 min

Lab File: BM033927.D

Acq: 31 Jan 2022 15:35




Tgt Ion:165 Resp: 160669


Ion Ratio Lower Upper

165 100

89 64.3 55.7 83.5

121 22.2 20.0 30.0









#50
Acenaphthylene
Concen: 39.179 ng/ul
RT: 14.289 min Scan# 1879
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:152 Resp: 1046589
Ion Ratio Lower Upper
152 100
151 20.3 15.8 23.8
153 13.2 10.1 15.1



#51
3-Nitroaniline
Concen: 38.123 ng/ul
RT: 14.466 min Scan# 1909
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion:138 Resp: 160067
Ion Ratio Lower Upper
138 100
108 13.0 10.8 16.2
92 126.8 110.6 166.0









#52
Acenaphthene
Concen: 38.656 ng/ul
RT: 14.630 min Scan# 1937
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:153 Resp: 679731
Ion Ratio Lower Upper
153 100
152 46.5 39.4 59.2
154 86.8 72.7 109.1



#53
2,4-Dinitrophenol
Concen: 37.784 ng/ul
RT: 14.671 min Scan# 1944
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion:184 Resp: 92908
Ion Ratio Lower Upper
184 100
63 74.9 73.8 110.8
154 73.6 63.5 95.3









#55
4-Nitrophenol
Concen: 33.112 ng/ul
RT: 14.771 min Scan# 1960
Delta R.T. 0.006 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:109 Resp: 131261
Ion Ratio Lower Upper
109 100
139 107.7 77.6 116.4
65 119.8 112.2 168.4



#56
Dibenzofuran
Concen: 37.903 ng/ul
RT: 14.965 min Scan# 1994
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion:168 Resp: 955906
Ion Ratio Lower Upper
168 100
139 41.6 34.0 51.0









#57
2,4-Dinitrotoluene
Concen: 37.212 ng/ul
RT: 14.924 min Scan# 1987
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:165 Resp: 227640
Ion Ratio Lower Upper
165 100
63 47.1 51.8 77.8#
182 2.8 2.6 4.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 38.165 ng/ul
RT: 15.189 min Scan# 2032
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion:232 Resp: 185965
Ion Ratio Lower Upper
232 100
131 52.3 41.4 62.2
130 3.0 2.0 3.0#
166 33.4 27.4 41.0





#59

Diethylphthalate

Concen: 35.619 ng/ul

RT: 15.383 min Scan# 2065

Delta R.T. -0.000 min

Lab File: BM033927.D

Acq: 31 Jan 2022 15:35

Instrument :

BNA_M

ClientSampleId :

SSTD040028





Tgt Ion:149 Resp: 802147

Ion Ratio Lower Upper

149 100

177 21.4 17.1 25.7

150 12.5 9.9 14.9



#61

Fluorene

Concen: 37.005 ng/ul

RT: 15.613 min Scan# 2104

Delta R.T. -0.000 min

Lab File: BM033927.D

Acq: 31 Jan 2022 15:35




Tgt Ion:166 Resp: 757985


Ion Ratio Lower Upper

166 100

165 98.1 77.9 116.9

167 13.3 10.3 15.5





#62

4-Chlorophenyl-phenylether

Concen: 37.298 ng/ul

RT: 15.607 min Scan# 2103

Delta R.T. -0.000 min

Lab File: BM033927.D

Acq: 31 Jan 2022 15:35

Instrument :

BNA_M

ClientSampleId :

SSTD040028





Tgt Ion:204 Resp: 383434

Ion Ratio Lower Upper

204 100

206 33.2 27.8 41.8

141 68.9 54.9 82.3



#63

4-Nitroaniline

Concen: 39.533 ng/ul

RT: 15.624 min Scan# 2106

Delta R.T. 0.006 min

Lab File: BM033927.D

Acq: 31 Jan 2022 15:35




Tgt Ion:138 Resp: 159912


Ion Ratio Lower Upper

138 100

92 52.4 48.0 72.0

108 93.4 89.5 134.3









#66
4,6-Dinitro-2-methylphenol
Concen: 41.202 ng/ul
RT: 15.689 min Scan# 2116
Delta R.T. 0.006 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:198 Resp: 127566
Ion Ratio Lower Upper
198 100
182 4.3 4.2 6.2
77 26.4 26.3 39.5



#67
N-Nitrosodiphenylamine
Concen: 41.059 ng/ul
RT: 15.818 min Scan# 2139
Delta R.T. 0.006 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion:169 Resp: 645645
Ion Ratio Lower Upper
169 100
168 66.5 52.9 79.3
167 35.8 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 43.301 ng/ul
RT: 16.495 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:248 Resp: 222840
Ion Ratio Lower Upper
248 100
250 97.4 79.0 118.4
141 90.9 69.7 104.5



#69
Hexachlorobenzene
Concen: 40.307 ng/ul
RT: 16.618 min Scan# 2275
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion:284 Resp: 252614
Ion Ratio Lower Upper
284 100
142 38.2 34.3 51.5
249 32.9 26.1 39.1





#70

Atrazine

Concen: 37.822 ng/ul

RT: 16.765 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM033927.D

Acq: 31 Jan 2022 15:35

Instrument :

BNA_M

ClientSampleId :

SSTD040028





Tgt Ion:200 Resp: 236716

Ion Ratio Lower Upper

200 100

173 27.6 20.6 30.8

215 46.2 36.9 55.3



#71

Pentachlorophenol

Concen: 38.778 ng/ul

RT: 16.954 min Scan# 2332

Delta R.T. -0.000 min

Lab File: BM033927.D

Acq: 31 Jan 2022 15:35




Tgt Ion:266 Resp: 158982


Ion Ratio Lower Upper

266 100

264 62.5 49.1 73.7

268 63.0 47.8 71.8









#72
Phenanthrene
Concen: 38.806 ng/u1
RT: 17.348 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:178 Resp: 1146342
Ion Ratio Lower Upper
178 100
179 15.4 11.7 17.5
176 19.1 15.1 22.7



#74
Anthracene
Concen: 38.899 ng/u1
RT: 17.436 min Scan# 2414
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion:178 Resp: 1166972
Ion Ratio Lower Upper
178 100
179 15.2 12.3 18.5
176 18.2 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 46.470 ng/uL
RT: 13.377 min Scan# 11
Delta R.T. 0.006 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:216 Resp: 362062
Ion Ratio Lower Upper
216 100
214 78.2 65.4 98.0
218 48.5 38.8 58.2



#76
Pentachlorobenzene
Concen: 43.559 ng/uL
RT: 14.889 min Scan# 1981
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion:250 Resp: 318778
Ion Ratio Lower Upper
250 100
248 62.8 51.6 77.4
252 63.6 52.0 78.0









#77
 Carbazole
 Concen: 37.309 ng/ul
 RT: 17.701 min Scan# 2459
 Delta R.T. -0.000 min
 Lab File: BM033927.D
 Acq: 31 Jan 2022 15:35




Instrument :
 BNA_M
 ClientSampleId :
 SSTD040028


Tgt Ion:167 Resp: 1004049
 Ion Ratio Lower Upper
 167 100
 166 21.4 16.9 25.3
 139 14.0 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 37.763 ng/ul
 RT: 18.271 min Scan# 2556
 Delta R.T. -0.000 min
 Lab File: BM033927.D
 Acq: 31 Jan 2022 15:35

Tgt Ion:149 Resp: 1279780
 Ion Ratio Lower Upper
 149 100
 150 8.7 7.0 10.6
 104 5.3 5.1 7.7









#80
Fluoranthene
Concen: 43.347 ng/u1
RT: 19.353 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:202 Resp: 1260045
Ion Ratio Lower Upper
202 100
101 11.0 8.2 12.4
100 8.6 6.9 10.3



#82
Pyrene
Concen: 42.336 ng/u1
RT: 19.718 min Scan# 2802
Delta R.T. 0.006 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion:202 Resp: 1272262
Ion Ratio Lower Upper
202 100
101 12.2 9.8 14.6
100 10.1 8.6 13.0









#83
 Butylbenzylphthalate
 Concen: 41.067 ng/ul
 RT: 20.606 min Scan# 2953
 Delta R.T. -0.000 min
 Lab File: BM033927.D
 Acq: 31 Jan 2022 15:35




Instrument :
 BNA_M
 ClientSampleId :
 SSTD040028


Tgt Ion:149 Resp: 533903
 Ion Ratio Lower Upper
 149 100
 91 73.5 69.3 103.9
 206 19.0 16.1 24.1



#84
 3,3'-Dichlorobenzidine
 Concen: 37.453 ng/ul
 RT: 21.383 min Scan# 3085
 Delta R.T. -0.000 min
 Lab File: BM033927.D
 Acq: 31 Jan 2022 15:35

Tgt Ion:252 Resp: 377324
 Ion Ratio Lower Upper
 252 100
 254 64.7 49.4 74.2
 126 11.3 8.2 12.4









#85
Benzo(a)anthracene
Concen: 38.792 ng/ul
RT: 21.453 min Scan# 3097
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:228 Resp: 1150056
Ion Ratio Lower Upper
228 100
229 19.3 15.0 22.4
226 27.1 21.4 32.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 37.447 ng/ul
RT: 21.383 min Scan# 3085
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion:149 Resp: 763530
Ion Ratio Lower Upper
149 100
167 27.2 21.8 32.8
279 4.3 3.2 4.8





#87

Chrysene

Concen: 38.953 ng/ul

RT: 21.506 min Scan# 3106

Delta R.T. -0.000 min

Lab File: BM033927.D

Acq: 31 Jan 2022 15:35

Instrument :

BNA_M

ClientSampleId :

SSTD040028





Tgt Ion:228 Resp: 1121188

Ion Ratio Lower Upper

228 100

226 29.8 24.5 36.7

229 19.5 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 34.282 ng/ul




RT: 22.300 min Scan# 3241


Delta R.T. -0.000 min

Lab File: BM033927.D

Acq: 31 Jan 2022 15:35

Tgt Ion:149 Resp: 1271913









#90
Benzo(b)fluoranthene
Concen: 39.376 ng/ul
RT: 23.130 min Scan# 3382
Delta R.T. -0.000 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35




Instrument :
BNA_M
ClientSampleId :
SSTD040028


Tgt Ion:252 Resp: 1209932
Ion Ratio Lower Upper
252 100
253 22.5 17.7 26.5
125 9.7 7.8 11.6



#91
Benzo(k)fluoranthene
Concen: 37.564 ng/ul
RT: 23.183 min Scan# 3391
Delta R.T. 0.006 min
Lab File: BM033927.D
Acq: 31 Jan 2022 15:35

Tgt Ion:252 Resp: 1074345
Ion Ratio Lower Upper
252 100
253 21.5 17.6 26.4
125 9.7 7.4 11.2





#93

Benzo(a)pyrene

Concen: 39.974 ng/ul

RT: 23.759 min Scan# 34

Delta R.T. 0.006 min

Lab File: BM033927.D

Acq: 31 Jan 2022 15:35

Instrument :

BNA_M

ClientSampleId :

SSTD040028





Tgt Ion:252 Resp: 1148727

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.0

125 10.9 7.8 11.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 46.905 ng/ul

RT: 26.365 min Scan# 3932

Delta R.T. 0.006 min

Lab File: BM033927.D

Acq: 31 Jan 2022 15:35




Tgt Ion:276 Resp: 1356483


Ion Ratio Lower Upper

276 100

138 24.0 16.5 24.7

277 25.4 20.6 31.0









#95
 Dibenzo(a,h)anthracene
 Concen: 47.345 ng/ul
 RT: 26.377 min Scan# 3933
 Delta R.T. 0.006 min
 Lab File: BM033927.D
 Acq: 31 Jan 2022 15:35

Instrument :
 BNA_M
 ClientSampleId :
 SSTD040028

Tgt Ion:278 Resp: 1179880
 Ion Ratio Lower Upper
 278 100
 139 15.0 11.0 16.4
 279 23.9 18.1 27.1



#96
 Benzo(g,h,i)perylene
 Concen: 48.621 ng/ul
 RT: 27.129 min Scan# 4062
 Delta R.T. 0.006 min
 Lab File: BM033927.D
 Acq: 31 Jan 2022 15:35

Tgt Ion:276 Resp: 1147445
 Ion Ratio Lower Upper
 276 100
 138 20.3 15.6 23.4
 277 23.5 18.7 28.1

