

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM013122\
 Data File : BM033928.D
 Acq On : 31 Jan 2022 16:12
 Operator : CG/JU
 Sample : SSTD08029
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD080029

Quant Time: Jan 31 16:48:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM013122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 31 16:44:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.931	152	120133	20.000	ng/ul	0.00
20) Naphthalene-d8	10.742	136	476044	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.571	164	274807	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.307	188	516384	20.000	ng/ul	0.00
79) Chrysene-d12	21.471	240	453239	20.000	ng/ul	0.00
88) Perylene-d12	23.865	264	462000	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.290	96	94917	29.684	ng/uL	0.00
4) Pyridine-d5	3.719	84	674557	75.968	ng/ul	0.00
7) Phenol-d5	7.090	99	795592	73.734	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.260	67	473969	72.134	ng/ul	0.00
11) 2-Chlorophenol-d4	7.460	132	654416	79.329	ng/ul	0.00
15) 4-Methylphenol-d8	8.648	113	622601	69.706	ng/ul	0.00
21) Nitrobenzene-d5	9.101	128	315105	83.817	ng/ul	0.00
24) 2-Nitrophenol-d4	9.825	143	353606	88.749	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.366	165	608790	79.701	ng/ul	0.00
31) 4-Chloroaniline-d4	10.878	131	800086	70.465	ng/ul	0.00
46) Dimethylphthalate-d6	13.983	166	1619434	74.795	ng/ul	0.00
49) Acenaphthylene-d8	14.266	160	2065806	80.875	ng/ul	0.00
54) 4-Nitrophenol-d4	14.765	143	291736	70.207	ng/ul	0.01
60) Fluorene-d10	15.560	176	1363124	75.887	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.677	200	276403	89.985	ng/ul	0.00
73) Anthracene-d10	17.407	188	2009820	80.210	ng/ul	0.00
81) Pyrene-d10	19.689	212	2120012	86.615	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.712	264	1968566	81.892	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.325	88	101851	30.515	ng/uL#	88
5) Pyridine	3.737	79	689894	75.763	ng/ul	85
6) Benzaldehyde	7.060	77	290145	46.600	ng/ul	85
8) Phenol	7.119	94	812920	73.338	ng/ul#	84
10) Bis(2-Chloroethyl)ether	7.354	93	629197	72.876	ng/ul	89
12) 2-Chlorophenol	7.495	128	671773	77.525	ng/ul#	88
13) 2-Methylphenol	8.378	108	612142	71.819	ng/ul	97
14) 2,2'-oxybis(1-Chloropr...	8.466	45	837153	68.921	ng/ul#	95
16) Acetophenone	8.760	105	976218	69.469	ng/ul#	88
17) N-Nitroso-di-n-propyla...	8.760	70	489134	67.216	ng/ul#	89
18) 4-Methylphenol	8.719	108	646219	69.364	ng/ul	98
19) Hexachloroethane	9.019	117	293642	79.910	ng/ul#	86
22) Nitrobenzene	9.142	77	751376	78.821	ng/ul#	90
23) Isophorone	9.678	82	1339340	73.513	ng/ul	96
25) 2-Nitrophenol	9.854	139	369407	85.934	ng/ul#	82
26) 2,4-Dimethylphenol	9.919	107	764148	76.853	ng/ul	96
27) Bis(2-Chloroethoxy)met...	10.154	93	808268	74.016	ng/ul	99
29) 2,4-Dichlorophenol	10.395	162	606488	78.917	ng/ul	98
30) Naphthalene	10.795	128	2041356	77.418	ng/ul	98
32) 4-Chloroaniline	10.901	127	813282	70.865	ng/ul	98
33) Hexachlorobutadiene	11.089	225	439880	83.053	ng/ul	98
34) Caprolactam	11.695	113	107744	40.995	ng/ul	80
35) 4-Chloro-3-methylphenol	12.030	107	673189	71.765	ng/ul	89
36) 2-Methylnaphthalene	12.407	142	1341934	73.896	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM013122\
 Data File : BM033928.D
 Acq On : 31 Jan 2022 16:12
 Operator : CG/JU
 Sample : SSTD08029
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD080029

Quant Time: Jan 31 16:48:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM013122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 31 16:44:50 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.624	142	1364402	72.937	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.772	216	713800	86.925	ng/ul	99
40) Hexachlorocyclopentadiene	12.760	237	540929	100.318	ng/ul	98
41) 2,4,6-Trichlorophenol	13.007	196	459064	86.991	ng/ul	98
42) 2,4,5-Trichlorophenol	13.083	196	489949	83.918	ng/ul	93
43) 1,1'-Biphenyl	13.413	154	1777634	82.486	ng/ul	98
44) 2-Chloronaphthalene	13.454	162	1352426	83.517	ng/ul	99
45) 2-Nitroaniline	13.654	65	430280	81.672	ng/ul#	77
47) Dimethylphthalate	14.030	163	1618040	74.748	ng/ul	100
48) 2,6-Dinitrotoluene	14.148	165	341651	82.341	ng/ul	90
50) Acenaphthylene	14.295	152	2177588	80.270	ng/ul	99
51) 3-Nitroaniline	14.471	138	300520	70.479	ng/ul	88
52) Acenaphthene	14.636	153	1391947	77.948	ng/ul	97
53) 2,4-Dinitrophenol	14.677	184	209489	83.892	ng/ul#	86
55) 4-Nitrophenol	14.783	109	276818	68.763	ng/ul	83
56) Dibenzofuran	14.965	168	1949183	76.106	ng/ul	100
57) 2,4-Dinitrotoluene	14.930	165	475541	76.548	ng/ul#	79
58) 2,3,4,6-Tetrachlorophenol	15.195	232	395307	79.887	ng/ul	97
59) Diethylphthalate	15.389	149	1645113	71.934	ng/ul	100
61) Fluorene	15.618	166	1564688	75.221	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.607	204	794826	76.132	ng/ul	98
63) 4-Nitroaniline	15.630	138	280814	68.360	ng/ul	87
66) 4,6-Dinitro-2-methylph...	15.689	198	275775	89.337	ng/ul	96
67) N-Nitrosodiphenylamine	15.818	169	1309446	83.520	ng/ul	98
68) 4-Bromophenyl-phenylether	16.501	248	464652	90.557	ng/ul	99
69) Hexachlorobenzene	16.618	284	520489	83.296	ng/ul	94
70) Atrazine	16.771	200	489895	78.508	ng/ul	98
71) Pentachlorophenol	16.959	266	341003	83.424	ng/ul	97
72) Phenanthrene	17.348	178	2310852	78.460	ng/ul	98
74) Anthracene	17.442	178	2342138	78.304	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.377	216	758981	97.703	ng/ul	97
76) Pentachlorobenzene	14.895	250	663753	90.968	ng/ul	98
77) Carbazole	17.706	167	2044457	76.196	ng/ul	98
78) Di-n-butylphthalate	18.271	149	2625226	77.694	ng/ul	99
80) Fluoranthene	19.359	202	2526812	86.847	ng/ul	99
82) Pyrene	19.718	202	2561184	85.150	ng/ul	99
83) Butylbenzylphthalate	20.606	149	1135196	87.240	ng/ul	87
84) 3,3'-Dichlorobenzidine	21.383	252	659889	65.440	ng/ul	97
85) Benzo(a)anthracene	21.453	228	2332404	78.601	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.383	149	1615317	79.151	ng/ul	99
87) Chrysene	21.506	228	2281912	79.208	ng/ul	99
89) Di-n-octyl phthalate	22.306	149	2726337	71.958	ng/ul	100
90) Benzo(b)fluoranthene	23.136	252	2438014	77.697	ng/ul	100
91) Benzo(k)fluoranthene	23.183	252	2294006	78.545	ng/ul	100
93) Benzo(a)pyrene	23.765	252	2387483	81.358	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.377	276	2827956	95.758	ng/ul	97
95) Dibenzo(a,h)anthracene	26.388	278	2455152	96.474	ng/ul	97
96) Benzo(g,h,i)perylene	27.147	276	2376160	98.597	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM013122\
 Data File : BM033928.D
 Acq On : 31 Jan 2022 16:12
 Operator : CG/JU
 Sample : SSTD08029
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD080029

Quant Time: Jan 31 16:48:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM013122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 31 16:44:50 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 30.515 ng/uL
RT: 3.325 min Scan# 15
Delta R.T. -0.000 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion: 88 Resp: 101851
Ion Ratio Lower Upper
88 100
43 31.6 36.2 54.4#
58 79.0 68.5 102.7



#5
Pyridine
Concen: 75.763 ng/uL
RT: 3.737 min Scan# 85
Delta R.T. -0.000 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion: 79 Resp: 689894
Ion Ratio Lower Upper
79 100
52 79.5 78.1 117.1
51 39.3 36.0 54.0









#6
Benzaldehyde
Concen: 46.600 ng/ul
RT: 7.060 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion: 77 Resp: 290145
Ion Ratio Lower Upper
77 100
105 93.8 65.6 98.4
106 90.1 60.6 90.8



#8
Phenol
Concen: 73.338 ng/ul
RT: 7.119 min Scan# 660
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion: 94 Resp: 812920
Ion Ratio Lower Upper
94 100
65 31.1 31.7 47.5#
66 44.4 45.4 68.0#









#10
Bis(2-Chloroethyl)ether
Concen: 72.876 ng/ul
RT: 7.354 min Scan# 700
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion: 93 Resp: 629197
Ion Ratio Lower Upper
93 100
63 76.1 71.7 107.5
95 32.2 26.3 39.5



#12
2-Chlorophenol
Concen: 77.525 ng/ul
RT: 7.495 min Scan# 724
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion:128 Resp: 671773
Ion Ratio Lower Upper
128 100
64 48.7 49.4 74.2#
130 32.0 26.8 40.2





#13

2-Methylphenol

Concen: 71.819 ng/ul

RT: 8.378 min Scan# 873

Delta R.T. 0.006 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12

Instrument :

BNA_M

ClientSampleId :





SSTD080029

Tgt Ion:108 Resp: 612142

Ion Ratio Lower Upper

108 100

107 90.4 70.2 105.4



#14

2,2'-oxybis(1-Chloropropane)

Concen: 68.921 ng/ul

RT: 8.466 min Scan# 889

Delta R.T. 0.006 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12




Tgt Ion: 45 Resp: 837153


Ion Ratio Lower Upper

45 100

77 15.4 9.9 14.9#

79 11.3 8.3 12.5









#16
Acetophenone
Concen: 69.469 ng/ul
RT: 8.760 min Scan# 91
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:105 Resp: 976218
Ion Ratio Lower Upper
105 100
77 79.9 73.1 109.7
51 30.1 30.5 45.7#



#17
N-Nitroso-di-n-propylamine
Concen: 67.216 ng/ul
RT: 8.760 min Scan# 939
Delta R.T. 0.018 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion: 70 Resp: 489134
Ion Ratio Lower Upper
70 100
42 60.6 56.9 85.3
101 10.5 6.9 10.3#
130 21.4 13.8 20.8#









#18
4-Methylphenol
Concen: 69.364 ng/ul
RT: 8.719 min Scan# 91
Delta R.T. 0.018 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:108 Resp: 646219
Ion Ratio Lower Upper
108 100
107 111.8 87.4 131.2



#19
Hexachloroethane
Concen: 79.910 ng/ul
RT: 9.019 min Scan# 983
Delta R.T. -0.000 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion:117 Resp: 293642
Ion Ratio Lower Upper
117 100
201 99.6 69.6 104.4
199 61.2 40.6 60.8#









#22
Nitrobenzene
Concen: 78.821 ng/ul
RT: 9.142 min Scan# 1004
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion: 77 Resp: 751376
Ion Ratio Lower Upper
77 100
123 44.2 29.0 43.6#
65 14.1 11.4 17.2



#23
Isophorone
Concen: 73.513 ng/ul
RT: 9.678 min Scan# 1095
Delta R.T. 0.012 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion: 82 Resp: 1339340
Ion Ratio Lower Upper
82 100
95 8.3 6.5 9.7
138 17.0 11.8 17.8









#25
2-Nitrophenol
Concen: 85.934 ng/ul
RT: 9.854 min Scan# 1125
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:139 Resp: 369407
Ion Ratio Lower Upper
139 100
65 56.2 59.0 88.6#
109 37.6 37.2 55.8



#26
2,4-Dimethylphenol
Concen: 76.853 ng/ul
RT: 9.919 min Scan# 1136
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion:107 Resp: 764148
Ion Ratio Lower Upper
107 100
121 46.7 34.8 52.2
122 78.3 60.0 90.0





#27

Bis(2-Chloroethoxy)methane

Concen: 74.016 ng/ul

RT: 10.154 min Scan# 1176

Delta R.T. 0.006 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12

Instrument :

BNA_M

ClientSampleId :

SSTD080029





Tgt Ion: 93 Resp: 808268

Ion Ratio Lower Upper

93 100

95 32.0 25.3 37.9

123 11.3 7.9 11.9



#29

2,4-Dichlorophenol

Concen: 78.917 ng/ul

RT: 10.395 min Scan# 1217

Delta R.T. 0.006 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12




Tgt Ion:162 Resp: 606488


Ion Ratio Lower Upper

162 100

164 64.2 50.0 75.0

98 37.4 30.6 45.8









#30
Naphthalene
Concen: 77.418 ng/ul
RT: 10.795 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:128 Resp: 2041356
Ion Ratio Lower Upper
128 100
129 11.0 9.1 13.7
127 13.7 10.2 15.4



#32
4-Chloroaniline
Concen: 70.865 ng/ul
RT: 10.901 min Scan# 1303
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion:127 Resp: 813282
Ion Ratio Lower Upper
127 100
129 32.8 25.4 38.0









#33
Hexachlorobutadiene
Concen: 83.053 ng/ul
RT: 11.089 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:225 Resp: 439880
Ion Ratio Lower Upper
225 100
223 63.5 52.2 78.2
227 64.4 52.8 79.2



#34
Caprolactam
Concen: 40.995 ng/ul
RT: 11.695 min Scan# 1438
Delta R.T. 0.029 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion:113 Resp: 107744
Ion Ratio Lower Upper
113 100
55 166.4 157.9 236.9
56 139.9 131.8 197.6









#35
4-Chloro-3-methylphenol
Concen: 71.765 ng/ul
RT: 12.030 min Scan# 1494
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:107 Resp: 673189
Ion Ratio Lower Upper
107 100
144 24.4 17.1 25.7
142 71.5 49.4 74.2



#36
2-Methylnaphthalene
Concen: 73.896 ng/ul
RT: 12.407 min Scan# 1559
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion:142 Resp: 1341934
Ion Ratio Lower Upper
142 100
141 91.1 71.1 106.7





#37

1-Methylnaphthalene

Concen: 72.937 ng/ul

RT: 12.624 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12

Instrument :

BNA_M

ClientSampleId :

SSTD080029





Tgt Ion:142 Resp: 1364402

Ion Ratio Lower Upper

142 100

141 92.1 75.5 113.3

116 4.3 4.0 6.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 86.925 ng/ul

RT: 12.772 min Scan# 1621

Delta R.T. -0.000 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12

Tgt Ion:216 Resp: 713800




Ion Ratio Lower Upper


216 100

214 78.0 61.3 91.9

179 22.7 18.0 27.0

108 20.4 16.6 25.0









#40
Hexachlorocyclopentadiene
Concen: 100.318 ng/ul
RT: 12.760 min Scan# 1619
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:237 Resp: 540929
Ion Ratio Lower Upper
237 100
235 61.9 50.6 76.0
272 12.4 9.6 14.4



#41
2,4,6-Trichlorophenol
Concen: 86.991 ng/ul
RT: 13.007 min Scan# 1661
Delta R.T. -0.000 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion:196 Resp: 459064
Ion Ratio Lower Upper
196 100
198 95.7 78.9 118.3
200 30.0 23.7 35.5









#42
2,4,5-Trichlorophenol
Concen: 83.918 ng/ul
RT: 13.083 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:196 Resp: 489949
Ion Ratio Lower Upper
196 100
198 96.5 82.2 123.2
200 30.5 27.9 41.9



#43
1,1'-Biphenyl
Concen: 82.486 ng/ul
RT: 13.413 min Scan# 1730
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion:154 Resp: 1777634
Ion Ratio Lower Upper
154 100
153 40.9 33.1 49.7
76 11.7 10.7 16.1









#44
2-Chloronaphthalene
Concen: 83.517 ng/ul
RT: 13.454 min Scan# 11
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:162 Resp: 1352426
Ion Ratio Lower Upper
162 100
164 32.1 26.2 39.4
127 43.0 33.7 50.5



#45
2-Nitroaniline
Concen: 81.672 ng/ul
RT: 13.654 min Scan# 1771
Delta R.T. 0.012 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion: 65 Resp: 430280
Ion Ratio Lower Upper
65 100
92 69.4 47.8 71.6
138 102.8 61.0 91.6#









#47
Dimethylphthalate
Concen: 74.748 ng/ul
RT: 14.030 min Scan# 1835
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:163 Resp: 1618040
Ion Ratio Lower Upper
163 100
194 4.0 3.3 4.9
164 10.1 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 82.341 ng/ul
RT: 14.148 min Scan# 1855
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion:165 Resp: 341651
Ion Ratio Lower Upper
165 100
89 60.6 55.7 83.5
121 21.8 20.0 30.0





#50

Acenaphthylene

Concen: 80.270 ng/ul

RT: 14.295 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12

Instrument :

BNA_M

ClientSampleId :

SSTD080029





Tgt Ion:152 Resp: 2177588

Ion Ratio Lower Upper

152 100

151 20.0 15.8 23.8

153 13.3 10.1 15.1



#51

3-Nitroaniline

Concen: 70.479 ng/ul

RT: 14.471 min Scan# 1910

Delta R.T. 0.006 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12




Tgt Ion:138 Resp: 300520


Ion Ratio Lower Upper

138 100

108 12.3 10.8 16.2

92 123.4 110.6 166.0









#52
Acenaphthene
Concen: 77.948 ng/ul
RT: 14.636 min Scan# 1938
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:153 Resp: 1391947
Ion Ratio Lower Upper
153 100
152 47.4 39.4 59.2
154 88.3 72.7 109.1



#53
2,4-Dinitrophenol
Concen: 83.892 ng/ul
RT: 14.677 min Scan# 1945
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion:184 Resp: 209489
Ion Ratio Lower Upper
184 100
63 73.5 73.8 110.8#
154 72.9 63.5 95.3









#55
4-Nitrophenol
Concen: 68.763 ng/ul
RT: 14.783 min Scan# 1960
Delta R.T. 0.018 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:109 Resp: 276818
Ion Ratio Lower Upper
109 100
139 109.9 77.6 116.4
65 116.9 112.2 168.4



#56
Dibenzofuran
Concen: 76.106 ng/ul
RT: 14.965 min Scan# 1994
Delta R.T. -0.000 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion:168 Resp: 1949183
Ion Ratio Lower Upper
168 100
139 42.7 34.0 51.0









#57
2,4-Dinitrotoluene
Concen: 76.548 ng/ul
RT: 14.930 min Scan# 1988
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:165 Resp: 475541
Ion Ratio Lower Upper
165 100
63 47.8 51.8 77.8#
182 2.8 2.6 4.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 79.887 ng/ul
RT: 15.195 min Scan# 2033
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion:232 Resp: 395307
Ion Ratio Lower Upper
232 100
131 49.3 41.4 62.2
130 2.7 2.0 3.0
166 32.2 27.4 41.0





#59

Diethylphthalate

Concen: 71.934 ng/ul

RT: 15.389 min Scan# 2065

Delta R.T. 0.006 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12

Instrument :

BNA_M

ClientSampleId :

SSTD080029





Tgt Ion:149 Resp: 1645113

Ion Ratio Lower Upper

149 100

177 21.3 17.1 25.7

150 12.7 9.9 14.9



#61

Fluorene

Concen: 75.221 ng/ul

RT: 15.618 min Scan# 2105

Delta R.T. 0.006 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12




Tgt Ion:166 Resp: 1564688


Ion Ratio Lower Upper

166 100

165 97.7 77.9 116.9

167 13.1 10.3 15.5









#62
4-Chlorophenyl-phenylether
Concen: 76.132 ng/ul
RT: 15.607 min Scan# 2103
Delta R.T. -0.000 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:204 Resp: 794826
Ion Ratio Lower Upper
204 100
206 31.7 27.8 41.8
141 68.0 54.9 82.3



#63
4-Nitroaniline
Concen: 68.360 ng/ul
RT: 15.630 min Scan# 2107
Delta R.T. 0.012 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion:138 Resp: 280814
Ion Ratio Lower Upper
138 100
92 54.1 48.0 72.0
108 94.7 89.5 134.3









#66
4,6-Dinitro-2-methylphenol
Concen: 89.337 ng/ul
RT: 15.689 min Scan# 2116
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:198 Resp: 275775
Ion Ratio Lower Upper
198 100
182 4.2 4.2 6.2
77 30.3 26.3 39.5



#67
N-Nitrosodiphenylamine
Concen: 83.520 ng/ul
RT: 15.818 min Scan# 2139
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion:169 Resp: 1309446
Ion Ratio Lower Upper
169 100
168 67.6 52.9 79.3
167 36.4 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 90.557 ng/ul
RT: 16.501 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:248 Resp: 464652
Ion Ratio Lower Upper
248 100
250 98.4 79.0 118.4
141 84.8 69.7 104.5



#69
Hexachlorobenzene
Concen: 83.296 ng/ul
RT: 16.618 min Scan# 2275
Delta R.T. -0.000 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion:284 Resp: 520489
Ion Ratio Lower Upper
284 100
142 37.6 34.3 51.5
249 34.3 26.1 39.1





#70

Atrazine

Concen: 78.508 ng/ul

RT: 16.771 min Scan# 21

Delta R.T. 0.006 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12

Instrument :

BNA_M

ClientSampleId :

SSTD080029





Tgt Ion:200 Resp: 489895

Ion Ratio Lower Upper

200 100

173 26.5 20.6 30.8

215 45.1 36.9 55.3



#71

Pentachlorophenol

Concen: 83.424 ng/ul

RT: 16.959 min Scan# 2333

Delta R.T. 0.006 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12




Tgt Ion:266 Resp: 341003


Ion Ratio Lower Upper

266 100

264 63.1 49.1 73.7

268 62.1 47.8 71.8









#72
Phenanthrene
Concen: 78.460 ng/ul
RT: 17.348 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:178 Resp: 2310852
Ion Ratio Lower Upper
178 100
179 15.5 11.7 17.5
176 19.3 15.1 22.7



#74
Anthracene
Concen: 78.304 ng/ul
RT: 17.442 min Scan# 2415
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion:178 Resp: 2342138
Ion Ratio Lower Upper
178 100
179 15.4 12.3 18.5
176 18.7 14.5 21.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 97.703 ng/uL

RT: 13.377 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12

Instrument :

BNA_M

ClientSampleId :

SSTD080029





Tgt Ion:216 Resp: 758981

Ion Ratio Lower Upper

216 100

214 77.5 65.4 98.0

218 47.7 38.8 58.2



#76

Pentachlorobenzene

Concen: 90.968 ng/uL

RT: 14.895 min Scan# 1982

Delta R.T. 0.006 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12




Tgt Ion:250 Resp: 663753


Ion Ratio Lower Upper

250 100

248 63.0 51.6 77.4

252 63.7 52.0 78.0









#77
 Carbazole
 Concen: 76.196 ng/ul
 RT: 17.706 min Scan# 2459
 Delta R.T. 0.006 min
 Lab File: BM033928.D
 Acq: 31 Jan 2022 16:12




Instrument :
 BNA_M
 ClientSampleId :
 SSTD080029


Tgt Ion:167 Resp: 2044457
 Ion Ratio Lower Upper
 167 100
 166 22.2 16.9 25.3
 139 14.1 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 77.694 ng/ul
 RT: 18.271 min Scan# 2556
 Delta R.T. -0.000 min
 Lab File: BM033928.D
 Acq: 31 Jan 2022 16:12

Tgt Ion:149 Resp: 2625226
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.0 10.6
 104 5.7 5.1 7.7









#80
Fluoranthene
Concen: 86.847 ng/u1
RT: 19.359 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:202 Resp: 2526812
Ion Ratio Lower Upper
202 100
101 10.4 8.2 12.4
100 7.9 6.9 10.3



#82
Pyrene
Concen: 85.150 ng/u1
RT: 19.718 min Scan# 2802
Delta R.T. 0.006 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion:202 Resp: 2561184
Ion Ratio Lower Upper
202 100
101 12.4 9.8 14.6
100 10.5 8.6 13.0









#83
Butylbenzylphthalate
Concen: 87.240 ng/ul
RT: 20.606 min Scan# 2953
Delta R.T. -0.000 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:149 Resp: 1135196
Ion Ratio Lower Upper
149 100
91 72.3 69.3 103.9
206 19.2 16.1 24.1



#84
3,3'-Dichlorobenzidine
Concen: 65.440 ng/ul
RT: 21.383 min Scan# 3085
Delta R.T. -0.000 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion:252 Resp: 659889
Ion Ratio Lower Upper
252 100
254 63.9 49.4 74.2
126 11.9 8.2 12.4









#85
Benzo(a)anthracene
Concen: 78.601 ng/ul
RT: 21.453 min Scan# 3097
Delta R.T. -0.000 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:228 Resp: 2332404
Ion Ratio Lower Upper
228 100
229 19.6 15.0 22.4
226 27.2 21.4 32.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 79.151 ng/ul
RT: 21.383 min Scan# 3085
Delta R.T. -0.000 min
Lab File: BM033928.D
Acq: 31 Jan 2022 16:12

Tgt Ion:149 Resp: 1615317
Ion Ratio Lower Upper
149 100
167 26.7 21.8 32.8
279 4.2 3.2 4.8





#87

Chrysene

Concen: 79.208 ng/ul

RT: 21.506 min Scan# 3106

Delta R.T. -0.000 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12

Instrument :

BNA_M

ClientSampleId :

SSTD080029





Tgt Ion:228 Resp: 2281912

Ion Ratio Lower Upper

228 100

226 30.1 24.5 36.7

229 19.8 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 71.958 ng/ul




RT: 22.306 min Scan# 3242


Delta R.T. 0.006 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12

Tgt Ion:149 Resp: 2726337





#90

Benzo(b)fluoranthene

Concen: 77.697 ng/ul

RT: 23.136 min Scan# 3382

Delta R.T. 0.006 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12

Instrument :

BNA_M

ClientSampleId :

SSTD080029





Tgt Ion:252 Resp: 2438014

Ion Ratio Lower Upper

252 100

253 21.9 17.7 26.5

125 9.8 7.8 11.6



#91

Benzo(k)fluoranthene

Concen: 78.545 ng/ul

RT: 23.183 min Scan# 3391

Delta R.T. 0.006 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12




Tgt Ion:252 Resp: 2294006


Ion Ratio Lower Upper

252 100

253 22.2 17.6 26.4

125 9.4 7.4 11.2





#93

Benzo(a)pyrene

Concen: 81.358 ng/ul

RT: 23.765 min Scan# 34

Delta R.T. 0.012 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12

Instrument :

BNA_M

ClientSampleId :

SSTD080029





Tgt Ion:252 Resp: 2387483

Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.0

125 10.4 7.8 11.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 95.758 ng/ul

RT: 26.377 min Scan# 3934

Delta R.T. 0.018 min

Lab File: BM033928.D

Acq: 31 Jan 2022 16:12




Tgt Ion:276 Resp: 2827956


Ion Ratio Lower Upper

276 100

138 23.6 16.5 24.7

277 25.5 20.6 31.0









#95
 Dibenzo(a,h)anthracene
 Concen: 96.474 ng/ul
 RT: 26.388 min Scan# 3936
 Delta R.T. 0.018 min
 Lab File: BM033928.D
 Acq: 31 Jan 2022 16:12

Instrument :
 BNA_M
 ClientSampleId :
 SSTD080029

Tgt Ion:278 Resp: 2455152
 Ion Ratio Lower Upper
 278 100
 139 15.1 11.0 16.4
 279 23.6 18.1 27.1



#96
 Benzo(g,h,i)perylene
 Concen: 98.597 ng/ul
 RT: 27.147 min Scan# 4065
 Delta R.T. 0.023 min
 Lab File: BM033928.D
 Acq: 31 Jan 2022 16:12

Tgt Ion:276 Resp: 2376160
 Ion Ratio Lower Upper
 276 100
 138 19.7 15.6 23.4
 277 23.6 18.7 28.1

