

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM020124\
 Data File : BM043949.D
 Acq On : 01 Feb 2024 16:24
 Operator : MA/JU
 Sample : P1239-12
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0FC2

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 02/02/2024
 Supervised By :Sohil Jodhani 02/02/2024

Quant Time: Feb 01 17:01:20 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM013124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 31 22:50:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.926	152	9214	0.400	ng/ul	0.00
4) Naphthalene-d8	10.726	136	27990	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.566	164	16769	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.317	188	34419m	0.400	ng/ul	0.00
17) Chrysene-d12	21.517	240	25763m	0.400	ng/ul	0.00
23) Perylene-d12	23.926	264	25950m	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.365	96	38616	3.257	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.316	152	14013	0.382	ng/ul	0.00
18) Fluoranthene-d10	19.350	212	34429m	0.426	ng/ul	0.00
Target Compounds						Qvalue
5) Naphthalene	10.776	128	2468	0.031	ng/ul#	57
12) Fluorene	15.621	166	2220	0.032	ng/ul#	71
14) Pentachlorophenol	16.963	266	1280	0.157	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM020124\
Data File : BM043949.D
Acq On : 01 Feb 2024 16:24
Operator : MA/JU
Sample : P1239-12
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0FC2

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 02/02/2024
Supervised By :Sohil Jodhani 02/02/2024

Quant Time: Feb 01 17:01:20 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM013124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 31 22:50:05 2024
Response via : Initial Calibration

