

Data Path : Z:\HPCHEM1\BNA M\DATA\BM020216\
 Data File : BM004165.D
 Acq On : 02 Feb 2016 15:55
 Operator : SJ/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 03 00:53:11 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM012216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 22 14:46:41 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	129	0.00
2	1,4-Dioxane	0.547	0.448	18.1	125	0.00
3	n-Nitrosodimethylamine	0.446	0.432	3.1	137	-0.02
4 S	2-Fluorophenol	1.043	0.975	6.5	133	-0.02
5 S	Phenol-d6	1.259	1.217	3.3	139	0.00
6	bis(2-Chloroethyl)ether	1.167	1.162	0.4	142	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	134	0.00
8 S	Nitrobenzene-d5	0.219	0.222	-1.4	151#	0.00
9	Nitrobenzene	0.250	0.251	-0.4	152#	0.00
10	Naphthalene	1.165	1.110	4.7	139	0.00
11	Hexachlorobutadiene	0.180	0.167	7.2	137	0.00
12	2-Methylnaphthalene	0.762	0.732	3.9	141	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	144	0.00
14 S	2,4,6-Tribromophenol	0.145	0.134	7.6	153#	0.00
15 S	2-Fluorobiphenyl	1.532	1.377	10.1	140	0.00
16	Acenaphthylene	2.158	2.035	5.7	150	0.00
17	Acenaphthene	1.634	1.380	15.5	137	0.00
18	Fluorene	1.987	1.723	13.3	139	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	130	0.00
20	4-Bromophenyl-phenylether	0.163	0.187	-14.7	153#	0.00
21	Hexachlorobenzene	0.171	0.193	-12.9	154#	0.00
22	Pentachlorophenol	0.043	0.050	-16.3	203#	0.00
23	Phenanthrene	1.074	1.178	-9.7	155#	0.00
24	Anthracene	1.142	1.207	-5.7	148	0.00
25	Fluoranthene	1.355	1.379	-1.8	143	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	133	0.00
27	Benzydine	0.241	0.235	2.5	147	0.00
28	Pvrene	1.423	1.420	0.2	147	0.00
29 S	Terphenyl-d14	0.643	0.642	0.2	149	0.00
30	Benzo(a)anthracene	1.518	1.536	-1.2	151#	0.00
31	3,3'-Dichlorobenzidine	0.338	0.292	13.6	140	0.00
32	Chrysene	1.372	1.112	19.0	122	0.00
33	Bis(2-ethylhexyl)phthalate	0.595	0.503	15.5	159#	-0.02
34	Indeno(1,2,3-cd)pyrene	1.463	1.135	22.4	117	-0.01
35 I	Perylene-d12	1.000	1.000	0.0	122	0.00
36	Benzo(b)fluoranthene	1.551	1.400	9.7	118	-0.01
37	Benzo(k)fluoranthene	1.404	1.441	-2.6	141	0.00
38 C	Benzo(a)pyrene	1.366	1.291	5.5	129	0.00
39	Dibenzo(a,h)anthracene	1.227	1.049	14.5	116	-0.01
40	Benzo(a,h,i)perylene	1.339	1.146	14.4	117	-0.01

Data Path : Z:\HPCHEM1\BNA M\DATA\BM020216\
Data File : BM004165.D
Acq On : 02 Feb 2016 15:55
Operator : SJ/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Feb 03 00:53:11 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM012216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 22 14:46:41 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0