

Data Path : Z:\HPCHEM1\BNA M\DATA\BM020216\
 Data File : BM004165.D
 Acq On : 02 Feb 2016 15:55
 Operator : SJ/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 03 00:53:11 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM012216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 22 14:46:41 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	129	0.00
2	1,4-Dioxane	1.000	0.922	7.8	125	0.00
3	n-Nitrosodimethylamine	1.000	0.970	3.0	137	-0.02
4 S	2-Fluorophenol	1.000	0.935	6.5	133	-0.02
5 S	Phenol-d6	1.000	0.966	3.4	139	0.00
6	bis(2-Chloroethyl)ether	1.000	0.996	0.4	142	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	134	0.00
8 S	Nitrobenzene-d5	1.000	1.011	-1.1	151	0.00
9	Nitrobenzene	1.000	0.988	1.2	152	0.00
10	Naphthalene	1.000	0.953	4.7	139	0.00
11	Hexachlorobutadiene	1.000	0.932	6.8	137	0.00
12	2-Methylnaphthalene	1.000	0.961	3.9	141	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	144	0.00
14 S	2,4,6-Tribromophenol	1.000	0.946	5.4	153	0.00
15 S	2-Fluorobiphenyl	1.000	0.899	10.1	140	0.00
16	Acenaphthylene	1.000	0.943	5.7	150	0.00
17	Acenaphthene	1.000	0.845	15.5	137	0.00
18	Fluorene	1.000	0.867	13.3	139	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	130	0.00
20	4-Bromophenyl-phenylether	1.000	1.148	-14.8	153	0.00
21	Hexachlorobenzene	1.000	1.127	-12.7	154	0.00
22	Pentachlorophenol	1.000	1.167	-16.7	203	0.00
23	Phenanthrene	1.000	1.122	-12.2	155	0.00
24	Anthracene	1.000	1.057	-5.7	148	0.00
25	Fluoranthene	1.000	1.018	-1.8	143	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	133	0.00
27	Benzydine	1.000	1.075	-7.5	147	0.00
28	Pvrene	1.000	1.012	-1.2	147	0.00
29 S	Terphenyl-d14	1.000	1.012	-1.2	149	0.00
30	Benzo(a)anthracene	1.000	1.013	-1.3	151	0.00
31	3,3'-Dichlorobenzidine	1.000	0.970	3.0	140	0.00
32	Chrysene	1.000	0.806	19.4	122	0.00
33	Bis(2-ethylhexyl)phthalate	1.000	1.042	-4.2	159	-0.02
34	Indeno(1,2,3-cd)pyrene	1.000	0.773	22.7	117	-0.01
35 I	Perylene-d12	5.000	5.000	0.0	122	0.00
36	Benzo(b)fluoranthene	1.000	0.903	9.7	118	-0.01
37	Benzo(k)fluoranthene	1.000	1.027	-2.7	141	0.00
38 C	Benzo(a)pyrene	1.000	0.945	5.5	129	0.00
39	Dibenzo(a,h)anthracene	1.000	0.855	14.5	116	-0.01
40	Benzo(a,h,i)perylene	1.000	0.856	14.4	117	-0.01

Data Path : Z:\HPCHEM1\BNA M\DATA\BM020216\
Data File : BM004165.D
Acq On : 02 Feb 2016 15:55
Operator : SJ/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Feb 03 00:53:11 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM012216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 22 14:46:41 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)
----------	--------	-------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0