


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM020217\
 Data File : BM008952.D
 Acq On : 02 Feb 2017 08:28
 Operator : SJ/MA
 Sample : DFTPP63
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP63

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM121416.M
 Title : SVOA CALIBRATION
 Last Update : Mon Dec 19 15:00:56 2016



AutoFind: Scans 2572, 2573, 2574; Background Corrected with Scan 2566

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	76.5	77533	PASS
68	69	0.00	2	1.7	1111	PASS
69	198	0.00	100	63.9	64813	PASS
70	69	0.00	2	0.3	171	PASS
127	198	10	80	48.6	49298	PASS
197	198	0.00	2	0.9	935	PASS
198	198	100	100	100.0	101362	PASS
199	198	5	9	6.7	6811	PASS
275	198	10	60	30.6	31066	PASS
365	198	1	100	5.9	6005	PASS
441	443	0.01	100	85.9	12919	PASS
442	198	50	100	74.1	75117	PASS
443	442	15	24	20.0	15044	PASS

m/z	Abundance
36.90	399.0
38.00	1614.0
39.10	10587.0
40.10	475.0
41.20	476.0
43.80	378.0
45.10	370.0
48.90	982.0
50.00	26752.0
51.00	106624.0
52.00	5186.0
54.90	839.0
56.00	3753.0
57.00	6916.0
58.00	376.0
61.00	1203.0
62.10	1702.0
63.00	4914.0
64.10	680.0
65.00	1605.0
67.90	1210.0
69.00	81304.0
73.10	2130.0
74.00	12074.0
75.10	13988.0
76.00	6016.0
77.10	96208.0
78.10	8295.0
78.90	7732.0
80.00	3973.0
81.00	7852.0
82.00	1588.0
82.90	991.0
84.90	1462.0
86.00	3165.0
87.10	1835.0
87.90	389.0
90.90	1136.0
92.00	1620.0
93.00	16784.0
93.90	899.0
95.10	319.0
96.00	772.0
97.90	13024.0
99.00	8838.0
100.00	411.0
101.10	3627.0
102.00	367.0
103.00	1370.0
104.00	2688.0
105.10	2063.0
105.80	1267.0
107.00	28816.0
108.10	2491.0
109.00	909.0
110.00	34672.0
111.00	8626.0
111.80	850.0
115.90	2180.0
117.00	45776.0
117.90	2803.0
120.10	680.0
122.00	1691.0
123.00	3272.0
123.90	702.0
125.00	782.0
125.90	336.0
127.00	61856.0
127.90	4560.0
129.00	30552.0
130.00	2393.0
131.10	857.0
133.10	500.0
134.10	1019.0
135.10	2871.0
136.10	1614.0
137.00	1671.0
137.80	463.0
139.90	358.0
140.90	5089.0
142.10	581.0
142.90	909.0
144.00	450.0
146.00	1797.0
147.10	3779.0
148.00	8854.0
149.00	2044.0
150.00	429.0
151.00	376.0
152.90	1316.0
154.00	765.0

Instrument :
BNA_M
ClientSampleId :
DFTPP63

155.00	2854.0
156.00	2808.0
157.20	579.0
157.80	625.0
158.90	1139.0
160.00	2136.0
160.90	2820.0
162.00	469.0
163.00	543.0
164.80	1706.0
166.10	1062.0
167.00	13601.0
168.00	8408.0
169.00	1891.0
170.20	353.0
171.10	442.0
172.00	793.0
173.00	977.0
173.90	1601.0
175.00	2802.0
175.80	515.0
176.90	2224.0
177.80	496.0
179.00	6910.0
180.00	4717.0
180.80	1320.0
181.10	1289.0
181.90	504.0
183.90	794.0
184.90	3503.0
186.00	22536.0
187.00	6918.0
187.80	395.0
188.20	607.0
189.10	2416.0
190.00	435.0
190.90	1345.0
191.90	2691.0
192.90	3968.0
194.20	321.0
196.00	2275.0
196.70	851.0
197.90	114728.0
199.00	8015.0
199.90	1068.0
202.90	1640.0
204.00	6027.0
205.00	10937.0
206.00	40256.0
207.00	6399.0
208.00	2429.0
209.00	819.0
210.20	897.0
211.00	1620.0
211.70	336.0
215.00	589.0
215.80	1345.0
217.00	14854.0
218.00	2206.0
219.80	470.0
220.90	2908.0
222.10	1163.0
223.10	2757.0
224.00	28424.0
225.00	5115.0
226.00	527.0
227.00	13341.0
228.00	3112.0
229.00	2090.0
231.00	1211.0
234.10	1050.0
234.90	1289.0
236.00	848.0
236.80	929.0
238.80	798.0
239.90	990.0
240.70	501.0
242.10	2068.0
243.00	1179.0
244.00	16760.0
245.00	2327.0
246.00	3287.0
246.80	451.0
252.10	433.0
253.00	693.0
253.80	817.0
255.00	91360.0
256.00	14788.0
257.00	2221.0
258.00	9710.0
259.00	1247.0
260.10	332.0
265.00	3332.0
272.00	351.0
273.00	2367.0

Instrument :
BNA_M
ClientSampleId :
DFTPP63

274.00	7558.0
275.00	31800.0
276.00	4770.0
277.00	4093.0
277.90	907.0
281.20	332.0
282.80	566.0
284.00	620.0
284.90	785.0
289.00	360.0
292.10	471.0
293.00	481.0
295.20	551.0
296.00	17984.0
296.90	1531.0
303.00	1226.0
313.90	1011.0
314.80	1631.0
316.00	1252.0
320.90	844.0
323.00	2749.0
324.10	556.0
326.80	428.0
328.00	493.0
332.00	390.0
332.90	304.0
334.00	1770.0
335.00	686.0
341.20	657.0
346.00	771.0
352.10	1067.0
352.80	999.0
354.10	1118.0
365.00	5995.0
365.90	882.0
372.10	836.0
373.10	535.0
390.20	403.0
402.00	409.0
403.00	1545.0
421.10	473.0
422.10	779.0
423.10	3053.0
424.10	845.0
432.70	566.0
441.00	13329.0
442.00	65992.0
443.00	14942.0
444.00	959.0

Instrument :
BNA_M
ClientSampleId :
DFTPP63

m/z	Abundance
36.80	347.0
38.00	1309.0
39.00	6393.0
40.10	727.0
41.00	861.0
43.00	393.0
44.90	396.0
48.90	682.0
50.10	24520.0
51.00	84392.0
52.00	4300.0
55.10	403.0
56.00	2495.0
57.00	5187.0
58.90	370.0
60.90	980.0
61.80	931.0
62.90	3114.0
63.90	352.0
65.10	1516.0
67.90	1559.0
69.00	76656.0
70.00	514.0
73.10	1899.0
74.10	9893.0
75.00	15421.0
76.00	4746.0
77.10	91632.0
78.00	5452.0
79.00	5959.0
80.00	5214.0
81.10	6830.0
82.00	1183.0
82.90	1449.0
83.80	377.0
84.90	1018.0
86.00	2951.0
86.90	772.0
90.90	1152.0
92.00	1913.0
93.00	15313.0
94.00	990.0
96.00	329.0
98.00	12943.0
99.00	5927.0
100.00	640.0
100.90	2289.0
102.90	1431.0
103.90	2487.0
105.00	2621.0
105.80	897.0
107.00	24400.0
108.00	3690.0
109.00	991.0
110.00	34880.0
110.90	6041.0
111.80	650.0
112.30	709.0
116.10	1831.0
117.00	44096.0
117.90	2770.0
119.80	385.0
121.90	1740.0
123.00	2676.0
123.90	1411.0
124.90	974.0
127.00	57888.0
128.00	4710.0
129.00	30776.0
130.00	2425.0
131.20	407.0
133.10	552.0
134.00	1557.0
135.10	3062.0
136.10	894.0
136.80	624.0
137.20	765.0
140.00	711.0
140.90	6106.0
142.00	1505.0
142.80	969.0
144.00	371.0
145.90	1087.0
147.10	3021.0
147.90	9978.0
148.90	1634.0
150.10	371.0
151.00	409.0
152.90	1308.0
153.80	1308.0
155.10	2809.0

Instrument :
BNA_M
ClientSampleId :
DFTPP63

156.00	3974.0
157.10	945.0
158.00	933.0
159.10	486.0
159.90	1557.0
161.00	2656.0
162.00	847.0
164.10	368.0
164.90	1156.0
166.00	1395.0
167.00	14586.0
168.00	6576.0
169.10	709.0
170.90	498.0
171.70	639.0
173.10	1737.0
174.10	2098.0
175.00	4317.0
176.10	1107.0
176.90	2671.0
177.70	798.0
179.00	6938.0
180.00	5473.0
181.00	2597.0
182.90	313.0
184.10	567.0
185.00	2996.0
186.00	23928.0
186.90	7765.0
187.90	1111.0
189.00	2134.0
190.20	408.0
191.00	1053.0
191.90	3086.0
193.20	2949.0
193.90	716.0
195.00	647.0
196.10	3527.0
197.10	2290.0
197.90	123744.0
198.90	7938.0
200.10	692.0
203.00	1655.0
204.00	8287.0
205.00	13149.0
206.00	44608.0
207.00	7169.0
207.90	1957.0
209.00	1330.0
209.90	1125.0
210.90	2549.0
213.00	362.0
214.90	918.0
215.90	1626.0
216.90	14478.0
217.90	2142.0
221.00	2358.0
221.90	2866.0
223.10	3501.0
224.00	28280.0
225.00	7051.0
225.80	757.0
227.00	16912.0
228.00	2420.0
229.00	2901.0
229.90	362.0
231.00	1006.0
232.80	408.0
233.90	1218.0
235.10	1265.0
235.70	396.0
236.10	304.0
237.00	859.0
241.10	930.0
242.00	1402.0
243.10	1538.0
244.00	18952.0
245.00	2979.0
246.00	6395.0
247.10	881.0
249.00	369.0
253.10	719.0
254.00	1017.0
255.00	106296.0
256.00	14541.0
256.90	806.0
258.00	8094.0
259.00	1396.0
260.10	480.0
264.90	4258.0
266.10	856.0
271.90	492.0
273.10	3029.0
274.00	10178.0
275.00	37024.0

Instrument :

BNA_M

ClientSampleId :

DFTPP63

276.00	4762.0
277.00	6553.0
278.00	977.0
278.80	486.0
281.00	364.0
282.00	539.0
282.90	300.0
284.10	310.0
284.90	1019.0
289.00	648.0
290.00	366.0
292.20	568.0
293.00	850.0
294.10	466.0
296.00	18968.0
296.90	2373.0
302.10	525.0
303.00	1606.0
304.00	791.0
308.00	487.0
313.80	957.0
315.00	2149.0
316.00	1467.0
321.20	636.0
323.10	3633.0
323.90	882.0
327.00	778.0
327.80	637.0
331.90	379.0
333.00	973.0
334.10	3416.0
335.00	470.0
341.10	553.0
342.10	550.0
345.90	870.0
352.10	1038.0
353.10	578.0
354.00	1011.0
365.00	7374.0
366.00	1432.0
371.10	313.0
372.10	1951.0
373.20	370.0
382.90	623.0
402.20	408.0
403.00	956.0
420.90	955.0
422.00	454.0
423.00	5487.0
423.90	1080.0
437.80	333.0
441.00	15276.0
442.00	88072.0
443.00	17928.0
444.10	1345.0

Instrument :
BNA_M
ClientSampleId :
DFTPP63

m/z	Abundance
37.90	721.0
39.00	3610.0
39.90	323.0
41.00	408.0
43.10	324.0
44.00	487.0
50.10	13611.0
51.00	41584.0
52.00	2096.0
55.20	335.0
55.90	973.0
57.10	2221.0
60.80	380.0
62.10	672.0
63.10	1599.0
64.90	447.0
68.00	566.0
69.00	36480.0
73.00	1916.0
73.90	3758.0
75.10	6117.0
76.10	2458.0
77.00	44600.0
77.90	2397.0
79.10	3539.0
80.10	2726.0
80.90	2832.0
82.10	845.0
82.90	555.0
84.90	382.0
86.00	1388.0
86.90	671.0
90.90	1064.0
92.00	1483.0
93.00	6473.0
93.90	313.0
98.00	6884.0
99.00	4822.0
100.90	1770.0
102.00	336.0
102.70	444.0
103.90	1217.0
105.00	1623.0
105.90	556.0
107.00	14334.0
108.00	1827.0
109.10	340.0
110.00	18776.0
111.00	2919.0
112.20	555.0
114.80	408.0
116.10	1453.0
117.00	22744.0
118.00	1060.0
119.90	319.0
121.70	1196.0
123.00	1596.0
123.80	1057.0
125.00	526.0
127.00	28152.0
127.90	2362.0
129.00	15939.0
129.90	1168.0
130.90	509.0
134.00	833.0
135.00	1396.0
135.90	674.0
137.00	1151.0
137.90	345.0
141.00	2686.0
141.70	387.0
142.90	951.0
145.80	416.0
147.00	2018.0
147.90	3654.0
148.80	879.0
153.00	440.0
154.00	858.0
154.90	1241.0
156.00	1606.0
157.00	518.0
157.80	521.0
159.80	664.0
161.00	1234.0
161.80	551.0
164.90	1661.0
166.10	1180.0
167.00	5551.0
168.00	4813.0
170.90	316.0
171.90	468.0

Instrument :
BNA_M
ClientSampleId :
DFTPP63

173.00	796.0
174.00	847.0
175.10	2084.0
176.00	536.0
177.00	767.0
177.90	650.0
179.00	4837.0
180.00	3395.0
180.80	1158.0
183.80	511.0
184.90	1576.0
186.00	13202.0
187.00	2326.0
188.00	317.0
189.00	968.0
190.90	701.0
192.00	1008.0
193.00	2069.0
194.00	574.0
196.00	2339.0
197.00	517.0
197.90	65616.0
198.90	4480.0
200.00	336.0
202.70	1018.0
203.90	4537.0
205.00	8011.0
206.00	24592.0
207.00	4900.0
208.00	2318.0
209.10	833.0
211.10	1253.0
215.80	1152.0
216.90	7781.0
217.90	1902.0
219.90	570.0
221.20	1816.0
222.10	1728.0
222.90	1851.0
224.10	13861.0
225.00	4320.0
225.80	619.0
227.00	8303.0
227.90	1814.0
229.00	2065.0
231.10	393.0
233.90	606.0
234.90	663.0
235.80	1069.0
237.00	1242.0
240.80	619.0
242.10	1000.0
242.90	1245.0
244.00	11326.0
245.00	2094.0
245.90	4661.0
247.00	488.0
249.00	722.0
252.90	670.0
253.80	598.0
254.10	513.0
255.00	56952.0
255.90	8639.0
257.00	750.0
257.90	5618.0
258.90	980.0
263.70	315.0
264.90	3424.0
265.70	343.0
272.00	471.0
273.00	2060.0
274.00	5023.0
275.00	24376.0
276.00	2371.0
277.00	3800.0
277.90	705.0
283.00	672.0
284.90	705.0
292.90	647.0
294.00	318.0
296.00	10046.0
296.80	1270.0
303.00	891.0
304.10	634.0
313.90	805.0
314.90	1486.0
316.10	943.0
321.10	788.0
323.10	2289.0
324.00	470.0
326.80	408.0
327.60	371.0
334.10	1834.0
335.00	470.0
341.00	693.0

Instrument :
BNA_M
ClientSampleId :
DFTPP63

346.10	748.0
346.90	345.0
351.70	751.0
352.20	362.0
352.80	421.0
354.00	1486.0
365.00	4647.0
365.80	607.0
371.90	1644.0
382.90	433.0
402.10	1122.0
403.00	1171.0
403.90	525.0
420.90	1272.0
422.00	792.0
423.00	4879.0
424.20	818.0
441.00	10152.0
442.00	71288.0
443.00	12263.0
444.00	1225.0

Instrument :
BNA_M
ClientSampleId :
DFTPP63

Scan 2566 (17.680 min): BM008952.D
DFTPP63

m/z	Abundance
41.10	341.0
73.00	1497.0
133.10	524.0
193.30	316.0
207.00	1090.0
254.80	481.0
281.10	698.0

Instrument :
BNA_M
ClientSampleId :
DFTPP63