


Data Path : Z:\HPCHEM1\BNA_M\Data\BM020315\
Data File : BM000408.D
Acq On : 03 Feb 2015 08:45
Operator : TP/IZ
Sample : DFTPP15
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP15

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM020315.M
Title : SVOA CALIBRATION
Last Update : Tue Feb 03 12:57:43 2015



AutoFind: Scans 2532, 2533, 2534; Background Corrected with Scan 2526

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.6	143450	PASS
68	69	0.00	2	1.7	2026	PASS
69	198	0.00	100	37.2	122525	PASS
70	69	0.00	2	0.6	727	PASS
127	198	10	80	47.9	157824	PASS
197	198	0.00	2	0.7	2293	PASS
198	198	100	100	100.0	329301	PASS
199	198	5	9	6.7	21973	PASS
275	198	10	60	25.3	83376	PASS
365	198	1	100	4.0	13093	PASS
441	443	0.01	100	77.5	43162	PASS
442	198	50	100	87.7	288810	PASS
443	442	15	24	19.3	55725	PASS

m/z	Abundance
36.00	939.0
37.00	497.0
38.00	1711.0
39.10	10026.0
40.00	303808.0
43.00	306.0
44.00	5132.0
49.00	669.0
50.10	34232.0
51.10	138496.0
52.10	7034.0
53.00	672.0
55.00	991.0
56.00	3683.0
57.10	8017.0
57.90	344.0
58.90	347.0
61.00	1178.0
62.10	1914.0
63.10	5092.0
64.00	683.0
65.10	2245.0
66.10	477.0
68.00	1778.0
69.00	110000.0
70.00	715.0
73.00	2055.0
74.10	13084.0
75.10	18576.0
76.10	6467.0
77.10	118240.0
78.00	7716.0
79.00	9923.0
80.00	7140.0
81.00	12076.0
81.90	2451.0
83.00	2058.0
84.00	623.0
85.00	1678.0
86.00	3522.0
87.00	2103.0
88.00	441.0
91.00	3156.0
92.10	3634.0
93.00	19480.0
94.10	1141.0
96.00	740.0
97.10	553.0
98.00	13835.0
99.00	11907.0
99.90	1135.0
101.00	5164.0
102.20	438.0
103.10	1996.0
104.00	3884.0
105.00	3299.0
106.20	1295.0
107.00	46912.0
108.00	5998.0
109.10	1080.0
110.00	70864.0
111.00	10928.0
112.00	1512.0
113.00	555.0
116.00	2587.0
117.00	41272.0
118.00	3483.0
118.90	613.0
119.90	763.0
122.00	2608.0
123.00	5515.0
124.00	1747.0
124.90	1528.0
125.90	668.0
127.00	132096.0
128.00	11471.0
129.00	60104.0
129.90	4491.0
131.10	1227.0
132.00	818.0
133.00	346.0
134.10	2110.0
135.00	5672.0
136.00	2349.0
137.00	3098.0
138.10	423.0
138.90	389.0
140.00	878.0
141.00	9097.0
142.10	2897.0
143.00	2312.0

Instrument :
BNA_M
ClientSampleId :
DFTPP15

144.00	957.0
145.00	512.0
146.10	1672.0
147.00	4062.0
148.00	10095.0
148.90	1959.0
149.80	427.0
150.90	958.0
151.80	719.0
152.90	1855.0
154.00	1343.0
155.00	4052.0
156.00	5666.0
157.00	1271.0
158.00	1465.0
159.00	985.0
160.00	2647.0
161.00	3619.0
162.20	1115.0
162.90	496.0
164.10	448.0
165.00	2822.0
166.00	2834.0
167.00	14175.0
168.00	8207.0
168.90	1854.0
169.90	545.0
171.20	710.0
172.10	1124.0
173.00	1765.0
173.90	3027.0
175.00	6405.0
176.00	918.0
177.10	2214.0
177.90	847.0
179.00	12719.0
180.00	7988.0
181.00	3106.0
182.00	1004.0
182.70	330.0
184.10	1066.0
185.10	5196.0
186.00	40512.0
187.10	11815.0
188.20	890.0
188.90	2211.0
190.10	485.0
191.10	1680.0
192.00	3339.0
192.90	3299.0
194.00	1122.0
194.90	673.0
196.10	5920.0
196.70	1380.0
197.10	1671.0
198.00	264192.0
199.00	18440.0
200.00	1979.0
201.60	583.0
202.10	722.0
203.00	2146.0
204.10	11035.0
205.00	17576.0
206.10	72016.0
207.00	12678.0
208.00	3552.0
209.10	2031.0
210.00	1064.0
211.10	3368.0
215.00	1263.0
216.00	1982.0
217.00	19944.0
218.10	3047.0
218.90	581.0
219.80	384.0
221.10	5275.0
221.90	2488.0
223.10	4215.0
224.00	41184.0
225.10	9220.0
226.00	1019.0
227.00	18488.0
228.10	3258.0
229.10	4001.0
229.90	789.0
231.10	1399.0
234.00	1468.0
235.00	1262.0
236.20	715.0
237.00	3762.0
238.00	323.0
240.10	910.0
241.00	682.0
242.10	1982.0
243.00	1791.0

Instrument :
BNA_M
ClientSampleId :
DFTPP15

244.10	24368.0
245.10	4111.0
246.00	6778.0
247.00	1461.0
249.00	1281.0
252.00	618.0
253.00	857.0
254.10	1394.0
255.00	154624.0
256.00	22024.0
257.00	1972.0
258.00	10647.0
259.00	1117.0
260.00	453.0
265.00	4080.0
266.00	611.0
270.10	555.0
271.00	643.0
271.90	809.0
273.10	4471.0
274.00	10801.0
275.00	60568.0
276.10	8472.0
277.00	5314.0
277.90	1056.0
279.00	344.0
281.00	832.0
281.90	495.0
283.00	1037.0
284.00	372.0
285.00	1026.0
288.90	341.0
293.00	1753.0
294.10	304.0
294.90	414.0
295.20	367.0
296.00	19144.0
297.10	2887.0
303.00	1633.0
304.00	923.0
307.90	334.0
309.90	640.0
313.90	814.0
315.00	2485.0
316.00	920.0
321.00	834.0
321.80	367.0
323.10	5622.0
324.00	1203.0
326.80	846.0
327.10	777.0
332.00	327.0
333.10	795.0
334.10	3495.0
335.10	652.0
340.90	621.0
346.00	1808.0
346.90	390.0
351.90	1810.0
353.00	1656.0
354.00	1723.0
354.90	420.0
365.00	9163.0
366.00	1352.0
371.00	515.0
372.00	2891.0
372.80	556.0
383.00	1311.0
402.00	1066.0
403.00	1166.0
404.00	804.0
420.90	1487.0
421.90	1638.0
423.10	8353.0
424.10	3720.0
425.00	1165.0
441.00	28040.0
442.00	171840.0
443.00	32064.0
444.00	3017.0

Instrument :
BNA_M
ClientSampleId :
DFTPP15

m/z	Abundance
35.90	1054.0
37.00	747.0
38.10	2412.0
39.10	11832.0
40.00	291328.0
41.10	373.0
43.10	382.0
44.00	4848.0
49.10	1203.0
50.10	42848.0
51.10	168064.0
52.10	8695.0
55.10	953.0
56.00	5064.0
57.00	11186.0
58.00	397.0
61.00	2709.0
62.00	2183.0
63.00	6457.0
64.10	824.0
65.10	3095.0
66.00	348.0
68.10	2406.0
69.00	150656.0
70.00	803.0
73.00	1960.0
74.00	17248.0
75.00	24032.0
76.00	7662.0
77.10	158912.0
78.10	10637.0
79.00	12971.0
80.00	9348.0
81.00	13640.0
82.00	2898.0
83.00	2463.0
84.00	324.0
85.00	1847.0
86.00	3652.0
87.00	2250.0
88.20	955.0
88.80	317.0
91.00	3393.0
92.10	3986.0
93.00	23632.0
94.10	2031.0
95.00	790.0
96.10	1199.0
97.10	564.0
98.00	18760.0
99.00	14439.0
100.20	870.0
101.00	7336.0
102.00	455.0
103.00	2548.0
104.00	5002.0
105.00	4592.0
106.10	2385.0
107.00	63160.0
108.10	8052.0
109.10	1799.0
110.00	96496.0
111.00	15887.0
112.00	2141.0
113.10	691.0
115.00	402.0
115.90	3698.0
117.00	58376.0
118.00	4004.0
119.00	654.0
119.90	995.0
121.10	498.0
122.00	5212.0
123.00	7573.0
124.00	2645.0
124.90	2911.0
126.20	1184.0
127.00	191424.0
128.00	14912.0
129.00	84080.0
130.00	8688.0
131.00	1926.0
132.10	521.0
133.10	806.0
134.00	2555.0
135.00	7462.0
136.00	2551.0
137.00	2767.0
138.10	1111.0
138.90	554.0
140.10	1133.0

Instrument :
BNA_M
ClientSampleId :
DFTPP15

141.00	12498.0
142.00	3265.0
143.10	2950.0
144.00	1344.0
145.00	465.0
146.10	1951.0
147.00	5475.0
148.00	14771.0
149.00	3201.0
150.10	640.0
151.20	1705.0
151.90	1162.0
153.00	3910.0
154.00	2621.0
155.00	7484.0
156.10	8878.0
157.10	1023.0
158.00	2234.0
159.00	1382.0
160.00	3760.0
161.00	4829.0
161.90	1602.0
163.00	845.0
164.10	372.0
165.00	4103.0
166.00	4038.0
167.00	21928.0
168.00	12097.0
169.00	2834.0
170.00	1176.0
171.10	1317.0
172.10	2123.0
173.10	2342.0
174.10	3788.0
175.00	9256.0
176.00	1577.0
177.10	3665.0
178.10	1315.0
179.00	16227.0
180.00	13751.0
181.00	5432.0
182.10	1415.0
182.70	503.0
184.00	1633.0
185.10	7713.0
186.10	63712.0
187.00	20368.0
188.00	2191.0
189.00	5421.0
190.10	588.0
191.10	3413.0
192.10	6146.0
193.10	6022.0
194.10	1199.0
195.00	713.0
196.00	8145.0
197.10	2738.0
198.00	398272.0
199.00	26752.0
200.00	2121.0
201.60	1517.0
203.00	3441.0
204.00	15278.0
205.10	25608.0
206.10	112808.0
207.10	17136.0
208.00	5097.0
209.00	1702.0
210.20	1662.0
211.00	3632.0
211.80	828.0
213.90	441.0
215.00	1844.0
216.10	2946.0
217.00	33288.0
218.00	4275.0
219.00	764.0
219.80	619.0
221.10	8550.0
222.10	3239.0
223.00	8387.0
224.10	56936.0
225.10	15243.0
226.00	2823.0
227.00	31928.0
228.10	4569.0
229.00	5843.0
230.00	840.0
231.10	1474.0
231.90	611.0
232.90	515.0
234.10	1379.0
235.00	1824.0
236.00	1760.0
237.10	2563.0

Instrument :
BNA_M
ClientSampleId :
DFTPP15

238.20	382.0
238.90	1661.0
240.00	772.0
241.00	2203.0
242.10	3915.0
243.10	3964.0
244.10	43584.0
245.10	6627.0
246.00	10237.0
247.00	1903.0
248.00	416.0
249.00	1350.0
251.20	481.0
252.00	544.0
252.90	962.0
254.10	2375.0
255.00	247488.0
256.10	34448.0
257.10	3747.0
258.00	12447.0
258.90	2805.0
260.00	592.0
264.00	336.0
265.00	6228.0
265.90	847.0
267.20	316.0
269.80	559.0
270.90	322.0
271.90	662.0
273.00	6083.0
274.00	17016.0
275.10	99160.0
276.10	12344.0
277.00	9386.0
278.00	1558.0
279.00	382.0
281.20	564.0
282.90	1109.0
283.90	1169.0
285.10	2294.0
289.10	305.0
290.00	320.0
291.10	334.0
291.90	1061.0
293.10	1960.0
294.10	426.0
295.00	896.0
296.00	32048.0
297.00	5266.0
298.00	491.0
301.00	709.0
302.00	963.0
303.10	3799.0
304.00	1341.0
308.10	531.0
308.70	367.0
309.90	300.0
312.90	411.0
314.10	1069.0
315.00	3200.0
316.00	1840.0
321.10	736.0
322.00	609.0
323.00	9670.0
324.10	1707.0
324.90	342.0
327.00	1554.0
328.00	870.0
331.80	544.0
333.10	1014.0
334.10	7164.0
335.00	1850.0
341.00	989.0
342.00	466.0
346.00	2649.0
346.70	384.0
347.10	355.0
352.00	4277.0
353.00	1597.0
354.00	3616.0
354.90	636.0
365.00	15897.0
366.00	1975.0
370.90	593.0
372.00	6263.0
373.00	1159.0
383.00	1607.0
384.10	785.0
385.00	306.0
389.80	732.0
391.10	737.0
402.00	2958.0
403.10	2608.0
404.00	1235.0
405.10	312.0

Instrument :
BNA_M
ClientSampleId :
DFTPP15

421.10	2306.0
422.00	2768.0
423.10	17064.0
424.00	5108.0
424.90	342.0
437.30	347.0
439.10	397.0
441.00	48880.0
442.10	329728.0
443.00	69816.0
444.00	5406.0
445.00	450.0

Instrument :
BNA_M
ClientSampleId :
DFTPP15

m/z	Abundance
36.00	1370.0
37.10	742.0
38.00	1920.0
39.10	8585.0
40.00	300544.0
41.10	356.0
44.00	5401.0
45.10	440.0
48.80	514.0
49.20	707.0
50.10	33304.0
51.10	123792.0
52.10	5340.0
55.10	664.0
56.00	3695.0
57.00	7342.0
58.00	399.0
61.10	1772.0
62.10	2322.0
63.00	4210.0
64.10	844.0
65.10	2210.0
68.00	1896.0
69.00	106920.0
70.00	664.0
73.10	1992.0
74.10	11219.0
75.00	17952.0
76.10	7266.0
77.10	121688.0
78.10	8373.0
79.00	9454.0
80.00	7562.0
81.00	10655.0
82.00	2199.0
83.00	2287.0
85.00	1565.0
86.00	3160.0
87.00	1292.0
87.90	626.0
91.00	2781.0
92.00	2424.0
93.00	17984.0
94.00	981.0
95.00	721.0
96.00	566.0
98.00	14845.0
99.00	10624.0
100.00	962.0
101.00	6588.0
101.90	307.0
103.00	2873.0
104.00	3259.0
105.00	3752.0
106.10	1299.0
107.00	50168.0
108.00	7033.0
109.00	1259.0
110.00	72408.0
111.00	11986.0
112.20	1018.0
113.00	629.0
115.10	327.0
116.10	3332.0
117.00	42024.0
117.90	3318.0
119.10	550.0
120.10	917.0
120.90	304.0
122.00	2855.0
123.00	5116.0
124.00	2555.0
125.10	2146.0
126.20	409.0
127.00	149952.0
128.00	10507.0
129.00	64152.0
130.00	4177.0
131.00	595.0
132.00	463.0
133.00	382.0
133.90	1712.0
135.00	4570.0
136.10	2139.0
136.90	2302.0
138.10	619.0
140.10	1291.0
141.00	10452.0
142.10	3275.0
143.00	2331.0
143.80	435.0

Instrument :
BNA_M
ClientSampleId :
DFTPP15

145.10	433.0
146.00	1670.0
147.00	4350.0
148.10	12047.0
149.10	1606.0
150.00	637.0
151.00	1577.0
151.90	849.0
153.00	3339.0
154.00	1255.0
155.00	4962.0
156.10	7314.0
157.00	1483.0
158.10	1636.0
158.90	1294.0
160.10	3341.0
161.00	4620.0
161.90	1378.0
162.80	527.0
164.00	396.0
165.00	3099.0
166.10	1896.0
167.00	16576.0
168.00	8255.0
168.90	2028.0
170.10	497.0
171.10	592.0
172.10	2397.0
173.10	2973.0
174.00	3992.0
175.00	7590.0
176.10	1312.0
176.90	2901.0
178.10	1543.0
179.00	14562.0
180.00	10220.0
181.00	4501.0
182.10	985.0
182.90	394.0
184.00	1400.0
185.00	6505.0
186.10	49824.0
187.00	16656.0
188.00	1498.0
189.00	3930.0
190.10	440.0
190.90	2058.0
192.00	4524.0
193.00	5226.0
193.90	1151.0
196.00	7880.0
197.00	2471.0
198.00	325440.0
199.00	20728.0
199.90	1968.0
200.80	445.0
201.40	1147.0
201.90	562.0
203.00	2666.0
204.10	13132.0
205.10	21880.0
206.10	95160.0
207.10	13747.0
208.10	4038.0
208.90	1461.0
210.00	1849.0
211.10	3464.0
211.80	462.0
215.10	1195.0
216.00	2244.0
217.00	26952.0
218.10	2934.0
219.00	397.0
220.10	806.0
221.10	6419.0
222.00	3475.0
223.00	6270.0
224.10	51344.0
225.10	13418.0
226.10	1547.0
227.00	28264.0
228.00	4583.0
229.00	5000.0
230.00	735.0
231.10	2225.0
232.00	338.0
233.20	641.0
234.00	1134.0
235.00	1480.0
236.10	1450.0
237.00	2450.0
238.10	489.0
239.10	1072.0
240.00	731.0
241.00	1424.0

Instrument :
BNA_M
ClientSampleId :
DFTPP15

242.00	2701.0
243.10	3025.0
244.10	35280.0
245.10	5899.0
246.00	9755.0
247.00	1827.0
247.80	448.0
248.20	421.0
248.90	2073.0
251.00	640.0
252.10	934.0
253.10	1098.0
254.10	2557.0
255.00	225216.0
256.10	31928.0
257.10	2232.0
258.00	13629.0
259.00	1934.0
260.10	378.0
264.10	531.0
265.00	5478.0
265.80	364.0
269.80	336.0
270.00	341.0
270.80	415.0
272.00	942.0
273.00	6046.0
274.00	17704.0
275.00	90400.0
276.00	11936.0
277.00	8415.0
278.10	1751.0
278.90	333.0
281.10	465.0
282.10	302.0
283.00	803.0
284.20	627.0
285.00	1858.0
289.00	520.0
290.10	305.0
292.90	1951.0
294.00	502.0
295.10	444.0
296.00	30448.0
297.00	3461.0
301.90	556.0
303.00	2657.0
304.00	842.0
307.90	618.0
308.90	445.0
313.90	1799.0
315.00	2948.0
316.10	1657.0
320.90	1350.0
321.90	723.0
323.10	9132.0
324.00	1963.0
326.90	1846.0
328.10	1393.0
332.10	842.0
333.10	1122.0
334.00	6343.0
335.10	1779.0
339.90	473.0
341.00	855.0
342.30	419.0
345.80	1489.0
347.10	478.0
351.10	422.0
352.00	2955.0
353.00	2271.0
354.00	2702.0
354.90	926.0
365.00	14221.0
366.00	2338.0
371.00	662.0
372.10	5835.0
372.90	1753.0
382.90	1234.0
383.90	814.0
385.00	429.0
389.90	808.0
390.90	448.0
392.20	494.0
401.00	412.0
402.00	2230.0
403.10	3105.0
404.10	1470.0
414.70	324.0
421.00	3543.0
421.90	2626.0
423.10	16896.0
424.10	4175.0
424.80	663.0
425.20	605.0

Instrument :
BNA_M
ClientSampleId :
DFTPP15

438.30	308.0
439.60	352.0
441.00	52568.0
442.00	364864.0
443.10	65296.0
444.00	6785.0

Instrument :
BNA_M
ClientSampleId :
DFTPP15

Scan 2526 (17.545 min): BM000408.D
DFTPP15

m/z	Abundance
36.00	1476.0
39.10	365.0
40.00	340608.0
44.00	4966.0
73.10	856.0
115.00	407.0
167.10	955.0
168.10	468.0
193.20	590.0
207.00	2290.0
208.10	857.0
281.20	502.0

Instrument :
BNA_M
ClientSampleId :
DFTPP15