


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM020317\
 Data File : BM008967.D
 Acq On : 03 Feb 2017 15:29
 Operator : SJ/MA
 Sample : DFTPP64
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP64

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM013017.M
 Title : SVOA CALIBRATION
 Last Update : Fri Feb 03 22:25:40 2017



AutoFind: Scans 2571, 2572, 2573; Background Corrected with Scan 2566

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	73.5	170240	PASS
68	69	0.00	2	1.7	2694	PASS
69	198	0.00	100	67.2	155549	PASS
70	69	0.00	2	0.6	884	PASS
127	198	10	80	49.0	113413	PASS
197	198	0.00	2	1.1	2539	PASS
198	198	100	100	100.0	231637	PASS
199	198	5	9	6.9	15953	PASS
275	198	10	60	29.1	67488	PASS
365	198	1	100	5.7	13111	PASS
441	443	0.01	100	83.6	23775	PASS
442	198	50	100	61.2	141802	PASS
443	442	15	24	20.1	28437	PASS

Scan 2566 (17.680 min): BM008967.D
DFTPP64

m/z	Abundance
40.00	418.0
43.90	448.0
44.80	521.0
55.20	449.0
73.10	986.0
202.90	312.0
207.00	1976.0
208.10	366.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64

m/z	Abundance
37.20	906.0
38.00	2298.0
39.10	9784.0
40.10	938.0
41.20	514.0
43.00	351.0
43.90	753.0
47.90	365.0
49.10	985.0
50.00	39912.0
51.00	133312.0
52.10	7436.0
53.00	632.0
55.00	1350.0
56.10	4465.0
57.00	9547.0
58.00	879.0
61.00	1669.0
61.90	2707.0
63.00	5419.0
64.10	1192.0
65.10	1976.0
65.80	740.0
67.00	818.0
68.10	2341.0
69.00	120152.0
69.90	660.0
70.20	640.0
72.90	2263.0
74.00	17216.0
75.00	23496.0
76.00	8135.0
77.10	127376.0
78.00	7250.0
79.00	11383.0
80.00	8752.0
81.00	10905.0
81.90	2563.0
82.90	1425.0
85.00	1751.0
85.90	3785.0
87.10	1592.0
88.00	314.0
88.90	358.0
91.00	2436.0
92.10	1933.0
93.00	20256.0
94.00	1515.0
94.90	549.0
96.00	567.0
96.90	350.0
98.00	17072.0
99.00	10511.0
99.90	881.0
101.00	5716.0
102.80	1818.0
104.00	3810.0
105.00	2720.0
106.00	1017.0
107.00	38368.0
107.90	4632.0
109.00	2202.0
110.00	50280.0
111.00	9831.0
112.10	1472.0
113.00	388.0
115.90	1474.0
117.00	55728.0
118.00	3150.0
119.70	536.0
121.80	2810.0
122.90	4135.0
123.90	1493.0
124.90	1751.0
127.00	81528.0
128.00	7382.0
129.00	43256.0
130.00	3059.0
130.90	944.0
132.10	376.0
133.90	927.0
135.00	3680.0
136.00	1515.0
137.00	1520.0
138.10	575.0
139.80	361.0
141.00	5087.0
142.00	1535.0
143.00	1769.0
145.00	384.0
145.80	1560.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64

147.00	2915.0
148.00	9423.0
149.10	1202.0
151.10	1075.0
152.80	724.0
154.10	829.0
155.00	3753.0
155.90	5292.0
157.00	699.0
158.00	2033.0
158.70	884.0
159.90	1937.0
161.10	2118.0
162.10	572.0
165.10	2320.0
165.90	2604.0
167.00	15657.0
168.00	8962.0
168.80	1047.0
169.90	442.0
171.20	604.0
172.00	1693.0
173.00	1169.0
174.00	3426.0
175.00	4596.0
176.10	988.0
177.10	1766.0
177.90	1198.0
178.90	8964.0
179.90	7268.0
181.00	2993.0
181.70	302.0
184.20	734.0
185.00	5351.0
186.00	28400.0
187.00	9161.0
188.10	1253.0
189.00	2610.0
190.80	2036.0
192.00	2962.0
193.00	4297.0
193.80	847.0
195.00	454.0
196.00	4542.0
196.90	1607.0
197.90	147840.0
199.00	12477.0
199.90	1328.0
201.60	950.0
203.00	1688.0
203.90	8716.0
205.00	13877.0
206.00	54080.0
207.10	8822.0
208.00	4251.0
208.80	549.0
209.10	457.0
210.10	802.0
211.00	2492.0
213.00	363.0
214.80	1620.0
216.10	1602.0
217.00	18984.0
217.90	1830.0
220.10	366.0
221.10	4005.0
222.10	1855.0
223.00	2525.0
224.00	30368.0
225.00	8749.0
225.80	785.0
227.00	17840.0
227.90	2635.0
229.00	3489.0
230.00	495.0
230.90	1326.0
233.90	1798.0
234.90	1487.0
235.80	1009.0
237.00	963.0
238.90	541.0
240.00	457.0
241.00	1386.0
241.90	1824.0
243.00	1639.0
244.00	21032.0
245.10	3117.0
246.00	7497.0
247.10	1313.0
248.90	915.0
253.00	1105.0
254.10	797.0
255.00	111272.0
256.00	16544.0
256.90	1700.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64

257.90	8443.0
258.80	1173.0
259.10	1365.0
263.80	542.0
265.00	3813.0
266.00	447.0
272.00	532.0
272.90	2080.0
274.00	10207.0
275.00	41032.0
276.00	5523.0
276.90	5913.0
277.90	1546.0
280.80	513.0
283.20	607.0
284.90	496.0
291.10	362.0
291.80	318.0
293.00	1196.0
295.10	588.0
296.00	18896.0
297.10	1892.0
301.00	428.0
303.00	2443.0
303.70	419.0
314.00	679.0
315.00	1660.0
315.90	988.0
321.10	447.0
323.00	4807.0
324.20	633.0
327.00	1220.0
334.10	3511.0
335.10	686.0
341.20	413.0
346.00	981.0
351.10	349.0
352.10	1311.0
353.20	389.0
354.10	544.0
354.90	313.0
359.30	354.0
365.00	7096.0
365.90	1318.0
370.90	599.0
372.10	2559.0
383.10	426.0
392.00	302.0
402.00	1420.0
403.00	953.0
421.00	959.0
422.00	615.0
423.00	4202.0
424.10	983.0
441.00	13086.0
442.10	70464.0
443.00	14545.0
444.00	1244.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64

m/z	Abundance
37.00	1297.0
38.10	2444.0
39.00	13764.0
40.00	1563.0
40.80	448.0
43.10	674.0
43.90	733.0
44.90	515.0
47.90	379.0
49.00	1876.0
50.00	55960.0
51.00	198720.0
52.10	11429.0
52.90	822.0
55.00	970.0
56.00	7145.0
57.00	11508.0
57.90	453.0
61.00	2289.0
62.10	3305.0
63.00	8072.0
63.90	1463.0
65.00	3913.0
65.90	400.0
68.10	2829.0
69.00	185664.0
70.00	955.0
73.00	3430.0
74.00	23880.0
75.00	33504.0
76.10	10444.0
77.10	202240.0
78.00	12343.0
79.00	15556.0
80.00	12921.0
81.00	18584.0
82.00	3011.0
83.00	3902.0
84.00	359.0
85.00	2320.0
86.00	5149.0
87.00	3089.0
88.10	814.0
89.10	694.0
91.10	3373.0
92.00	4805.0
93.00	32576.0
94.00	2341.0
96.00	1014.0
96.90	1136.0
98.00	27840.0
99.00	16960.0
100.00	2077.0
101.00	6821.0
102.10	481.0
102.80	1558.0
104.00	6103.0
105.00	5215.0
106.10	1675.0
107.00	65024.0
108.00	8832.0
109.00	1800.0
110.00	82704.0
111.00	15325.0
112.10	1754.0
113.10	1379.0
114.80	610.0
115.90	4649.0
117.00	86120.0
118.00	4692.0
118.80	628.0
119.90	785.0
120.90	702.0
122.00	4801.0
122.90	6739.0
124.00	3388.0
125.00	1480.0
125.80	616.0
127.00	134720.0
128.00	9788.0
129.00	73880.0
130.00	5039.0
131.00	1930.0
132.10	605.0
133.90	2683.0
135.00	5059.0
136.00	2630.0
136.90	2981.0
138.00	699.0
139.90	1751.0
141.00	9548.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64

142.00	3169.0
143.10	2662.0
144.00	707.0
144.90	883.0
146.00	2192.0
147.00	4812.0
148.00	19160.0
149.00	3896.0
150.10	764.0
151.00	1162.0
151.90	928.0
153.00	3214.0
154.00	2468.0
154.90	5990.0
156.00	6998.0
157.10	795.0
158.00	2256.0
158.90	2080.0
160.00	4053.0
160.90	5272.0
162.10	1267.0
162.80	519.0
163.20	500.0
163.90	536.0
165.00	3328.0
166.00	2301.0
167.00	25768.0
168.00	14812.0
169.10	1955.0
169.80	1131.0
171.10	1252.0
172.00	2360.0
172.80	2685.0
174.10	4276.0
175.00	7099.0
176.00	1458.0
176.90	3723.0
178.00	1627.0
178.90	17240.0
180.00	8613.0
181.00	4271.0
182.20	716.0
183.00	580.0
184.00	1604.0
185.00	8863.0
186.00	50128.0
187.00	14322.0
188.00	2301.0
189.00	3841.0
190.00	914.0
190.80	2448.0
192.00	6168.0
193.00	7134.0
193.90	1087.0
194.80	1229.0
195.90	7267.0
196.90	2440.0
197.90	276928.0
199.00	18048.0
199.80	1537.0
201.10	691.0
202.10	636.0
203.00	3280.0
204.00	12943.0
205.00	25560.0
206.00	89992.0
207.10	15475.0
208.00	5794.0
209.00	1669.0
209.90	2138.0
211.00	4405.0
214.00	384.0
215.00	2358.0
216.00	2124.0
216.90	28520.0
218.00	3405.0
218.90	480.0
220.90	5948.0
222.00	3622.0
223.00	5523.0
224.00	51720.0
225.00	13227.0
226.10	2000.0
227.00	32192.0
228.00	5420.0
229.00	5582.0
229.80	550.0
231.00	2136.0
232.00	570.0
234.00	2096.0
234.90	1759.0
235.90	1882.0
237.10	2226.0
237.80	396.0
239.10	1003.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64

239.80	1036.0
241.00	2221.0
242.00	3033.0
243.00	2568.0
244.00	34968.0
245.00	5570.0
246.00	10428.0
246.90	2266.0
248.00	350.0
249.00	1204.0
250.80	309.0
252.00	558.0
252.90	1179.0
254.10	2163.0
255.00	196096.0
256.00	29680.0
257.00	2098.0
258.00	16496.0
258.90	3229.0
260.00	382.0
264.10	603.0
265.00	6679.0
265.90	1088.0
270.90	460.0
272.20	885.0
273.00	6301.0
274.00	15962.0
275.00	80536.0
276.00	11870.0
277.00	10096.0
277.90	1785.0
281.10	518.0
282.10	446.0
283.10	1333.0
283.90	516.0
285.00	1935.0
285.90	338.0
291.00	321.0
293.00	2233.0
294.20	826.0
294.90	664.0
296.00	34616.0
296.90	5396.0
300.90	474.0
302.00	833.0
303.10	3819.0
304.10	1522.0
304.90	325.0
308.90	752.0
310.00	341.0
314.00	1670.0
314.90	4018.0
316.00	1417.0
321.10	698.0
321.90	394.0
323.10	5808.0
324.10	1195.0
326.90	2146.0
328.00	854.0
332.10	630.0
333.00	726.0
334.00	4816.0
335.00	1626.0
339.10	333.0
341.00	1513.0
345.90	2452.0
347.10	620.0
352.00	2303.0
353.10	2203.0
353.90	1736.0
364.00	321.0
365.00	14655.0
366.00	2164.0
369.90	308.0
370.90	407.0
372.10	2672.0
373.00	952.0
382.70	620.0
383.80	708.0
390.00	586.0
392.00	396.0
401.90	1794.0
403.00	2505.0
404.10	595.0
421.00	1476.0
421.80	1181.0
423.00	7198.0
424.10	1829.0
425.00	334.0
441.00	25864.0
442.00	153408.0
443.10	30056.0
444.00	2873.0
444.70	631.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64

m/z	Abundance
37.10	898.0
38.10	2089.0
39.00	13898.0
40.00	1365.0
41.00	726.0
43.10	323.0
43.90	523.0
45.10	406.0
48.90	1618.0
50.00	51768.0
51.10	178688.0
52.10	8221.0
52.90	353.0
55.00	849.0
55.90	5567.0
57.00	12823.0
60.00	501.0
61.00	3003.0
62.00	2287.0
63.00	6661.0
63.90	1670.0
65.10	2064.0
65.90	424.0
67.10	436.0
68.00	2912.0
69.00	160832.0
70.00	1037.0
73.00	3554.0
74.10	21552.0
75.00	28336.0
76.10	11156.0
77.10	177280.0
78.10	10863.0
79.00	13115.0
80.00	10299.0
81.00	15052.0
82.00	3349.0
83.00	3546.0
85.00	3082.0
86.00	5920.0
87.00	2134.0
88.00	706.0
88.80	357.0
91.10	2876.0
92.00	3506.0
93.00	30800.0
94.00	1562.0
95.00	622.0
96.10	1140.0
97.00	695.0
98.00	22680.0
99.00	14305.0
100.00	1170.0
101.00	6827.0
101.80	529.0
103.00	2307.0
104.00	6141.0
105.10	4465.0
106.00	1919.0
107.00	59352.0
108.00	6190.0
109.00	1482.0
110.00	78920.0
111.00	12388.0
112.00	1775.0
113.10	666.0
114.80	362.0
115.20	339.0
115.90	3602.0
117.00	82360.0
118.00	5961.0
119.10	922.0
119.90	853.0
120.90	750.0
121.90	4332.0
123.00	6090.0
124.00	3957.0
124.90	2358.0
125.90	702.0
127.00	123992.0
128.00	11007.0
129.00	61808.0
129.90	5813.0
131.00	2134.0
132.20	597.0
132.90	518.0
134.00	1935.0
135.00	6222.0
136.10	2458.0
137.00	3044.0
137.90	630.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64

139.00	428.0
140.00	846.0
141.00	9945.0
142.00	2655.0
143.10	2045.0
144.00	875.0
145.00	1337.0
146.10	2441.0
147.00	6226.0
148.00	20064.0
149.00	4011.0
150.00	1148.0
150.90	1477.0
151.90	797.0
153.00	3969.0
153.90	2416.0
154.90	6574.0
156.00	8801.0
156.80	346.0
157.20	758.0
157.90	2778.0
158.80	1376.0
159.90	2920.0
160.90	4768.0
161.90	2036.0
162.90	650.0
163.90	414.0
165.10	4378.0
166.00	3584.0
167.00	29584.0
167.90	13920.0
169.00	2514.0
170.00	596.0
171.00	542.0
172.10	2202.0
173.00	2395.0
173.90	3965.0
175.00	8264.0
176.00	1952.0
176.80	4307.0
177.90	2281.0
179.00	15911.0
180.00	10809.0
181.00	5220.0
182.30	463.0
182.90	485.0
183.80	1337.0
185.00	8464.0
186.00	49448.0
187.00	16117.0
187.90	1724.0
189.00	4645.0
189.90	549.0
190.90	2870.0
192.10	5420.0
193.00	5016.0
193.90	1581.0
195.00	527.0
196.00	8881.0
197.00	3571.0
198.00	270144.0
199.00	17336.0
200.00	2559.0
201.70	658.0
203.00	4505.0
204.00	15593.0
205.00	23008.0
206.00	92784.0
207.10	12748.0
208.00	4896.0
208.90	2480.0
210.00	2022.0
210.90	4208.0
211.90	347.0
213.00	464.0
215.10	1296.0
216.10	3704.0
217.00	34480.0
218.00	4363.0
218.90	319.0
219.90	494.0
221.00	7016.0
222.00	3962.0
223.00	7188.0
224.00	61360.0
225.00	14614.0
226.00	1955.0
227.00	35816.0
227.90	4763.0
229.00	5120.0
229.90	655.0
231.10	1222.0
232.20	632.0
233.00	582.0
234.00	2745.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64

235.20	1862.0
236.00	2031.0
236.90	2813.0
237.90	376.0
238.90	1403.0
240.10	742.0
240.90	1446.0
242.00	2542.0
243.10	3369.0
244.00	42832.0
245.00	5183.0
246.00	13214.0
247.00	2050.0
248.20	340.0
249.00	981.0
250.00	444.0
250.90	844.0
252.00	393.0
253.00	1473.0
254.10	2808.0
255.00	226752.0
256.00	30800.0
257.10	3374.0
258.00	19160.0
259.00	2587.0
261.00	450.0
264.10	329.0
264.90	7443.0
266.00	1464.0
271.10	594.0
272.00	854.0
273.00	8112.0
274.00	18904.0
275.00	80896.0
276.00	10177.0
277.00	10359.0
277.90	1578.0
278.80	307.0
281.00	579.0
282.00	523.0
282.90	710.0
284.00	765.0
285.00	1923.0
288.90	461.0
290.10	302.0
290.90	755.0
291.90	522.0
292.80	1047.0
294.10	1015.0
295.00	928.0
296.00	38032.0
297.00	4586.0
298.00	755.0
301.00	565.0
302.00	435.0
303.00	3710.0
304.00	1049.0
307.90	425.0
309.00	381.0
309.80	541.0
314.10	1599.0
315.00	4687.0
316.00	1462.0
317.00	587.0
321.00	883.0
321.70	585.0
322.20	953.0
323.00	9241.0
324.00	1727.0
326.90	2822.0
328.00	1297.0
332.00	998.0
333.10	1075.0
334.00	8341.0
335.10	1776.0
340.90	1055.0
346.00	2115.0
346.90	985.0
352.10	2703.0
352.90	1611.0
354.10	2404.0
355.10	393.0
365.00	17584.0
366.00	2389.0
370.80	591.0
372.00	3131.0
373.00	649.0
383.00	1601.0
390.10	556.0
402.00	1465.0
403.10	1775.0
404.20	922.0
421.00	2074.0
422.00	1929.0
423.00	10095.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64

424.10	2216.0
425.00	413.0
441.00	32376.0
442.00	201536.0
443.10	40712.0
444.00	2825.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64