


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM020322\
 Data File : BM033993.D
 Acq On : 03 Feb 2022 09:50
 Operator : CG/JU
 Sample : DFTPP036
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP036

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM013122.M
 Title : SVOA CALIBRATION
 Last Update : Tue Feb 01 02:16:28 2022



AutoFind: Scans 2465, 2466, 2467; Background Corrected with Scan 2459

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.6	152589	PASS
68	69	0.00	2	2.0	3017	PASS
69	198	0.00	100	37.4	151635	PASS
70	69	0.00	2	0.6	885	PASS
127	198	10	80	46.8	189589	PASS
197	198	0.00	2	0.7	2763	PASS
198	198	100	100	100.0	405355	PASS
199	198	5	9	6.9	27989	PASS
275	198	10	60	28.6	115859	PASS
365	198	1	100	4.2	16992	PASS
441	198	0.01	100	12.9	52235	PASS
442	198	50	100	83.2	337067	PASS
443	442	15	24	19.2	64773	PASS

m/z	Abundance
39.00	155.0
40.00	462.0
41.00	158.0
44.10	865.0
49.00	448.0
73.00	731.0
77.00	242.0
84.00	254.0
96.00	341.0
115.10	161.0
119.00	156.0
133.10	316.0
135.00	161.0
147.10	187.0
177.00	154.0
191.00	314.0
193.00	250.0
207.10	2824.0
208.00	553.0
208.90	314.0
221.20	198.0
266.70	164.0
281.10	572.0
282.00	173.0
282.90	158.0

Instrument :
BNA_M
ClientSampleId :
DFTPP036

m/z	Abundance
35.90	153.0
37.00	873.0
38.00	1703.0
39.10	11190.0
40.00	916.0
41.00	437.0
43.00	210.0
44.10	1010.0
45.00	319.0
49.10	1394.0
50.10	42080.0
51.10	163136.0
52.10	7970.0
53.10	265.0
55.00	995.0
56.00	4768.0
57.00	10800.0
58.00	395.0
58.90	199.0
61.00	1847.0
62.00	2476.0
63.10	6095.0
64.00	803.0
65.10	3205.0
66.00	277.0
66.90	259.0
68.10	3168.0
69.00	156992.0
70.10	809.0
71.00	150.0
73.10	1816.0
74.10	17080.0
75.10	25640.0
76.10	9094.0
77.10	172992.0
78.10	12352.0
79.00	12804.0
80.00	9423.0
81.00	13119.0
82.00	3491.0
83.00	2901.0
84.00	438.0
85.00	2261.0
86.00	4061.0
87.00	2012.0
88.10	710.0
88.90	418.0
91.00	3274.0
92.00	3926.0
93.10	22416.0
94.00	1218.0
95.10	323.0
96.00	1195.0
98.10	17768.0
99.00	13866.0
100.00	1374.0
101.00	6856.0
101.90	370.0
103.00	2440.0
104.00	5583.0
105.00	4696.0
106.00	1409.0
107.10	58288.0
108.00	8278.0
109.00	1764.0
110.00	99528.0
111.00	15490.0
112.00	2001.0
113.00	717.0
114.10	166.0
115.00	421.0
116.10	3448.0
117.00	51328.0
118.00	3599.0
119.00	516.0
120.10	738.0
121.00	297.0
122.00	3662.0
123.00	6045.0
124.00	2910.0
125.00	2345.0
127.10	186816.0
128.10	14317.0
129.00	74960.0
130.00	6473.0
131.00	1432.0
132.10	673.0
133.00	704.0
134.00	2425.0
135.00	6166.0
136.00	2673.0
137.10	3577.0

Instrument :
BNA_M
ClientSampleId :
DFTPP036

138.10	747.0
139.10	370.0
140.00	1150.0
141.00	11407.0
142.10	3311.0
143.00	2119.0
144.00	468.0
145.10	466.0
146.00	1576.0
147.00	5675.0
148.00	12613.0
149.00	2450.0
150.00	736.0
151.00	1358.0
151.80	772.0
153.00	3480.0
154.00	2236.0
155.10	4971.0
156.00	8916.0
157.00	1471.0
158.00	1859.0
159.00	1492.0
160.00	3223.0
161.00	4901.0
162.00	1516.0
163.10	325.0
164.10	464.0
165.00	3739.0
166.00	3560.0
167.00	19136.0
168.00	10544.0
169.00	1313.0
170.10	634.0
171.10	702.0
172.00	1920.0
173.00	2212.0
174.00	4244.0
175.10	8048.0
176.00	1984.0
177.00	3983.0
178.10	1597.0
179.00	15426.0
180.00	11038.0
181.00	4454.0
182.00	757.0
182.90	441.0
184.00	1510.0
185.00	7594.0
186.10	56080.0
187.10	16392.0
188.10	1853.0
189.10	3718.0
190.00	670.0
191.00	2120.0
192.00	4495.0
193.00	5416.0
194.00	1188.0
195.10	511.0
196.00	10234.0
197.10	2594.0
198.00	378944.0
199.00	25824.0
200.10	2224.0
201.50	1316.0
203.00	2772.0
204.10	15472.0
205.10	25176.0
206.10	103232.0
207.10	18000.0
208.00	4370.0
209.00	1657.0
210.10	2128.0
211.10	3977.0
212.00	846.0
213.00	442.0
214.00	263.0
215.00	1170.0
216.00	2145.0
217.00	28320.0
218.00	3899.0
219.00	402.0
221.10	12446.0
221.90	3688.0
223.10	5949.0
224.10	57288.0
225.10	14425.0
226.00	1654.0
227.00	24504.0
228.00	3664.0
229.00	5596.0
230.00	719.0
231.10	1806.0
232.00	439.0
233.00	346.0
234.00	2016.0
235.00	1972.0

Instrument :
BNA_M
ClientSampleId :
DFTPP036

236.00	1357.0
237.10	2476.0
237.90	349.0
239.00	1030.0
240.10	671.0
241.00	1580.0
242.10	2804.0
243.10	3192.0
244.10	40104.0
245.10	5810.0
246.10	9681.0
247.00	2024.0
248.00	378.0
249.00	1583.0
250.10	337.0
251.10	365.0
252.10	308.0
253.00	917.0
254.20	2069.0
255.10	225344.0
256.10	33768.0
257.10	3203.0
258.00	13111.0
259.00	2239.0
260.00	458.0
261.00	245.0
264.00	459.0
265.00	6475.0
266.00	777.0
267.00	227.0
267.90	192.0
269.90	245.0
271.00	298.0
272.10	631.0
273.10	6013.0
274.10	17104.0
275.10	97576.0
276.00	13343.0
277.00	8769.0
278.00	1924.0
279.00	304.0
281.10	633.0
282.00	422.0
283.00	1189.0
284.00	935.0
285.10	1629.0
285.90	343.0
289.00	392.0
290.10	330.0
290.90	213.0
292.10	315.0
293.00	2038.0
294.10	634.0
295.00	854.0
296.00	29512.0
297.00	4248.0
297.90	201.0
301.10	463.0
302.00	496.0
303.10	2911.0
304.10	850.0
308.00	481.0
309.00	253.0
310.10	431.0
313.00	277.0
314.00	1384.0
315.00	3049.0
316.10	1540.0
317.10	261.0
321.00	1013.0
322.10	571.0
323.10	9089.0
324.10	1425.0
325.10	188.0
326.10	154.0
327.00	1832.0
328.00	720.0
329.00	333.0
332.00	578.0
332.90	824.0
334.10	6762.0
335.00	1515.0
336.10	217.0
339.00	178.0
341.10	849.0
341.90	281.0
346.00	2217.0
347.00	545.0
351.00	169.0
352.10	2630.0
353.00	1850.0
354.10	3261.0
355.10	648.0
359.10	152.0
365.00	13663.0
366.00	1713.0

Instrument :
BNA_M
ClientSampleId :
DFTPP036

370.00	236.0
370.90	652.0
372.10	4617.0
373.00	1383.0
374.00	158.0
377.00	168.0
383.00	1038.0
383.80	381.0
384.90	150.0
390.00	456.0
390.90	354.0
392.00	400.0
401.00	284.0
402.00	1862.0
403.00	2425.0
403.90	892.0
405.10	167.0
421.00	2152.0
422.10	2261.0
423.10	13976.0
424.00	4701.0
425.00	565.0
441.10	35624.0
442.10	235520.0
443.10	47632.0
444.10	4340.0
445.10	181.0

Instrument :
BNA_M
ClientSampleId :
DFTPP036

m/z	Abundance
37.00	814.0
38.10	2086.0
39.10	13000.0
40.00	800.0
41.00	390.0
43.00	197.0
44.00	753.0
45.00	413.0
47.90	187.0
49.10	1969.0
50.10	48176.0
51.10	181248.0
52.10	9216.0
53.10	377.0
55.00	1338.0
56.00	5506.0
57.10	12306.0
58.00	539.0
58.90	231.0
61.10	2342.0
62.00	2706.0
63.10	7786.0
64.00	1068.0
65.10	3211.0
65.90	265.0
66.80	162.0
68.10	3447.0
69.00	182080.0
70.10	1090.0
73.10	2169.0
74.10	19608.0
75.10	30144.0
76.10	10399.0
77.10	201536.0
78.10	13235.0
79.00	15162.0
80.10	10561.0
81.00	15906.0
82.00	3601.0
83.00	3152.0
84.00	736.0
85.00	2689.0
86.00	4957.0
87.10	2007.0
88.10	723.0
89.00	496.0
91.00	3239.0
92.00	3863.0
93.00	27048.0
94.00	1749.0
95.00	356.0
96.10	1114.0
98.00	21168.0
99.10	15445.0
100.00	1227.0
101.00	8369.0
102.10	576.0
103.00	3154.0
104.00	5981.0
105.00	5360.0
106.10	1851.0
107.00	69584.0
108.00	11009.0
109.20	1860.0
110.00	120744.0
111.00	19312.0
112.00	2148.0
112.90	789.0
114.20	180.0
114.90	315.0
116.10	4027.0
117.00	59752.0
118.00	4101.0
119.00	715.0
120.10	1025.0
121.10	316.0
122.00	4587.0
123.00	7487.0
124.00	3217.0
125.10	2876.0
127.10	230592.0
128.00	17264.0
129.00	93496.0
130.00	7982.0
131.00	1677.0
132.00	1039.0
133.00	551.0
134.00	2881.0
135.10	7407.0
136.00	3366.0
137.10	3867.0
138.00	858.0

Instrument :
BNA_M
ClientSampleId :
DFTPP036

139.00	580.0
140.00	1245.0
141.00	14359.0
142.00	4583.0
143.00	2952.0
144.00	763.0
145.00	530.0
146.00	2378.0
147.00	6859.0
148.00	14762.0
149.00	3109.0
150.00	732.0
151.10	1380.0
151.90	772.0
153.00	4167.0
154.00	2688.0
155.00	6939.0
156.10	10766.0
157.00	1788.0
158.00	2689.0
159.00	1774.0
160.00	4195.0
161.00	5845.0
162.00	1936.0
163.00	523.0
164.00	636.0
165.00	4919.0
166.10	4040.0
167.00	26048.0
168.00	11992.0
169.00	2204.0
170.00	1084.0
171.00	1018.0
171.90	2451.0
173.00	3065.0
174.10	5354.0
175.10	10093.0
176.10	2847.0
177.00	5096.0
178.00	1397.0
179.00	18512.0
180.10	13290.0
181.00	5757.0
182.10	941.0
183.00	507.0
184.10	1543.0
185.00	9274.0
186.10	71200.0
187.00	19912.0
188.00	2103.0
189.00	5091.0
190.10	832.0
191.00	2454.0
192.10	6603.0
193.10	7399.0
194.00	1452.0
195.00	595.0
196.10	13629.0
196.80	2528.0
197.20	3178.0
198.00	478336.0
199.00	33632.0
200.10	2819.0
201.60	2139.0
203.00	4047.0
204.10	19464.0
205.10	31312.0
206.10	136320.0
207.10	21512.0
208.10	5582.0
209.00	1650.0
210.10	2216.0
211.00	5219.0
211.80	1249.0
213.00	422.0
215.00	1493.0
216.10	2744.0
217.00	34984.0
218.00	4356.0
219.00	526.0
220.10	723.0
221.10	16960.0
222.00	4552.0
223.00	8452.0
224.10	74360.0
225.10	19504.0
226.10	2146.0
227.00	33936.0
228.10	4811.0
229.00	6600.0
230.00	703.0
231.10	2461.0
232.00	341.0
233.10	511.0
234.10	2039.0
235.00	2386.0

Instrument :
BNA_M
ClientSampleId :
DFTPP036

236.00	1706.0
237.10	2986.0
237.90	446.0
239.00	1443.0
240.00	1143.0
241.00	2185.0
242.10	3759.0
243.10	3652.0
244.10	55000.0
245.10	7693.0
246.10	12185.0
247.00	2581.0
248.00	450.0
249.00	1920.0
250.00	454.0
251.00	561.0
252.10	654.0
253.10	1241.0
254.10	2992.0
255.10	301376.0
256.10	42232.0
257.10	3665.0
258.00	18144.0
259.10	3155.0
259.90	397.0
260.90	541.0
263.00	184.0
264.00	548.0
265.00	7542.0
266.00	1214.0
267.00	256.0
270.10	284.0
271.00	681.0
272.10	682.0
273.00	8888.0
274.00	24032.0
275.10	137600.0
276.10	18392.0
277.00	12520.0
278.00	2204.0
279.00	478.0
280.90	522.0
282.10	598.0
283.10	1602.0
284.00	1309.0
285.10	2017.0
286.10	306.0
288.00	150.0
289.00	600.0
290.00	382.0
291.00	387.0
292.10	588.0
293.00	2867.0
294.00	832.0
296.00	39800.0
297.00	5258.0
297.90	457.0
301.00	542.0
302.00	586.0
303.10	4882.0
304.10	1086.0
304.90	186.0
308.00	728.0
309.10	421.0
310.10	510.0
311.10	163.0
313.00	271.0
314.00	2067.0
315.10	4595.0
316.00	2283.0
317.00	484.0
320.00	353.0
321.10	1282.0
321.90	738.0
323.10	12835.0
324.10	2482.0
325.00	426.0
326.00	264.0
327.00	2885.0
328.00	1125.0
328.90	317.0
332.00	961.0
333.00	1279.0
334.10	8209.0
335.00	2354.0
336.00	299.0
339.00	158.0
339.90	232.0
341.00	1445.0
342.10	310.0
346.00	2473.0
346.90	420.0
351.10	191.0
352.10	4501.0
353.10	2553.0
354.10	4188.0

Instrument :
BNA_M
ClientSampleId :
DFTPP036

355.10	688.0
357.10	160.0
359.10	394.0
363.90	238.0
365.00	19552.0
366.00	2576.0
370.00	364.0
371.00	952.0
372.10	6804.0
373.10	1474.0
374.00	232.0
377.10	150.0
383.00	1724.0
384.10	615.0
385.00	229.0
390.10	783.0
391.00	703.0
392.00	545.0
401.00	377.0
402.10	2655.0
403.10	3581.0
404.10	1598.0
405.10	418.0
421.00	3691.0
422.00	3021.0
423.10	24256.0
424.10	6425.0
425.10	898.0
438.80	163.0
441.10	60696.0
442.10	392704.0
443.10	74952.0
444.10	7264.0
445.10	484.0

Instrument :
BNA_M
ClientSampleId :
DFTPP036

m/z	Abundance
37.00	519.0
38.10	1398.0
39.10	7632.0
40.00	877.0
41.00	510.0
43.10	183.0
44.00	649.0
45.00	322.0
49.10	1096.0
50.10	29400.0
51.10	113384.0
52.10	5558.0
53.10	199.0
55.00	524.0
56.00	3179.0
57.00	8419.0
58.00	360.0
59.00	162.0
61.00	1582.0
62.00	1672.0
63.00	4548.0
64.10	559.0
65.00	1997.0
68.10	2435.0
69.00	115832.0
70.00	757.0
73.10	1625.0
74.10	12507.0
75.10	18656.0
76.10	6570.0
77.10	129544.0
78.10	9082.0
79.00	9954.0
80.00	7042.0
81.10	9661.0
82.10	2567.0
83.00	2181.0
84.00	504.0
85.00	1953.0
86.00	3486.0
87.10	1315.0
88.00	502.0
89.00	275.0
91.00	2240.0
92.00	2754.0
93.00	16936.0
94.00	1145.0
95.00	380.0
96.00	848.0
98.00	13294.0
99.00	10621.0
100.00	921.0
101.00	5810.0
101.90	316.0
103.00	1776.0
104.00	4111.0
105.00	3607.0
106.10	984.0
107.00	46160.0
108.00	6806.0
109.00	1302.0
110.00	77624.0
111.00	10655.0
112.00	1749.0
113.00	398.0
115.00	178.0
116.00	2736.0
117.00	41160.0
118.00	2970.0
119.00	429.0
120.00	481.0
121.10	321.0
122.00	3010.0
123.00	4983.0
124.00	2270.0
125.00	2054.0
127.10	151360.0
128.10	10716.0
129.00	62960.0
130.00	4963.0
130.90	1026.0
132.00	549.0
132.90	568.0
134.00	2034.0
135.00	5490.0
136.00	2026.0
137.00	2736.0
138.00	616.0
138.90	346.0
140.10	1043.0
141.00	9043.0
142.00	2504.0

Instrument :
BNA_M
ClientSampleId :
DFTPP036

143.00	1583.0
144.10	569.0
145.00	650.0
146.00	1393.0
147.00	5007.0
148.00	10912.0
149.00	2044.0
150.00	631.0
151.10	1073.0
151.80	604.0
153.00	2889.0
154.00	1978.0
155.10	5268.0
156.10	7740.0
157.00	1119.0
158.00	1936.0
159.00	1343.0
160.10	2955.0
161.00	3788.0
162.00	1241.0
163.00	631.0
164.10	440.0
165.00	3761.0
166.00	2756.0
167.00	16928.0
168.00	8470.0
169.00	1259.0
170.00	581.0
171.00	809.0
171.90	1566.0
173.10	2382.0
174.00	3908.0
175.10	7504.0
176.10	1748.0
177.00	3749.0
178.00	1203.0
179.00	14401.0
180.00	9431.0
181.00	4229.0
182.10	622.0
183.00	312.0
184.00	1060.0
185.00	6697.0
186.10	52456.0
187.10	14813.0
188.10	1446.0
189.00	3628.0
190.00	603.0
191.00	1876.0
192.10	4586.0
193.10	5588.0
194.00	1221.0
195.00	413.0
196.00	9498.0
197.10	2517.0
198.00	358784.0
199.00	24512.0
200.10	1978.0
201.50	1783.0
203.00	2808.0
204.10	14342.0
205.10	24328.0
206.10	101632.0
207.10	15751.0
208.00	4370.0
209.00	1731.0
210.10	1667.0
211.10	3907.0
212.00	864.0
213.00	275.0
215.00	1423.0
216.00	1959.0
217.00	26832.0
218.00	3628.0
218.90	233.0
220.10	456.0
221.10	12873.0
221.90	3488.0
223.00	6562.0
224.10	58440.0
225.10	14185.0
226.10	2026.0
227.00	26216.0
228.00	3624.0
229.10	4583.0
230.10	748.0
231.10	2186.0
232.10	346.0
232.90	294.0
234.10	1584.0
235.10	2057.0
236.00	936.0
237.10	2409.0
238.00	455.0
239.00	1211.0
240.00	856.0

Instrument :
BNA_M
ClientSampleId :
DFTPP036

241.00	1283.0
242.10	2929.0
243.00	2904.0
244.10	42792.0
245.10	5888.0
246.00	10338.0
247.00	1728.0
247.90	527.0
249.00	1797.0
250.00	267.0
250.90	362.0
251.90	324.0
253.00	1039.0
254.10	2103.0
255.10	240576.0
256.00	34744.0
257.10	2765.0
258.00	15077.0
259.00	2292.0
260.00	545.0
261.00	341.0
264.00	331.0
265.00	6454.0
266.00	650.0
267.00	309.0
270.10	247.0
271.00	782.0
272.10	575.0
273.00	7058.0
274.00	20448.0
275.10	112400.0
276.10	14272.0
277.00	10835.0
278.00	1540.0
278.90	223.0
281.00	764.0
282.00	400.0
283.10	1466.0
284.10	825.0
285.10	1697.0
286.00	337.0
288.00	289.0
289.00	465.0
290.00	299.0
291.00	291.0
292.00	438.0
293.00	2325.0
294.00	633.0
295.10	927.0
296.00	35920.0
297.00	4839.0
298.00	458.0
301.00	468.0
302.00	715.0
303.10	4150.0
304.00	905.0
308.00	479.0
309.00	265.0
310.00	506.0
312.90	307.0
314.10	1587.0
315.10	3885.0
316.10	2277.0
317.00	397.0
321.10	1069.0
322.10	583.0
323.10	11542.0
324.10	2068.0
325.00	239.0
325.90	202.0
327.00	2489.0
328.10	949.0
328.80	157.0
332.00	644.0
333.00	933.0
334.10	7548.0
335.10	2497.0
336.00	159.0
339.00	152.0
340.10	310.0
341.10	1251.0
342.10	504.0
346.00	2606.0
347.00	556.0
349.90	167.0
351.10	157.0
352.10	3424.0
353.10	2192.0
354.10	3394.0
355.00	655.0
355.90	163.0
359.10	192.0
364.10	229.0
365.00	17760.0
365.90	2207.0
370.00	520.0

Instrument :
BNA_M
ClientSampleId :
DFTPP036

371.10	922.0
372.10	6361.0
373.00	1587.0
374.20	171.0
383.00	1945.0
384.10	539.0
385.00	179.0
390.00	659.0
391.00	488.0
392.10	484.0
401.10	442.0
402.00	3042.0
403.10	3300.0
404.00	1559.0
405.10	335.0
415.00	261.0
421.00	3772.0
422.10	3290.0
423.10	22192.0
424.10	5677.0
425.10	737.0
439.50	204.0
441.10	60384.0
442.10	382976.0
443.10	71736.0
444.10	6732.0
445.10	253.0

Instrument :
BNA_M
ClientSampleId :
DFTPP036