

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM020415\
 Data File : BM000419.D
 Acq On : 04 Feb 2015 15:38
 Operator : TP/IZ
 Sample : PB81533BL
 Misc : SBLK10
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK10

Quant Time: Feb 05 07:59:19 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM020315.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 05 07:40:45 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	193580	20.00	ng/ul	0.00
18) Naphthalene-d8	10.54	136	851651	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.40	164	589913	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.14	188	1399380	20.00	ng/ul	0.00
75) Chrysene-d12	21.33	240	1410075	20.00	ng/ul	0.00
83) Perylene-d12	23.59	264	1133741	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.22	96	44338	13.91	ng/uL	0.00
5) Phenol-d5	6.92	99	1004253	72.13	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.09	67	628221	75.51	ng/ul	0.00
9) 2-Chlorophenol-d4	7.28	132	843717	73.77	ng/ul	0.00
13) 4-Methylphenol-d8	8.46	113	891626	74.47	ng/ul	0.00
19) Nitrobenzene-d5	8.91	128	453750	85.45	ng/ul	0.00
22) 2-Nitrophenol-d4	9.63	143	536261	88.11	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.16	165	976543	72.40	ng/ul	0.00
29) 4-Chloroaniline-d4	10.68	131	1388020	111.16	ng/ul	0.00
43) Dimethylphthalate-d6	13.81	166	3631785	83.05	ng/ul	0.00
46) Acenaphthylene-d8	14.09	160	4300646	76.61	ng/ul	0.00
51) 4-Nitrophenol-d4	14.60	143	565292	76.57	ng/ul	0.00
57) Fluorene-d10	15.40	176	3136777	73.45	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.52	200	562790	73.59	ng/ul	0.00
70) Anthracene-d10	17.24	188	4908703	76.21	ng/ul	0.00
76) Pyrene-d10	19.54	212	5298366	74.71	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.46	264	3942034	75.64	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM020415\
Data File : BM000419.D
Acq On : 04 Feb 2015 15:38
Operator : TP/IZ
Sample : PB81533BL
Misc : SBLK10
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK10

Quant Time: Feb 05 07:59:19 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM020315.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 05 07:40:45 2015
Response via : Initial Calibration

