

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM020420\
 Data File : BM024714.D
 Acq On : 04 Feb 2020 19:44
 Operator : CG/JU
 Sample : L1343-02
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 NW-SP

Quant Time: Feb 05 03:50:46 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM012320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 23 14:57:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.42	152	216353	20.00	ng	-0.02
21) Naphthalene-d8	10.18	136	774373	20.00	ng	-0.02
39) Acenaphthene-d10	14.07	164	496747	20.00	ng	-0.02
64) Phenanthrene-d10	16.83	188	1112307	20.00	ng	-0.02
76) Chrysene-d12	21.04	240	1287402	20.00	ng	-0.02
87) Perylene-d12	23.14	264	1509198	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.06	112	1557850	136.94	ng	-0.02
7) Phenol-d6	6.62	99	1857931	118.55	ng	-0.02
23) Nitrobenzene-d5	8.56	82	1226915	77.70	ng	-0.03
42) 2,4,6-Tribromophenol	15.57	330	1146739	141.92	ng	-0.02
45) 2-Fluorobiphenyl	12.69	172	3256432	89.12	ng	-0.02
79) Terphenyl-d14	19.49	244	6491222	94.13	ng	-0.02
Target Compounds						
50) Dimethylphthalate	13.55	163	103547	2.578	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM020420\
Data File : BM024714.D
Acq On : 04 Feb 2020 19:44
Operator : CG/JU
Sample : L1343-02
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
NW-SP

Quant Time: Feb 05 03:50:46 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM012320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 23 14:57:07 2020
Response via : Initial Calibration

