

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM020521\
 Data File : BM028664.D
 Acq On : 05 Feb 2021 15:03
 Operator : CG/JU
 Sample : SSTD0.877
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.877

Quant Time: Feb 05 15:58:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM020521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 05 15:47:53 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.91	152	477	0.40	ng/ul	0.00
2) Naphthalene-d8	10.71	136	1893	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.56	164	990	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.29	188	1968	0.40	ng/ul	0.00
16) Chrysene-d12	21.44	240	1823	0.40	ng/ul	0.00
20) Perylene-d12	23.68	264	1915	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.30	152	2116	0.66	ng/ul	0.00
14) Fluoranthene-d10	19.31	212	3987	0.76	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.77	128	3695	0.651	ng/ul	96
5) 2-Methylnaphthalene	12.38	142	2483	0.643	ng/ul	98
7) Acenaphthylene	14.28	152	3137	0.649	ng/ul	97
8) Acenaphthene	14.62	153	2479	0.685	ng/ul	99
9) Fluorene	15.60	166	2761	0.658	ng/ul	97
11) Pentachlorophenol	16.95	266	393	0.586	ng/ul	98
12) Phenanthrene	17.33	178	4213	0.678	ng/ul	97
13) Anthracene	17.43	178	4050	0.675	ng/ul	96
15) Fluoranthene	19.34	202	4800	0.696	ng/ul	95
17) Pyrene	19.69	202	4935	0.533	ng/ul	99
18) Benzo(a)anthracene	21.42	228	4676	0.636	ng/ul	97
19) Chrysene	21.47	228	4664	0.651	ng/ul	98
21) Benzo(b)fluoranthene	23.01	252	4951	0.615	ng/ul	87
22) Benzo(k)fluoranthene	23.05	252	4988	0.633	ng/ul#	88
23) Benzo(a)pyrene	23.59	252	4438	0.631	ng/ul#	84
24) Indeno(1,2,3-cd)pyrene	25.96	276	5775	0.758	ng/ul#	100
25) Dibenzo(a,h)anthracene	25.97	278	4802	0.757	ng/ul#	89
26) Benzo(g,h,i)perylene	26.65	276	4852	0.759	ng/ul	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM020521\
Data File : BM028664.D
Acq On : 05 Feb 2021 15:03
Operator : CG/JU
Sample : SST00.877
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.877

Quant Time: Feb 05 15:58:50 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM020521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 05 15:47:53 2021
Response via : Initial Calibration

