

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM020521\
 Data File : BM028666.D
 Acq On : 05 Feb 2021 16:21
 Operator : CG/JU
 Sample : SSTD1.678
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD1.678

Quant Time: Feb 05 17:00:59 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM020521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 05 15:47:53 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.91	152	576	0.40	ng/ul	0.00
2) Naphthalene-d8	10.71	136	2393	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.55	164	1341	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.29	188	2742	0.40	ng/ul	0.00
16) Chrysene-d12	21.44	240	2603	0.40	ng/ul	0.00
20) Perylene-d12	23.68	264	2605	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.30	152	5528	1.37	ng/ul	0.00
14) Fluoranthene-d10	19.31	212	11421	1.56	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.76	128	9214	1.285	ng/ul	93
5) 2-Methylnaphthalene	12.38	142	6422	1.316	ng/ul	98
7) Acenaphthylene	14.28	152	8582	1.311	ng/ul	94
8) Acenaphthene	14.62	153	6641	1.354	ng/ul	98
9) Fluorene	15.60	166	7637	1.343	ng/ul	97
11) Pentachlorophenol	16.95	266	1195	1.280	ng/ul	98
12) Phenanthrene	17.33	178	11812	1.364	ng/ul	94
13) Anthracene	17.42	178	11600	1.388	ng/ul	94
15) Fluoranthene	19.34	202	13670	1.423	ng/ul	93
17) Pyrene	19.70	202	13944	1.055	ng/ul	96
18) Benzo(a)anthracene	21.42	228	13418	1.278	ng/ul	95
19) Chrysene	21.47	228	13302	1.300	ng/ul	95
21) Benzo(b)fluoranthene	23.00	252	13698	1.251	ng/ul	78
22) Benzo(k)fluoranthene	23.05	252	13756	1.283	ng/ul#	79
23) Benzo(a)pyrene	23.58	252	12074	1.261	ng/ul#	74
24) Indeno(1,2,3-cd)pyrene	25.95	276	15209	1.468	ng/ul#	100
25) Dibenzo(a,h)anthracene	25.96	278	12770	1.480	ng/ul#	81
26) Benzo(g,h,i)perylene	26.65	276	12526	1.441	ng/ul#	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM020521\
Data File : BM028666.D
Acq On : 05 Feb 2021 16:21
Operator : CG/JU
Sample : SSTD1.678
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD1.678

Quant Time: Feb 05 17:00:59 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM020521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 05 15:47:53 2021
Response via : Initial Calibration

