

Data Path : Z:\HPCHEM1\BNA M\DATA\BM020915\
 Data File : BM000461.D
 Acq On : 10 Feb 2015 10:41
 Operator : TP/IZ
 Sample : G1249-16
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 A4FM9

Quant Time: Feb 10 11:27:36 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM01.2-EPA-BM020615.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 10 05:48:08 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.72	152	221708	20.00	ng/ul	0.00
18) Naphthalene-d8	10.52	136	908701	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.37	164	548797	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.12	188	1183490	20.00	ng/ul	0.00
75) Chrysene-d12	21.31	240	1154731	20.00	ng/ul	0.00
83) Perylene-d12	23.56	264	1080020	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	0.00	96	0	0.00	ng/uL	
5) Phenol-d5	6.89	99	62985	4.28	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.06	67	196952	21.74	ng/ul	0.00
9) 2-Chlorophenol-d4	7.26	132	220474	17.21	ng/ul	0.00
13) 4-Methylphenol-d8	8.43	113	131309	10.76	ng/ul	0.00
19) Nitrobenzene-d5	8.88	128	130856	22.99	ng/ul	0.00
22) 2-Nitrophenol-d4	9.60	143	147494	22.92	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.13	165	271585	19.59	ng/ul	0.00
29) 4-Chloroaniline-d4	10.65	131	47902	3.76	ng/ul	0.00
43) Dimethylphthalate-d6	13.78	166	934229	24.70	ng/ul	0.00
46) Acenaphthylene-d8	14.06	160	1141887	22.10	ng/ul	0.00
51) 4-Nitrophenol-d4	14.57	143	17509	2.83	ng/ul	0.00
57) Fluorene-d10	15.37	176	850936	22.36	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.49	200	131494	20.99	ng/ul	0.00
70) Anthracene-d10	17.22	188	1240397	23.04	ng/ul	0.00
76) Pyrene-d10	19.52	212	1361694	25.02	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.41	264	988001	20.48	ng/ul	0.00

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM020915\
Data File : BM000461.D
Acq On : 10 Feb 2015 10:41
Operator : TP/IZ
Sample : G1249-16
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampled :
A4FM9

Quant Time: Feb 10 11:27:36 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM01.2-EPA-BM020615.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Feb 10 05:48:08 2015
Response via : Initial Calibration

