



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

### **Inframark**

**Project Name: Fresh Kills Landfill Leachate Treatment Plant**

**Project # N/A**

**Chemtech Project # N1283**

**Test Name: Metals Group4**

### **A. Number of Samples and Date of Receipt:**

1 Water sample was received on 01/26/2022.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: Ammonia, BOD5 and Metals Group4. This data package contains results for Metals Group4.

### **C. Analytical Techniques:**

The analysis and digestion of Metals Group4 was based on method 200.7.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike analysis met criteria for all samples.

The Matrix Spike Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

### **E. Additional Comments:**

LLCCV & LLICV are not required for 200.7 method.

In analytical sequence LB118538, The % recovery was outside of acceptance limit for Zinc of CCV14 But No any sample analyze for this CCVs.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_