



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: N1287

MATRIX: /Water

METHOD: 1010B,1030,9012B,9034,9040C,9045D,SM4500-CN
C,E

- | | NA | NO | YES |
|--|----|----|-----|
| 1. Blank Contamination - If yes, list compounds and concentrations in each blank: | | ✓ | |
| 2. Matrix Spike Duplicate Recoveries Met Criteria | | ✓ | |
| If not met, list those compounds and their recoveries which fall outside the acceptable range. | | | |
| The Blank Spike met requirements for all samples. The Matrix Spike (WASTE-CHARACTER-2MS) analysis met criteria for all samples except for Cyanide due to matrix interferences. The Matrix Spike Duplicate (WASTE-CHARACTER-2MSD) analysis met criteria for all samples except for Cyanide due to matrix interferences. | | | |
| 3. Sample Duplicate Analysis Met QC Criteria | | | ✓ |
| If not met, list those compounds and their recoveries which fall outside the acceptable range. | | | |
| 4. Digestion Holding Time Met | | ✓ | |
| If not met, list number of days exceeded for each sample: | | | |
| The Holding Times were met for all samples except for WASTE-CHARACTER of Corrosivity, for WASTE-CHARACTER-2 of Corrosivity, Samples receive out of hold. | | | |

ADDITIONAL COMMENTS:

QA REVIEW

Date