


Benzoic acid

Response Ratio



$$\text{Response} = 2.390\text{e-}001 * \text{Amt} - 7.269\text{e-}002$$


Coef of Det (r^2) = 0.997663 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM020922.M

Calibration Table Last Updated: Wed Feb 09 17:52:36 2022

Butylbenzylphthalate

Response Ratio



$$\text{Response} = 9.299\text{e-}001 * \text{Amt} - 1.965\text{e-}001$$


Coef of Det (r^2) = 0.995823 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM020922.M

Calibration Table Last Updated: Wed Feb 09 17:52:36 2022

Bis(2-ethylhexyl)phthalate

Response Ratio



$$\text{Response} = 7.310\text{e-}001 * \text{Amt} - 1.388\text{e-}001$$

Coef of Det (r^2) = 0.995909 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM020922.M

Calibration Table Last Updated: Wed Feb 09 17:52:36 2022