

Data Path : Z:\HPCHEM1\BNA_M\Data\BM021115\
 Data File : BM000483.D
 Acq On : 11 Feb 2015 18:37
 Operator : TP/IZ
 Sample : SSTD0.403
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.403

Quant Time: Feb 11 23:59:16 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-SIM-BM021115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 11 14:38:48 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
3	Naphthalene	0.977	0.946	3.2	104	0.00
4 SURR	2-Methylnaphthalene-d10	0.536	0.510	4.9	100	0.00
5	2-Methylnaphthalene	0.683	0.652	4.5	102	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
7	Acenaphthylene	1.597	1.547	3.1	99	0.00
8 C	Acenaphthene	1.224	1.214	0.8	100	0.00
9	Fluorene	1.493	1.444	3.3	97	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
11	Pentachlorophenol	0.037	0.024	35.1#	65	0.00
12	Phenanthrene	1.129	1.075	4.8	93	0.00
13	Anthracene	1.014	0.972	4.1	94	0.00
14 SURR	Fluoranthene-d10	0.886	0.794	10.4	88	0.00
15 C	Fluoranthene	1.303	1.160	11.0	89	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	80	0.00
17	Pyrene	1.468	1.585	-8.0	88	0.00
18	Benzo(a)anthracene	1.175	1.086	7.6	79	0.00
19	Chrysene	1.356	1.342	1.0	81	0.00
20 I	Perylene-d12	1.000	1.000	0.0	77	0.00
21	Benzo(b)fluoranthene	1.377	1.327	3.6	76	0.00
22	Benzo(k)fluoranthene	1.464	1.430	2.3	76	0.00
23 C	Benzo(a)pyrene	1.296	1.268	2.2	78	0.00
24	Indeno(1,2,3-cd)pyrene	1.363	1.337	1.9	81	0.00
25	Dibenzo(a,h)anthracene	1.078	1.066	1.1	82	0.00
26	Benzo(g,h,i)perylene	1.181	1.171	0.8	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0