


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM021120\
 Data File : BM024831.D
 Acq On : 11 Feb 2020 11:52
 Operator : CG/JU
 Sample : DFTPP20
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP20

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM020620MA.M
 Title : SVOA CALIBRATION
 Last Update : Wed Feb 12 04:40:07 2020



AutoFind: Scans 2449, 2450, 2451; Background Corrected with Scan 2444

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	24.5	52445	PASS
68	69	0.00	2	1.8	1278	PASS
69	198	0.00	100	32.7	69954	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	40.4	86413	PASS
197	198	0.00	2	0.7	1451	PASS
198	198	100	100	100.0	213760	PASS
199	198	5	9	6.6	14159	PASS
275	198	10	60	31.5	67253	PASS
365	198	1	100	4.3	9240	PASS
441	198	0.01	100	14.7	31354	PASS
442	198	50	100	97.8	209096	PASS
443	442	15	24	19.6	40896	PASS

m/z	Abundance
39.90	776.0
44.00	2834.0
49.00	349.0
73.00	560.0
83.90	544.0
133.00	313.0
191.00	300.0
207.00	2094.0
208.00	568.0
281.10	856.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

m/z	Abundance
38.00	857.0
39.00	3202.0
40.00	616.0
44.00	2105.0
49.00	762.0
50.00	13817.0
51.00	46072.0
52.10	2373.0
54.80	319.0
56.00	1628.0
57.00	3406.0
60.90	713.0
61.90	794.0
63.10	1772.0
64.90	775.0
68.10	1128.0
69.00	61000.0
73.10	1400.0
74.00	6402.0
75.00	9613.0
76.10	3387.0
77.00	67160.0
78.00	4485.0
78.90	5025.0
80.00	4287.0
81.00	4899.0
81.90	1239.0
83.00	1032.0
84.00	367.0
84.90	1144.0
86.00	1425.0
87.00	932.0
88.00	403.0
91.00	1075.0
92.00	1234.0
93.00	9524.0
94.00	744.0
96.00	456.0
98.00	7137.0
99.00	5643.0
100.00	640.0
100.90	3071.0
103.00	1323.0
104.00	1856.0
105.00	1417.0
106.20	571.0
107.00	21736.0
108.00	3700.0
109.10	819.0
110.00	37864.0
111.00	5354.0
112.00	638.0
114.80	323.0
115.90	1276.0
117.00	17952.0
118.00	1947.0
119.00	427.0
119.90	353.0
122.00	1330.0
123.00	2473.0
123.90	933.0
124.90	774.0
125.80	303.0
127.00	73640.0
128.00	5803.0
129.00	28264.0
130.00	2157.0
130.80	509.0
132.00	327.0
133.90	820.0
135.00	2358.0
136.10	889.0
137.00	1583.0
139.90	375.0
141.00	4085.0
141.90	901.0
142.90	785.0
144.80	351.0
146.10	684.0
146.90	2067.0
147.90	4514.0
149.00	1263.0
151.20	626.0
151.90	356.0
153.00	1408.0
153.90	1220.0
155.00	2761.0
156.00	3575.0
157.30	534.0
157.80	708.0
159.00	523.0
160.00	1458.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

161.00	1699.0
161.90	634.0
162.10	634.0
163.10	375.0
164.90	2060.0
166.00	1258.0
167.00	8365.0
167.90	6047.0
169.00	571.0
169.80	409.0
171.80	932.0
173.00	982.0
173.90	1797.0
175.00	3380.0
176.00	1175.0
176.90	1539.0
178.00	773.0
178.90	5561.0
180.00	4707.0
181.00	1955.0
182.20	348.0
183.80	557.0
185.00	2749.0
186.00	21352.0
187.00	6624.0
188.00	861.0
189.00	1342.0
190.00	335.0
191.00	811.0
191.90	2048.0
192.90	2263.0
194.10	550.0
196.00	4672.0
197.10	1161.0
197.90	163072.0
198.90	10819.0
200.00	1005.0
201.30	733.0
203.00	1174.0
204.00	5840.0
205.00	10311.0
206.00	44104.0
207.00	8258.0
208.00	1864.0
208.80	794.0
209.80	577.0
210.80	1495.0
214.90	602.0
215.90	1028.0
216.90	12107.0
217.90	1926.0
221.00	6272.0
221.70	1799.0
223.00	2778.0
224.00	23048.0
225.00	6514.0
226.00	887.0
227.00	10294.0
227.90	1874.0
228.90	2361.0
229.90	303.0
231.00	976.0
233.90	712.0
235.00	456.0
235.90	640.0
236.90	1538.0
239.00	566.0
239.90	385.0
241.00	605.0
242.10	1033.0
243.10	1165.0
244.00	17256.0
245.10	2164.0
245.90	4215.0
247.10	657.0
248.90	850.0
252.80	449.0
254.20	975.0
255.00	98544.0
256.00	16928.0
256.90	1214.0
257.90	5693.0
258.90	814.0
264.90	2093.0
266.00	302.0
273.00	3106.0
274.00	7149.0
275.00	44760.0
276.00	6689.0
276.90	4083.0
277.90	776.0
281.00	904.0
281.80	331.0
282.20	324.0
283.00	757.0
284.00	481.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

285.00	697.0
292.10	305.0
293.00	676.0
293.90	342.0
296.00	11656.0
296.90	1612.0
301.90	433.0
303.00	1817.0
303.90	567.0
314.00	597.0
314.90	1646.0
316.00	1084.0
321.10	542.0
323.00	4400.0
324.10	791.0
326.90	745.0
328.00	421.0
333.90	3103.0
335.00	1105.0
341.10	625.0
346.00	757.0
352.00	1421.0
353.00	718.0
354.00	1089.0
355.20	411.0
365.00	5870.0
366.00	841.0
372.00	2399.0
373.00	670.0
383.10	640.0
402.00	704.0
403.00	1117.0
404.00	472.0
421.10	1190.0
422.10	820.0
423.00	6685.0
424.00	2132.0
424.90	356.0
441.00	17632.0
442.00	112664.0
443.00	22488.0
444.00	2098.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

m/z	Abundance
37.00	563.0
38.10	807.0
39.00	3908.0
39.90	920.0
43.90	2545.0
49.00	956.0
50.00	17912.0
51.00	66320.0
52.10	2682.0
55.90	2429.0
57.00	4421.0
57.90	386.0
60.90	948.0
61.90	1221.0
63.00	3121.0
64.00	451.0
65.10	1561.0
67.90	1530.0
69.00	88336.0
73.10	1554.0
74.00	9619.0
75.00	14142.0
76.00	4475.0
77.00	91360.0
78.00	6734.0
79.00	7123.0
80.00	5285.0
81.00	7195.0
81.90	2073.0
83.10	1451.0
83.90	490.0
85.00	1340.0
86.00	2547.0
87.00	1011.0
87.90	322.0
91.00	1518.0
92.10	2095.0
93.00	12238.0
93.90	830.0
96.00	749.0
98.00	10768.0
99.00	9117.0
99.90	649.0
101.00	4727.0
102.90	1402.0
104.00	2339.0
105.00	2657.0
106.10	787.0
107.00	30184.0
108.00	5727.0
109.10	922.0
110.00	52600.0
111.00	8009.0
112.00	972.0
112.90	360.0
116.00	1918.0
117.00	28136.0
117.90	2397.0
118.90	352.0
120.00	544.0
122.00	2022.0
123.00	3068.0
123.90	1276.0
125.00	1499.0
127.00	107912.0
128.00	7937.0
129.00	43024.0
130.00	3637.0
130.90	878.0
132.10	575.0
133.00	448.0
133.90	1416.0
135.00	3849.0
136.00	1583.0
136.90	1882.0
139.90	678.0
141.00	5839.0
142.00	1756.0
142.90	1088.0
144.10	469.0
145.00	378.0
145.90	1690.0
146.90	3076.0
147.90	7867.0
148.90	1922.0
149.90	547.0
150.80	761.0
151.40	604.0
152.90	1891.0
153.90	992.0
155.00	2811.0
156.00	4898.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

156.90	927.0
157.90	1115.0
158.90	941.0
159.90	2471.0
161.00	2663.0
161.90	719.0
163.90	310.0
165.00	2221.0
166.00	1704.0
167.00	11709.0
168.00	8104.0
168.90	872.0
169.90	494.0
171.10	503.0
171.90	1088.0
173.00	1457.0
173.90	3140.0
175.00	5510.0
176.00	1387.0
176.90	2516.0
178.00	1157.0
178.90	9767.0
180.00	7101.0
181.00	3438.0
182.00	735.0
183.90	836.0
185.00	4379.0
186.00	35112.0
187.00	9837.0
188.20	938.0
189.00	1829.0
189.80	409.0
191.00	1565.0
192.00	3015.0
193.00	3964.0
194.00	961.0
195.00	499.0
196.00	7831.0
197.00	1723.0
197.90	258688.0
198.90	18224.0
200.00	1369.0
201.50	1276.0
203.00	2301.0
204.00	10263.0
205.00	17264.0
206.00	70024.0
207.00	12594.0
208.00	2819.0
209.00	1019.0
210.90	2897.0
211.90	459.0
215.00	1058.0
216.00	1705.0
217.00	19088.0
217.90	2279.0
218.90	443.0
220.20	697.0
221.00	10928.0
222.90	3907.0
224.00	40352.0
225.00	10278.0
226.00	779.0
227.00	16728.0
227.90	2083.0
229.00	2879.0
229.80	471.0
230.90	1377.0
232.90	657.0
234.00	1236.0
234.80	1092.0
235.90	661.0
237.10	1701.0
239.00	827.0
240.00	676.0
241.00	935.0
241.90	1937.0
243.10	2689.0
244.00	32752.0
245.00	4450.0
246.00	7124.0
247.00	1064.0
248.90	1551.0
249.90	338.0
250.80	329.0
252.00	462.0
253.10	906.0
254.10	1249.0
255.00	172544.0
256.00	23928.0
257.00	1838.0
258.00	9457.0
259.00	2230.0
259.90	314.0
260.90	301.0
264.00	309.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

264.90	3777.0
265.90	675.0
270.90	537.0
272.00	626.0
272.90	6120.0
274.00	13326.0
275.00	80696.0
276.00	10979.0
277.00	6370.0
277.90	1337.0
279.00	325.0
281.00	1061.0
283.00	963.0
284.10	725.0
285.00	1059.0
292.10	472.0
293.10	1600.0
293.90	727.0
294.90	595.0
296.00	22008.0
296.90	3259.0
301.90	432.0
302.90	2092.0
304.00	828.0
309.80	448.0
314.00	1007.0
315.00	2852.0
315.90	1292.0
321.00	952.0
322.00	391.0
323.00	7609.0
323.90	1449.0
327.00	1455.0
328.00	1050.0
331.80	372.0
332.90	519.0
334.00	5517.0
335.00	1607.0
341.00	1191.0
341.90	346.0
346.00	1756.0
346.80	400.0
352.00	2004.0
352.90	1168.0
354.00	3143.0
355.00	682.0
365.00	10967.0
365.90	1242.0
371.00	479.0
372.10	4426.0
373.00	1229.0
383.10	1333.0
390.10	521.0
391.00	558.0
401.00	506.0
402.00	1314.0
402.90	1774.0
404.10	883.0
421.00	1822.0
422.00	1853.0
423.00	14384.0
424.00	4458.0
425.00	594.0
441.00	36096.0
442.00	243136.0
443.00	46616.0
444.00	4946.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

m/z	Abundance
38.10	583.0
38.90	2706.0
40.00	968.0
44.00	2389.0
49.10	468.0
50.00	11893.0
51.00	44944.0
52.10	2374.0
54.90	391.0
56.00	1509.0
57.00	3583.0
61.00	830.0
62.00	957.0
63.00	2326.0
65.10	1179.0
68.10	1176.0
69.00	60528.0
73.00	1178.0
74.00	6248.0
75.00	9680.0
76.10	3806.0
77.00	64600.0
78.00	4454.0
79.00	5248.0
80.00	3886.0
81.00	5444.0
81.90	1501.0
83.00	1018.0
83.80	367.0
84.90	1082.0
85.90	1975.0
87.00	896.0
87.90	337.0
90.90	1081.0
92.00	1489.0
93.00	8611.0
93.90	716.0
96.00	431.0
98.00	6774.0
98.90	5434.0
100.00	590.0
101.00	2513.0
103.00	1097.0
104.00	1785.0
105.00	2035.0
105.80	511.0
107.00	22672.0
108.00	3690.0
108.90	729.0
110.00	39232.0
110.90	6307.0
112.10	739.0
116.00	1070.0
117.00	21488.0
117.90	1351.0
119.90	515.0
122.00	1917.0
123.00	2382.0
123.90	1318.0
125.00	1136.0
127.00	77688.0
128.00	5414.0
129.00	32040.0
130.00	3017.0
130.90	451.0
132.10	351.0
133.00	381.0
133.90	1145.0
135.00	2915.0
135.90	1116.0
137.00	1365.0
137.90	412.0
139.90	547.0
141.00	4944.0
142.00	1565.0
143.00	1160.0
146.00	520.0
147.00	2496.0
147.90	5152.0
148.90	1420.0
150.00	410.0
151.00	459.0
151.80	676.0
152.90	1374.0
153.80	1118.0
155.00	2413.0
156.00	3790.0
156.90	929.0
158.00	663.0
158.90	802.0
159.90	1421.0
160.90	2883.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

161.80	873.0
162.90	316.0
163.90	358.0
164.90	1578.0
166.10	1584.0
167.00	9404.0
168.00	5190.0
168.90	857.0
170.00	455.0
171.00	492.0
171.90	973.0
172.90	1519.0
174.00	2044.0
175.00	4494.0
176.00	1597.0
176.90	1845.0
178.10	539.0
178.90	8055.0
180.00	5800.0
180.90	2501.0
182.00	479.0
183.90	568.0
185.00	4163.0
186.00	29272.0
187.00	7928.0
187.90	926.0
189.00	1946.0
189.90	389.0
190.90	1224.0
192.00	2684.0
193.00	3044.0
194.00	692.0
194.90	373.0
195.90	5961.0
197.10	1471.0
197.90	219520.0
199.00	13434.0
200.00	1273.0
201.30	1076.0
203.00	1680.0
204.00	7685.0
205.00	13428.0
206.00	58536.0
207.00	10600.0
207.90	2269.0
208.90	1137.0
210.20	1092.0
211.10	2450.0
211.90	524.0
214.80	482.0
216.10	1200.0
216.90	17392.0
218.00	2322.0
221.00	10214.0
221.90	2175.0
223.00	3819.0
224.00	34864.0
225.00	8006.0
226.00	972.0
226.90	14687.0
227.90	2161.0
228.90	2609.0
230.00	499.0
231.00	1200.0
233.90	1017.0
235.00	989.0
235.80	692.0
237.00	1565.0
239.00	698.0
239.90	589.0
241.10	985.0
241.90	1670.0
243.00	2059.0
244.00	27560.0
245.00	3512.0
245.90	6409.0
247.00	1044.0
249.00	1091.0
252.00	397.0
252.90	703.0
255.00	151488.0
256.00	22368.0
256.90	1955.0
257.90	9377.0
259.00	1277.0
259.90	334.0
265.00	3844.0
266.10	494.0
270.90	376.0
272.00	342.0
273.00	4906.0
274.00	12093.0
275.00	76304.0
276.00	10794.0
277.00	6719.0
278.00	1131.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

281.00	1144.0
282.00	442.0
282.90	961.0
284.00	549.0
284.90	1200.0
292.80	1368.0
293.90	612.0
294.90	494.0
296.00	19312.0
297.00	3257.0
301.00	421.0
302.00	718.0
303.00	2707.0
303.80	560.0
308.00	318.0
314.10	1346.0
315.00	2016.0
316.00	1363.0
320.90	814.0
322.10	401.0
323.00	6400.0
324.20	974.0
327.00	1537.0
327.90	928.0
332.00	536.0
333.00	760.0
333.90	5252.0
335.00	1275.0
341.00	970.0
342.10	547.0
345.90	1581.0
346.90	392.0
352.00	2476.0
352.90	1742.0
354.00	2666.0
355.00	654.0
365.00	10884.0
366.00	1300.0
370.10	478.0
370.80	561.0
372.00	4027.0
373.10	828.0
383.00	1233.0
384.10	495.0
390.00	686.0
402.00	1691.0
403.00	2533.0
403.90	843.0
421.00	2097.0
422.10	1958.0
423.00	14638.0
424.00	3621.0
424.90	581.0
441.00	40336.0
442.00	271488.0
443.00	53584.0
444.10	4822.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20