


Data Path : Z:\HPCHEM1\BNA_M\Data\BM021215\
Data File : BM000499.D
Acq On : 12 Feb 2015 12:52
Operator : TP/IZ
Sample : DFTPP50
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP50

Integration File: RTEINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-SIM-BM021215.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Feb 12 16:45:55 2015



AutoFind: Scans 2525, 2526, 2527; Background Corrected with Scan 2518

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.7	38658	PASS
68	69	0.00	2	1.8	586	PASS
69	198	0.00	100	31.6	33370	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	43.3	45621	PASS
197	198	0.00	2	0.7	789	PASS
198	198	100	100	100.0	105453	PASS
199	198	5	9	7.0	7413	PASS
275	198	10	60	24.0	25320	PASS
365	198	1	100	3.2	3380	PASS
441	443	0.01	100	79.3	16045	PASS
442	198	50	100	97.7	103042	PASS
443	442	15	24	19.6	20229	PASS

Scan 2518 (17.498 min): BM000499.D
DFTPP50

m/z	Abundance
40.10	490.0
44.00	2403.0
207.00	709.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

m/z	Abundance
38.00	497.0
39.10	3442.0
39.90	626.0
44.00	2607.0
49.00	493.0
50.10	11765.0
51.10	48080.0
52.00	2808.0
55.00	333.0
56.00	1374.0
57.00	3252.0
61.00	426.0
62.00	622.0
63.10	1603.0
64.10	411.0
65.10	994.0
68.10	667.0
69.00	41104.0
72.90	417.0
74.00	4741.0
75.00	6958.0
76.10	2553.0
77.10	45272.0
78.10	3673.0
78.90	3621.0
80.00	2487.0
81.00	4055.0
82.00	1174.0
82.90	807.0
85.00	760.0
86.00	1156.0
87.10	408.0
91.10	1018.0
92.10	1037.0
93.00	5795.0
93.90	611.0
95.90	412.0
98.00	4561.0
99.00	4262.0
101.00	2216.0
102.90	790.0
104.00	1432.0
105.10	1255.0
106.00	399.0
107.00	15708.0
108.00	2453.0
109.00	421.0
110.00	28112.0
111.00	5125.0
112.00	521.0
115.90	1068.0
117.10	12598.0
118.00	826.0
120.00	364.0
121.90	1003.0
122.90	1603.0
124.00	1078.0
124.90	664.0
127.00	51888.0
128.00	3602.0
129.00	19808.0
130.00	2313.0
131.10	506.0
133.90	667.0
135.00	1747.0
136.00	731.0
137.00	1037.0
137.70	386.0
140.10	321.0
141.00	2203.0
141.90	832.0
143.10	699.0
146.00	548.0
147.00	1512.0
148.00	2897.0
149.00	795.0
151.10	544.0
152.90	1039.0
154.00	702.0
154.90	1474.0
156.10	2370.0
158.10	726.0
158.90	407.0
160.00	885.0
160.90	1612.0
162.10	340.0
165.00	1126.0
166.00	1159.0
167.00	4952.0
168.10	3818.0
168.90	528.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

172.00	653.0
173.00	577.0
174.00	1303.0
175.00	1699.0
176.00	525.0
177.00	944.0
178.10	458.0
179.00	4251.0
180.00	3242.0
180.90	1570.0
184.00	534.0
185.10	2029.0
186.00	15043.0
187.00	4596.0
188.20	605.0
188.90	1093.0
190.90	480.0
192.00	1144.0
193.00	1720.0
196.00	3460.0
197.10	954.0
198.00	115864.0
198.90	7531.0
199.90	489.0
201.20	503.0
202.10	305.0
202.90	732.0
204.10	4020.0
205.00	5320.0
206.10	28232.0
207.10	4875.0
208.00	1305.0
209.20	508.0
210.20	424.0
211.00	1360.0
211.80	654.0
215.10	512.0
216.10	854.0
217.00	8195.0
218.00	935.0
220.90	2848.0
221.70	908.0
222.00	885.0
223.00	1577.0
224.10	16223.0
225.10	4099.0
225.90	768.0
227.10	6252.0
227.90	1407.0
229.00	1449.0
231.20	562.0
234.10	568.0
234.80	428.0
235.90	325.0
236.90	1180.0
241.10	682.0
242.00	1113.0
243.10	1043.0
244.10	11585.0
245.10	1900.0
246.00	2312.0
247.00	480.0
249.00	558.0
253.00	408.0
254.10	653.0
255.00	62304.0
256.10	9694.0
257.00	1108.0
258.00	4460.0
259.20	391.0
265.10	1372.0
272.00	311.0
273.00	1898.0
274.00	4691.0
275.10	24488.0
276.00	3541.0
277.00	2785.0
277.90	749.0
285.00	498.0
293.00	768.0
296.00	8160.0
297.00	1118.0
302.90	726.0
315.00	663.0
316.00	554.0
323.10	2149.0
324.20	332.0
327.00	348.0
332.90	349.0
334.00	1522.0
335.10	424.0
345.90	527.0
352.10	946.0
353.10	563.0
354.00	934.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

365.00	3445.0
365.80	473.0
371.00	349.0
372.10	1309.0
373.10	404.0
401.80	517.0
403.00	732.0
421.00	634.0
422.10	678.0
423.10	3843.0
424.10	2102.0
424.90	344.0
441.00	13683.0
442.00	88240.0
443.00	17376.0
444.10	1478.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

m/z	Abundance
38.00	542.0
39.00	3948.0
40.00	796.0
44.00	2468.0
49.00	350.0
50.10	10183.0
51.10	44728.0
52.10	2228.0
54.90	338.0
56.00	1381.0
57.00	3662.0
61.00	652.0
62.00	542.0
63.00	1797.0
63.90	309.0
65.10	793.0
68.10	691.0
69.00	39144.0
73.10	586.0
74.00	3781.0
75.00	6119.0
76.00	2041.0
77.10	42992.0
78.10	2649.0
79.00	3435.0
80.00	2457.0
81.00	3031.0
81.90	778.0
82.90	789.0
84.90	668.0
85.90	1060.0
87.10	449.0
91.00	760.0
91.90	965.0
93.00	5793.0
94.00	467.0
96.00	334.0
98.00	4641.0
99.00	3376.0
100.00	326.0
101.00	1865.0
102.90	636.0
104.00	1442.0
105.00	1641.0
105.90	441.0
107.00	15271.0
108.00	2378.0
109.00	508.0
110.00	25968.0
111.00	4520.0
112.00	372.0
116.00	849.0
117.00	12558.0
118.00	930.0
122.00	1002.0
123.00	1744.0
124.10	872.0
125.00	708.0
125.80	370.0
127.00	53992.0
128.00	4279.0
129.00	20952.0
130.10	2372.0
131.00	478.0
132.00	364.0
134.10	387.0
135.00	1915.0
136.00	579.0
137.10	969.0
139.90	319.0
141.00	3330.0
142.00	871.0
143.00	767.0
146.00	544.0
147.10	1624.0
148.00	3708.0
149.00	1007.0
151.00	350.0
151.60	308.0
152.90	1042.0
153.90	739.0
155.00	1598.0
156.10	2571.0
157.10	554.0
158.00	440.0
159.00	603.0
160.10	955.0
160.90	1349.0
161.90	331.0
165.00	772.0
166.00	1196.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

167.00	5218.0
168.10	3331.0
169.10	546.0
170.90	389.0
172.00	512.0
173.10	846.0
174.00	1144.0
175.10	2600.0
176.10	548.0
177.00	1404.0
178.10	381.0
179.00	5048.0
180.00	3358.0
181.00	1239.0
183.80	407.0
185.00	2391.0
186.00	17104.0
187.10	4491.0
188.00	834.0
189.10	1269.0
191.00	1106.0
192.00	1335.0
193.10	2238.0
194.00	301.0
196.00	3619.0
197.00	845.0
198.00	126032.0
199.00	9450.0
199.90	671.0
201.40	453.0
202.10	330.0
202.90	1010.0
204.10	4931.0
205.00	7921.0
206.10	33424.0
207.10	4261.0
208.00	1220.0
209.00	436.0
210.00	748.0
210.90	1214.0
214.90	403.0
216.10	691.0
217.00	8437.0
218.00	1026.0
221.00	4219.0
221.80	1123.0
223.00	2318.0
224.10	19184.0
225.00	4823.0
226.10	640.0
227.00	8051.0
228.00	1185.0
229.10	1401.0
231.10	812.0
233.90	808.0
235.10	672.0
236.00	385.0
237.00	998.0
238.90	315.0
241.00	456.0
242.00	818.0
243.00	1037.0
244.10	12867.0
245.00	1525.0
246.10	3066.0
247.00	413.0
248.90	891.0
253.00	373.0
254.10	504.0
255.00	71168.0
256.00	11146.0
257.10	932.0
258.10	3887.0
259.00	888.0
265.00	1683.0
266.10	308.0
273.00	1949.0
274.00	5137.0
275.10	31528.0
276.00	4258.0
277.00	2394.0
278.00	674.0
285.10	556.0
293.10	770.0
296.00	8774.0
297.10	1268.0
303.10	1231.0
313.90	419.0
314.90	1235.0
316.00	947.0
323.10	3125.0
324.00	719.0
327.00	715.0
334.10	1805.0
335.20	388.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

345.90	354.0
352.20	593.0
353.00	632.0
354.10	934.0
365.00	4278.0
365.90	812.0
370.80	315.0
372.10	1943.0
383.00	444.0
402.00	914.0
402.90	877.0
404.20	327.0
420.90	1051.0
422.00	999.0
423.00	6021.0
424.10	2624.0
424.90	539.0
441.00	19736.0
442.00	125080.0
443.10	25392.0
444.00	1729.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

m/z	Abundance
38.10	377.0
39.10	1526.0
40.00	652.0
44.00	2614.0
50.10	6064.0
51.10	23168.0
52.10	1010.0
56.00	564.0
57.00	1601.0
61.10	353.0
62.10	340.0
63.00	841.0
64.90	398.0
68.10	402.0
69.00	19864.0
73.00	398.0
74.00	2301.0
75.00	3039.0
76.00	1261.0
77.10	23808.0
78.00	1253.0
79.10	1672.0
80.00	1416.0
81.00	1583.0
82.00	632.0
83.00	467.0
85.00	559.0
85.90	653.0
86.90	331.0
91.00	468.0
92.10	438.0
93.00	3285.0
98.00	2576.0
99.00	2650.0
100.00	313.0
101.00	1142.0
102.90	424.0
104.00	589.0
105.00	865.0
107.00	8229.0
108.10	1193.0
110.00	14310.0
111.00	2275.0
112.10	367.0
116.00	570.0
117.00	6698.0
117.80	454.0
118.20	518.0
122.10	587.0
122.90	978.0
124.10	393.0
125.00	428.0
127.00	30984.0
128.10	2727.0
129.00	12057.0
129.90	1323.0
133.90	406.0
135.00	869.0
135.80	480.0
137.00	598.0
141.00	1834.0
143.00	305.0
147.00	1024.0
148.00	1767.0
148.90	487.0
153.00	652.0
153.90	572.0
155.00	1157.0
156.10	1503.0
156.90	307.0
159.90	326.0
161.10	732.0
162.00	310.0
164.90	953.0
166.00	480.0
167.10	3165.0
168.00	2049.0
169.00	310.0
172.90	319.0
173.90	792.0
175.10	1227.0
176.30	446.0
177.00	592.0
179.00	2576.0
180.00	1850.0
181.10	646.0
185.10	1812.0
186.00	8920.0
187.00	3589.0
189.00	369.0
191.00	509.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

192.10	990.0
193.00	1335.0
196.20	1645.0
197.00	568.0
198.00	74464.0
199.00	5260.0
200.10	565.0
201.40	361.0
203.00	478.0
204.00	2982.0
205.00	4920.0
206.10	20080.0
207.00	3219.0
208.00	723.0
211.10	824.0
215.10	383.0
217.00	5748.0
218.00	738.0
221.20	2823.0
221.80	689.0
223.00	1159.0
224.10	11024.0
225.10	2452.0
226.10	447.0
227.00	4708.0
227.90	670.0
229.00	717.0
231.10	387.0
235.00	520.0
237.00	760.0
242.10	591.0
243.10	751.0
244.00	9066.0
245.00	1010.0
246.00	1842.0
246.90	353.0
248.90	365.0
254.10	381.0
255.00	45912.0
256.00	5708.0
257.10	566.0
258.00	2966.0
258.90	403.0
264.90	949.0
273.00	1574.0
274.00	3929.0
275.10	19944.0
276.00	2665.0
277.00	1603.0
283.20	376.0
284.90	320.0
292.80	513.0
296.00	5267.0
297.00	611.0
303.10	652.0
315.00	804.0
321.00	378.0
323.10	1787.0
324.00	448.0
327.00	429.0
334.00	1348.0
334.90	380.0
346.00	392.0
351.90	623.0
352.90	514.0
354.10	595.0
365.10	2419.0
366.00	371.0
372.10	991.0
402.00	518.0
403.00	780.0
403.90	391.0
421.00	921.0
422.00	675.0
423.00	4692.0
424.00	1687.0
441.00	14717.0
442.00	95808.0
443.00	17920.0
444.00	1620.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50