

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM021221\
 Data File : BM028731.D
 Acq On : 12 Feb 2021 15:41
 Operator : CG/JU
 Sample : SSTD3.205
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.2005

Quant Time: Feb 12 16:12:40 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM021221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 12 14:23:47 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.89	152	397	0.40	ng/ul	0.00
4) Naphthalene-d8	10.70	136	1537	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.54	164	788	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.28	188	1531	0.40	ng/ul	0.00
17) Chrysene-d12	21.42	240	1501	0.40	ng/ul	0.00
23) Perylene-d12	23.66	264	1478	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.20	96	1420	3.22	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	12.29	152	7066	3.18	ng/ul	0.00
18) Fluoranthene-d10	19.29	212	12673	2.67	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.24	88	1466	3.208	ng/ul#	76
5) Naphthalene	10.75	128	12795	2.910	ng/ul#	92
7) 2-Methylnaphthalene	12.36	142	8758	3.117	ng/ul	100
8) 1-Methylnaphthalene	12.58	142	8376	3.123	ng/ul	100
10) Acenaphthylene	14.26	152	11255	3.300	ng/ul	93
11) Acenaphthene	14.60	153	8531	2.924	ng/ul	97
12) Fluorene	15.59	166	9567	3.000	ng/ul	99
14) Pentachlorophenol	16.93	266	1350	3.918	ng/ul	99
15) Phenanthrene	17.32	178	14404	2.976	ng/ul	94
16) Anthracene	17.41	178	13802	3.264	ng/ul	94
19) Fluoranthene	19.32	202	16199	2.462	ng/ul	95
20) Pyrene	19.68	202	16504	2.539	ng/ul	97
21) Benzo(a)anthracene	21.41	228	16068	3.119	ng/ul	97
22) Chrysene	21.46	228	16397	2.902	ng/ul	96
24) Benzo(b)fluoranthene	22.98	252	17644	2.775	ng/ul	87
25) Benzo(k)fluoranthene	23.03	252	17774	2.952	ng/ul#	85
26) Benzo(a)pyrene	23.56	252	15505	2.972	ng/ul#	82
27) Indeno(1,2,3-cd)pyrene	25.91	276	20474	2.994	ng/ul#	100
28) Dibenzo(a,h)anthracene	25.92	278	17111	3.026	ng/ul#	88
29) Benzo(g,h,i)perylene	26.60	276	17195	2.806	ng/ul	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM021221\
Data File : BM028731.D
Acq On : 12 Feb 2021 15:41
Operator : CG/JU
Sample : SSTD3.205
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.2005

Quant Time: Feb 12 16:12:40 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM021221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 12 14:23:47 2021
Response via : Initial Calibration

