

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM021221\
 Data File : BM028731.D
 Acq On : 12 Feb 2021 15:41
 Operator : CG/JU
 Sample : SSTD3.205
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.2005

Quant Time: Feb 12 16:13:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM021221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 12 14:23:47 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.89	152	397	0.40	ng/ul	0.00
4) Naphthalene-d8	10.70	136	1537	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.54	164	788	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.28	188	1531	0.40	ng/ul	0.00
17) Chrysene-d12	21.42	240	1501	0.40	ng/ul	0.00
23) Perylene-d12	23.66	264	1478	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.20	96	1420	3.22	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	0.00	152	0d	0.00	ng/ul	
18) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul	
Target Compounds						
2) 1,4-Dioxane	3.24	88	1466	3.208	ng/ul#	76
14) Pentachlorophenol	16.93	266	1350	3.918	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM021221\
Data File : BM028731.D
Acq On : 12 Feb 2021 15:41
Operator : CG/JU
Sample : SSTD3.205
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.2005

Quant Time: Feb 12 16:13:29 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM021221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 12 14:23:47 2021
Response via : Initial Calibration

