

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM021221\
 Data File : BM028729.D
 Acq On : 12 Feb 2021 14:28
 Operator : CG/JU
 Sample : SST00.803
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST00.8003

Quant Time: Feb 12 15:02:52 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM021221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 12 14:23:47 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.89	152	382	0.40	ng/ul	0.00
4) Naphthalene-d8	10.70	136	1522	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.54	164	810	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.28	188	1598	0.40	ng/ul	0.00
17) Chrysene-d12	21.42	240	1442	0.40	ng/ul	0.00
23) Perylene-d12	23.66	264	1404	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.21	96	370	0.87	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.29	152	1848	0.84	ng/ul	0.00
18) Fluoranthene-d10	19.29	212	3382	0.74	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.24	88	363	0.826	ng/ul#	87
5) Naphthalene	10.75	128	3295	0.757	ng/ul	96
7) 2-Methylnaphthalene	12.36	142	2273	0.817	ng/ul	100
8) 1-Methylnaphthalene	12.58	142	2186	0.823	ng/ul	100
10) Acenaphthylene	14.26	152	2857	0.815	ng/ul	97
11) Acenaphthene	14.60	153	2249	0.750	ng/ul	97
12) Fluorene	15.59	166	2534	0.773	ng/ul	99
14) Pentachlorophenol	16.94	266	330	0.918	ng/ul	97
15) Phenanthrene	17.32	178	3887	0.769	ng/ul	96
16) Anthracene	17.41	178	3585	0.812	ng/ul	97
19) Fluoranthene	19.32	202	4282	0.677	ng/ul	98
20) Pyrene	19.68	202	4354	0.697	ng/ul	99
21) Benzo(a)anthracene	21.41	228	4046	0.817	ng/ul	99
22) Chrysene	21.46	228	4155	0.766	ng/ul	97
24) Benzo(b)fluoranthene	22.98	252	4303	0.712	ng/ul	93
25) Benzo(k)fluoranthene	23.03	252	4384	0.767	ng/ul	93
26) Benzo(a)pyrene	23.56	252	3759	0.759	ng/ul#	91
27) Indeno(1,2,3-cd)pyrene	25.91	276	4939	0.760	ng/ul#	100
28) Dibenzo(a,h)anthracene	25.92	278	4112	0.766	ng/ul	94
29) Benzo(g,h,i)perylene	26.60	276	4206	0.723	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM021221\
Data File : BM028729.D
Acq On : 12 Feb 2021 14:28
Operator : CG/JU
Sample : SST00.803
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.8003

Quant Time: Feb 12 15:02:52 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM021221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 12 14:23:47 2021
Response via : Initial Calibration

