

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021224\
 Data File : BM044050.D
 Acq On : 12 Feb 2024 11:05
 Operator : MA/JU
 Sample : PB155297BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB155297BL

Quant Time: Feb 12 12:07:39 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM020824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 09 02:42:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.910	152	480632	20.000	ng	0.00
21) Naphthalene-d8	10.710	136	1939678	20.000	ng	0.00
39) Acenaphthene-d10	14.545	164	1069939	20.000	ng	-0.01
64) Phenanthrene-d10	17.298	188	2066222	20.000	ng	0.00
76) Chrysene-d12	21.503	240	1431763	20.000	ng	0.00
86) Perylene-d12	23.897	264	1447277	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.481	112	4253137	128.703	ng	0.00
7) Phenol-d6	7.075	99	5267790	125.286	ng	0.00
23) Nitrobenzene-d5	9.081	82	3102234	84.868	ng	0.00
42) 2,4,6-Tribromophenol	16.045	330	1393764	120.313	ng	0.00
45) 2-Fluorobiphenyl	13.180	172	6262372	82.605	ng	0.00
79) Terphenyl-d14	19.945	244	7324103	89.675	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021224\
Data File : BM044050.D
Acq On : 12 Feb 2024 11:05
Operator : MA/JU
Sample : PB155297BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB155297BL

Quant Time: Feb 12 12:07:39 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM020824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 09 02:42:01 2024
Response via : Initial Calibration

