

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021225\
 Data File : BM049632.D
 Acq On : 12 Feb 2025 23:20
 Operator : RC/JU
 Sample : Q1336-06
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 E29D2

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 02/19/2025
 Supervised By :Jagrut Upadhyay 02/19/2025

Quant Time: Feb 14 17:22:32 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 12 16:40:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.813	152	7910	0.400	ng/ul	0.00
4) Naphthalene-d8	10.605	136	22244	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.450	164	13021	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.191	188	28941m	0.400	ng/ul	0.00
17) Chrysene-d12	21.365	240	19383m	0.400	ng/ul	0.00
23) Perylene-d12	23.651	264	17619m	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.252	96	47109	5.294	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.200	152	10549	0.345	ng/ul	0.00
18) Fluoranthene-d10	19.224	212	26179	0.444	ng/ul	0.00
Target Compounds						Qvalue
14) Pentachlorophenol	16.840	266	156m	0.026	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021225\
Data File : BM049632.D
Acq On : 12 Feb 2025 23:20
Operator : RC/JU
Sample : Q1336-06
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
E29D2

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 02/19/2025
Supervised By :Jagrut Upadhyay 02/19/2025

Quant Time: Feb 14 17:22:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 12 16:40:36 2025
Response via : Initial Calibration

