

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021225\
 Data File : BM049629.D
 Acq On : 12 Feb 2025 21:33
 Operator : RC/JU
 Sample : Q1336-03
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 E29C6

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 02/19/2025
 Supervised By :Jagrut Upadhyay 02/19/2025

Quant Time: Feb 14 16:32:46 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 12 16:40:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.813	152	5699	0.400	ng/ul	0.00
4) Naphthalene-d8	10.605	136	15679	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.450	164	9151	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.191	188	19420m	0.400	ng/ul	0.00
17) Chrysene-d12	21.365	240	12808m	0.400	ng/ul	0.00
23) Perylene-d12	23.654	264	11984m	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.252	96	39893	6.223	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.200	152	8492	0.394	ng/ul	0.00
18) Fluoranthene-d10	19.215	212	21003	0.538	ng/ul	0.00
Target Compounds						Qvalue
14) Pentachlorophenol	16.844	266	130m	0.033	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021225\
Data File : BM049629.D
Acq On : 12 Feb 2025 21:33
Operator : RC/JU
Sample : Q1336-03
Misc :
ALS Vial : 18 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
E29C6

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 02/19/2025
Supervised By :Jagrut Upadhyay 02/19/2025


Quant Time: Feb 14 16:32:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 12 16:40:36 2025
Response via : Initial Calibration





#14
 Pentachlorophenol
 Concen: 0.033 ng/ul m
 RT: 16.844 min Scan# 30
 Delta R.T. 0.004 min
 Lab File: BM049629.D
 Acq: 12 Feb 2025 21:33

Instrument :
 BNA_M
 ClientSampleId :
 E29C6



Tgt Ion:266 Resp: 130
 Ion Ratio Lower Upper
 266 100
 264 2.3 50.3 75.5
 268 3.1 49.2 73.8

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 02/19/2025
 Supervised By :Jagrut Upadhyay 02/19/2025

