

Data Path : Z:\HPCHEM1\BNA_M\Data\BM021315\
 Data File : BM000520.D
 Acq On : 13 Feb 2015 15:24
 Operator : TP/IZ
 Sample : SSTD0.406
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.491

Quant Time: Feb 13 16:43:04 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-SIM-BM021215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 12 16:45:55 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	116	0.00
3	Naphthalene	0.938	0.954	-1.7	118	0.00
4 SURR	2-Methylnaphthalene-d10	0.510	0.498	2.4	114	0.00
5	2-Methylnaphthalene	0.634	0.637	-0.5	117	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	111	0.00
7	Acenaphthylene	1.505	1.546	-2.7	118	0.00
8 C	Acenaphthene	1.155	1.213	-5.0	116	0.00
9	Fluorene	1.396	1.448	-3.7	117	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	114	0.00
11	Pentachlorophenol	0.051	0.042	17.6	102	-0.02
12	Phenanthrene	1.086	1.109	-2.1	119	0.00
13	Anthracene	0.977	0.998	-2.1	118	0.00
14 SURR	Fluoranthene-d10	0.841	0.834	0.8	112	0.00
15 C	Fluoranthene	1.211	1.250	-3.2	115	-0.01
16 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
17	Pyrene	1.439	1.489	-3.5	117	-0.01
18	Benzo(a)anthracene	1.128	1.130	-0.2	110	0.00
19	Chrysene	1.300	1.341	-3.2	110	0.00
20 I	Perylene-d12	1.000	1.000	0.0	103	0.00
21	Benzo(b)fluoranthene	1.352	1.491	-10.3	114	-0.01
22	Benzo(k)fluoranthene	1.345	1.290	4.1	102	0.00
23 C	Benzo(a)pyrene	1.264	1.281	-1.3	106	-0.01
24	Indeno(1,2,3-cd)pyrene	1.410	1.354	4.0	99	0.00
25	Dibenzo(a,h)anthracene	1.114	1.051	5.7	98	-0.01
26	Benzo(g,h,i)perylene	1.244	1.211	2.7	101	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0