

Data Path : Z:\HPCHEM1\BNA_M\Data\BM021315\
 Data File : BM000525.D
 Acq On : 13 Feb 2015 19:15
 Operator : TP/IZ
 Sample : SSTD0.407
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.407

Quant Time: Feb 14 01:40:41 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-SIM-BM021215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 12 16:45:55 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	119	0.00
3	Naphthalene	0.938	0.946	-0.9	120	0.00
4 SURR	2-Methylnaphthalene-d10	0.510	0.501	1.8	117	0.00
5	2-Methylnaphthalene	0.634	0.634	0.0	119	-0.01
6 I	Acenaphthene-d10	1.000	1.000	0.0	122	0.00
7	Acenaphthylene	1.505	1.465	2.7	122	-0.01
8 C	Acenaphthene	1.155	1.126	2.5	118	-0.02
9	Fluorene	1.396	1.356	2.9	120	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	121	0.00
11	Pentachlorophenol	0.051	0.038	25.5#	97	-0.02
12	Phenanthrene	1.086	1.086	0.0	123	0.00
13	Anthracene	0.977	0.984	-0.7	123	0.00
14 SURR	Fluoranthene-d10	0.841	0.841	0.0	120	0.00
15 C	Fluoranthene	1.211	1.243	-2.6	121	-0.01
16 I	Chrysene-d12	1.000	1.000	0.0	115	-0.01
17	Pyrene	1.439	1.468	-2.0	122	-0.01
18	Benzo(a)anthracene	1.128	1.134	-0.5	117	-0.01
19	Chrysene	1.300	1.332	-2.5	116	0.00
20 I	Perylene-d12	1.000	1.000	0.0	109	0.00
21	Benzo(b)fluoranthene	1.352	1.347	0.4	109	-0.01
22	Benzo(k)fluoranthene	1.345	1.416	-5.3	118	-0.01
23 C	Benzo(a)pyrene	1.264	1.287	-1.8	113	-0.01
24	Indeno(1,2,3-cd)pyrene	1.410	1.352	4.1	105	-0.01
25	Dibenzo(a,h)anthracene	1.114	1.061	4.8	105	-0.02
26	Benzo(g,h,i)perylene	1.244	1.205	3.1	106	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0