

Data Path : Z:\HPCHEM1\BNA_M\Data\BM021317\
 Data File : BM008988.D
 Acq On : 13 Feb 2017 18:12
 Operator : SJ/MA
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02042

Quant Time: Feb 14 00:08:09 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM020817.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 08 15:01:57 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	125	0.00
2	1,4-Dioxane	0.575	0.564	1.9	124	0.00
3 S	1,4-Dioxane-d8	0.520	0.502	3.5	113	0.00
4	Benzaldehyde	1.371	1.480	-8.0	122	0.00
5 S	Phenol-d5	1.777	1.810	-1.9	125	0.00
6	Phenol	1.909	1.930	-1.1	122	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.293	1.345	-4.0	124	0.00
8	Bis(2-Chloroethyl)ether	1.497	1.539	-2.8	123	0.00
9 S	2-Chlorophenol-d4	1.194	1.259	-5.4	127	0.00
10	2-Chlorophenol	1.251	1.293	-3.4	123	0.00
11	2-Methylphenol	1.320	1.333	-1.0	120	0.00
12	2,2'-oxybis(1-Chloropropane	2.397	2.504	-4.5	125	0.00
13 S	4-Methylphenol-d8	1.327	1.353	-2.0	126	0.00
14	Acetophenone	2.307	2.325	-0.8	121	0.00
15 P	N-Nitroso-di-n-propylamine	1.327	1.397	-5.3	122	0.00
16	4-Methylphenol	1.441	1.443	-0.1	119	0.00
17	Hexachloroethane	0.607	0.633	-4.3	128	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	123	0.00
19 S	Nitrobenzene-d5	0.137	0.142	-3.6	122	0.00
20	Nitrobenzene	0.478	0.500	-4.6	123	0.00
21	Isophorone	0.827	0.842	-1.8	121	0.00
22 S	2-Nitrophenol-d4	0.150	0.158	-5.3	121	0.00
23 C	2-Nitrophenol	0.165	0.169	-2.4	121	0.00
24	2,4-Dimethylphenol	0.417	0.424	-1.7	120	0.00
25	Bis(2-Chloroethoxy)methane	0.478	0.489	-2.3	119	0.00
26 S	2,4-Dichlorophenol-d3	0.320	0.320	0.0	118	0.00
27 C	2,4-Dichlorophenol	0.320	0.329	-2.8	120	0.00
28	Naphthalene	0.997	1.028	-3.1	123	0.00
29 S	4-Chloroaniline-d4	0.325	0.364	-12.0	126	0.00
30	4-Chloroaniline	0.341	0.384	-12.6	127	0.00
31 C	Hexachlorobutadiene	0.263	0.270	-2.7	123	0.00
32	Caprolactam	0.097	0.093	4.1	115	0.00
33 C	4-Chloro-3-methylphenol	0.363	0.377	-3.9	122	0.00
34	2-Methylnaphthalene	0.742	0.763	-2.8	123	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	120	0.00
36	1,2,4,5-Tetrachlorobenzene	0.653	0.657	-0.6	117	0.00
37	Hexachlorocyclopentadiene	0.417	0.422	-1.2	128	0.00
38 C	2,4,6-Trichlorophenol	0.372	0.384	-3.2	120	0.00
39	2,4,5-Trichlorophenol	0.406	0.427	-5.2	123	0.00
40	1,1'-Biphenyl	1.388	1.440	-3.7	120	0.00
41	2-Chloronaphthalene	1.093	1.131	-3.5	121	0.00
42	2-Nitroaniline	0.371	0.390	-5.1	118	0.00
43 S	Dimethylphthalate-d6	1.368	1.407	-2.9	118	0.00
44	Dimethylphthalate	1.368	1.398	-2.2	118	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM021317\
 Data File : BM008988.D
 Acq On : 13 Feb 2017 18:12
 Operator : SJ/MA
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02042

Quant Time: Feb 14 00:08:09 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM020817.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 08 15:01:57 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.249	0.270	-8.4	123	0.00
46 S	Acenaphthylene-d8	1.652	1.738	-5.2	120	0.00
47	Acenaphthylene	1.639	1.728	-5.4	122	0.00
48	3-Nitroaniline	0.248	0.264	-6.5	121	0.00
49 C	Acenaphthene	1.171	1.208	-3.2	120	0.00
50	2,4-Dinitrophenol	0.176	0.164	6.8	122	0.00
51 S	4-Nitrophenol-d4	0.241	0.240	0.4	122	0.00
52	4-Nitrophenol	0.290	0.290	0.0	118	0.00
53	Dibenzofuran	1.738	1.754	-0.9	117	0.00
54	2,4-Dinitrotoluene	0.379	0.401	-5.8	121	0.00
55	2,3,4,6-Tetrachlorophenol	0.377	0.395	-4.8	122	0.00
56	Diethylphthalate	1.405	1.427	-1.6	118	0.00
57 S	Fluorene-d10	1.271	1.299	-2.2	118	0.00
58	Fluorene	1.424	1.462	-2.7	118	0.00
59	4-Chlorophenyl-phenylether	0.767	0.785	-2.3	118	0.00
60	4-Nitroaniline	0.287	0.298	-3.8	121	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.119	0.119	0.0	122	0.00
63	4,6-Dinitro-2-methylphenol	0.128	0.123	3.9	118	0.00
64	N-Nitrosodiphenylamine	0.568	0.581	-2.3	117	0.00
65	4-Bromophenyl-phenylether	0.224	0.230	-2.7	118	0.00
66	Hexachlorobenzene	0.252	0.252	0.0	117	0.00
67	Atrazine	0.225	0.218	3.1	114	0.00
68 C	Pentachlorophenol	0.151	0.149	1.3	121	0.00
69	Phenanthrene	1.085	1.117	-2.9	118	0.00
70 S	Anthracene-d10	0.941	0.957	-1.7	116	0.00
71	Anthracene	1.109	1.139	-2.7	116	0.00
72	Carbazole	0.987	0.985	0.2	116	0.00
73	Di-n-butylphthalate	1.106	1.153	-4.2	119	0.00
74 C	Fluoranthene	1.323	1.343	-1.5	114	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
76 S	Pyrene-d10	0.857	0.901	-5.1	112	0.00
77	Pyrene	1.099	1.156	-5.2	114	0.00
78	Butylbenzylphthalate	0.388	0.426	-9.8	118	0.00
79	3,3'-Dichlorobenzidine	0.379	0.400	-5.5	111	0.00
80	Benzo(a)anthracene	1.133	1.175	-3.7	111	0.00
81	Bis(2-ethylhexyl)phthalate	0.570	0.624	-9.5	119	0.00
82	Chrysene	1.088	1.108	-1.8	110	0.00
83 I	Perylene-d12	1.000	1.000	0.0	108	0.00
84	Di-n-octyl phthalate	1.102	1.140	-3.4	115	0.00
85	Benzo(b)fluoranthene	1.196	1.255	-4.9	110	0.00
86	Benzo(k)fluoranthene	1.137	1.144	-0.6	106	0.00
87 S	Benzo(a)pyrene-d12	0.901	0.931	-3.3	109	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM021317\
Data File : BM008988.D
Acq On : 13 Feb 2017 18:12
Operator : SJ/MA
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02042

Quant Time: Feb 14 00:08:09 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM020817.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Feb 08 15:01:57 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.136	1.166	-2.6	107	0.00
89	Indeno(1,2,3-cd)pyrene	1.251	1.271	-1.6	107	0.00
90	Dibenzo(a,h)anthracene	1.078	1.089	-1.0	107	0.00
91	Benzo(g,h,i)perylene	1.051	1.054	-0.3	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0