




Quantitation Report (Qedit)

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM021318\
 Data File : BM014217.D
 Acq On : 13 Feb 2018 10:52
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 14 05:40:58 2018
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM020718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 07 17:16:49 2018
 Response via : Initial Calibration



TIC: BM014217.D

(69) Pentachlorophenol (C)




16.610min (-0.005) 0.00ng

response 60737

Ion	Exp%	Act%
265.70	100	100
268.00	65.00	45.64#
264.00	61.20	49.92
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM021318\
Data File : BM014217.D
Acq On : 13 Feb 2018 10:52
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 14 05:40:58 2018
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM020718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 07 17:16:49 2018
Response via : Initial Calibration



(76) Benzidine

19.227min (-0.000) 0.00ng

response 309900

Ion	Exp%	Act%
184.00	100	100
185.00	14.10	12.74
183.00	11.10	11.24
0.00	0.00	0.00