


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM021417\
 Data File : BM008992.D
 Acq On : 14 Feb 2017 08:55
 Operator : SJ/MA
 Sample : DFTPP36
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP36

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM121416.M
 Title : SVOA CALIBRATION
 Last Update : Mon Dec 19 15:00:56 2016



AutoFind: Scans 2571, 2572, 2573; Background Corrected with Scan 2565

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.1	282752	PASS
68	69	0.00	2	1.8	5365	PASS
69	198	0.00	100	44.3	290986	PASS
70	69	0.00	2	0.4	1039	PASS
127	198	10	80	42.3	277440	PASS
197	198	0.00	2	0.7	4416	PASS
198	198	100	100	100.0	656384	PASS
199	198	5	9	6.5	42925	PASS
275	198	10	60	28.2	185066	PASS
365	198	1	100	4.2	27645	PASS
441	443	0.01	100	83.9	74461	PASS
442	198	50	100	71.6	470208	PASS
443	442	15	24	18.9	88744	PASS

m/z	Abundance
37.00	839.0
38.10	3246.0
39.10	22024.0
40.00	2449.0
41.00	1580.0
43.20	355.0
44.00	916.0
44.90	754.0
47.00	338.0
49.00	2758.0
50.00	92448.0
51.00	354688.0
52.00	16608.0
52.90	1016.0
55.00	1795.0
56.00	11228.0
57.00	23656.0
58.00	1490.0
58.90	761.0
61.10	4655.0
62.10	5557.0
63.00	16320.0
63.90	1495.0
65.00	8473.0
65.90	1545.0
66.90	441.0
68.00	7160.0
69.00	358400.0
69.90	2163.0
71.10	540.0
71.80	364.0
73.10	4947.0
74.00	37600.0
75.00	57352.0
76.00	18680.0
77.10	404032.0
78.10	28512.0
79.00	28992.0
80.00	20312.0
81.00	29880.0
82.00	7524.0
83.00	6388.0
84.00	783.0
85.10	4952.0
85.90	8353.0
86.90	3536.0
88.00	1278.0
88.90	774.0
91.00	5833.0
92.00	8009.0
93.00	52016.0
94.00	3233.0
95.00	967.0
96.00	2272.0
97.00	1021.0
98.00	38576.0
99.00	30544.0
100.00	1999.0
101.00	15067.0
102.00	724.0
103.00	5101.0
104.00	11340.0
105.00	10168.0
105.90	2350.0
107.00	118552.0
108.00	15812.0
109.00	2593.0
110.00	179904.0
111.00	30608.0
112.00	4249.0
113.10	1032.0
113.90	606.0
116.00	5992.0
117.00	118560.0
118.00	10003.0
118.90	772.0
120.00	1632.0
121.10	324.0
122.10	8080.0
122.90	11205.0
124.00	4733.0
124.90	4697.0
126.10	908.0
127.00	327104.0
128.00	24672.0
129.00	139648.0
130.00	12241.0
130.90	1657.0
132.00	1527.0
132.90	1033.0
133.90	4162.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

135.00	12090.0
136.00	5820.0
137.00	5802.0
138.10	1283.0
138.90	1282.0
140.00	1747.0
141.00	18552.0
142.10	5000.0
143.00	5586.0
143.90	1242.0
145.00	1251.0
145.90	3925.0
147.10	9854.0
148.00	25344.0
148.90	4606.0
150.00	1251.0
150.90	2587.0
151.70	2173.0
153.00	5974.0
154.00	5357.0
155.00	11451.0
156.00	17760.0
157.10	2905.0
157.90	4761.0
158.90	4104.0
159.90	6010.0
161.00	8695.0
162.10	2755.0
162.90	597.0
164.20	1429.0
164.90	9192.0
166.00	5936.0
167.00	43328.0
168.00	24360.0
169.00	3818.0
170.10	1047.0
171.20	1605.0
171.90	4107.0
173.00	3370.0
174.00	7298.0
175.00	18016.0
176.10	4546.0
177.00	6225.0
178.10	3162.0
179.00	30152.0
180.00	20184.0
181.00	8634.0
182.10	1779.0
182.70	401.0
184.00	2651.0
185.00	15151.0
186.00	103344.0
187.00	32832.0
188.10	4461.0
188.90	9541.0
190.00	859.0
190.90	4856.0
191.90	9020.0
193.00	11203.0
194.00	2719.0
195.20	979.0
196.00	18136.0
196.70	3983.0
197.10	5454.0
197.90	719168.0
199.00	45280.0
200.00	3442.0
201.40	3430.0
203.00	6868.0
204.00	30448.0
205.00	50824.0
206.00	205632.0
207.10	29496.0
208.10	7765.0
209.00	3250.0
210.00	3984.0
211.10	7435.0
211.80	1152.0
212.90	908.0
215.00	2817.0
215.90	4436.0
217.00	60032.0
217.90	7373.0
218.90	788.0
219.90	923.0
221.00	22680.0
221.90	8612.0
223.00	12569.0
224.00	121632.0
225.00	31408.0
226.10	4259.0
227.00	56720.0
228.00	9515.0
229.00	9437.0
230.00	1032.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

230.90	4422.0
232.00	626.0
232.90	975.0
234.00	3382.0
234.90	3305.0
236.00	3684.0
237.10	3839.0
238.00	815.0
239.00	2817.0
240.00	1202.0
240.90	2979.0
242.00	5339.0
243.00	5886.0
244.00	85632.0
245.00	11987.0
246.00	20872.0
247.00	3657.0
248.00	1303.0
249.00	3473.0
250.00	449.0
251.10	1164.0
251.80	492.0
252.90	2244.0
254.10	3582.0
255.00	447488.0
256.00	69024.0
257.00	7659.0
258.00	29296.0
258.90	3863.0
260.00	746.0
263.10	566.0
264.10	1040.0
264.90	12216.0
266.00	1656.0
266.90	426.0
268.00	732.0
270.00	405.0
271.20	917.0
271.80	992.0
273.00	15592.0
274.00	36064.0
275.00	189952.0
276.00	25792.0
277.00	17672.0
278.00	3965.0
279.00	1171.0
279.90	357.0
281.20	1456.0
281.80	374.0
282.20	465.0
283.10	3002.0
283.90	1230.0
285.10	3540.0
285.90	451.0
288.10	436.0
289.00	854.0
290.00	373.0
290.80	350.0
292.00	868.0
293.00	3291.0
294.00	1307.0
295.00	1298.0
296.00	55688.0
297.00	8177.0
297.90	632.0
300.80	897.0
302.20	792.0
303.10	7218.0
304.10	1619.0
305.00	499.0
308.10	420.0
308.90	587.0
310.00	648.0
314.00	3472.0
314.90	5860.0
316.00	3905.0
317.00	366.0
320.90	1229.0
321.70	351.0
323.10	14719.0
324.10	2747.0
327.00	2893.0
328.00	1695.0
328.80	501.0
330.20	351.0
331.90	370.0
333.00	2144.0
334.10	10009.0
335.00	2617.0
341.00	1327.0
342.00	618.0
345.00	339.0
345.90	3621.0
347.00	1085.0
352.00	4852.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

353.10	4156.0
354.10	4821.0
355.10	751.0
363.10	345.0
365.00	24128.0
365.90	3936.0
366.90	345.0
370.00	811.0
370.80	1202.0
372.00	4969.0
373.20	1487.0
382.80	1376.0
383.00	1376.0
384.10	1096.0
385.00	500.0
389.90	652.0
391.00	450.0
392.00	322.0
401.10	494.0
402.00	3105.0
403.00	4276.0
404.10	2043.0
404.90	463.0
420.90	2946.0
422.10	2404.0
423.10	22008.0
424.10	7867.0
425.10	1075.0
437.90	423.0
438.90	690.0
441.10	63240.0
442.10	404480.0
443.10	76168.0
444.10	7711.0
445.00	398.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

m/z	Abundance
37.00	1527.0
38.10	3471.0
39.00	23488.0
40.00	2290.0
41.00	866.0
42.90	426.0
44.00	712.0
44.90	607.0
47.00	542.0
49.10	1808.0
50.00	84728.0
51.00	326400.0
52.10	14293.0
53.00	457.0
55.00	1471.0
56.00	9234.0
57.00	21608.0
57.90	1094.0
61.00	5380.0
62.10	4051.0
63.10	13667.0
64.10	2031.0
65.00	6100.0
66.00	1979.0
66.90	419.0
68.10	5311.0
69.00	340736.0
70.10	1656.0
73.10	3995.0
74.00	39208.0
75.00	58376.0
76.00	17416.0
77.10	388672.0
78.10	26056.0
79.00	28752.0
80.00	17968.0
81.00	27984.0
82.10	6287.0
83.00	4762.0
83.90	352.0
85.10	4918.0
86.00	8237.0
87.00	4432.0
88.00	1234.0
89.10	498.0
91.00	6823.0
92.10	7862.0
93.00	45624.0
94.00	2637.0
95.10	999.0
96.00	2416.0
97.10	532.0
98.00	37856.0
99.00	29632.0
99.90	2713.0
100.90	14410.0
102.00	833.0
103.00	5706.0
103.90	9844.0
105.00	9683.0
106.00	2784.0
107.00	116560.0
108.00	17744.0
108.90	3992.0
110.00	185600.0
111.00	34280.0
112.10	3844.0
113.00	2065.0
115.10	487.0
116.10	6832.0
117.00	121472.0
118.00	8826.0
119.10	569.0
120.00	911.0
121.00	425.0
122.00	9149.0
123.00	10792.0
123.90	5490.0
125.00	6465.0
126.10	1230.0
127.00	322944.0
128.00	25736.0
129.00	131200.0
130.00	11204.0
131.00	3058.0
131.90	1278.0
132.90	997.0
134.00	5795.0
135.00	11682.0
136.00	5879.0
137.00	6656.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

138.10	1604.0
139.10	1042.0
139.90	1491.0
141.00	18808.0
142.10	5206.0
143.00	4216.0
143.80	1204.0
144.10	1086.0
145.10	831.0
146.00	3321.0
147.00	14040.0
148.00	30696.0
149.00	4794.0
150.10	1154.0
151.00	2046.0
152.00	615.0
153.00	6934.0
154.00	3745.0
155.00	10700.0
156.00	15580.0
157.00	1669.0
157.90	4100.0
159.00	3084.0
160.00	6153.0
161.00	8510.0
162.10	2829.0
163.10	1143.0
164.00	693.0
165.00	6431.0
166.00	4814.0
167.00	49640.0
168.00	24400.0
169.00	3630.0
170.10	1517.0
171.00	2628.0
171.90	4256.0
173.00	4995.0
174.00	9277.0
175.10	15727.0
176.20	3385.0
176.90	8603.0
178.00	2823.0
179.00	33104.0
180.00	23920.0
181.00	8943.0
182.10	2438.0
182.70	509.0
184.10	2387.0
185.00	17936.0
186.00	118768.0
187.00	33608.0
188.00	5186.0
189.00	8857.0
189.90	1263.0
191.00	3738.0
192.00	10588.0
193.10	11560.0
194.00	3010.0
195.00	1409.0
196.00	23440.0
197.00	5258.0
198.00	794496.0
199.00	54496.0
200.10	3306.0
201.60	2759.0
202.90	4827.0
204.10	31232.0
205.00	54672.0
206.00	229184.0
207.00	30736.0
207.90	7375.0
209.00	3131.0
210.00	4517.0
210.90	6863.0
213.10	1018.0
215.00	3417.0
216.00	5787.0
217.00	65720.0
218.00	7789.0
219.10	1030.0
221.10	24024.0
221.90	9063.0
223.00	14529.0
224.00	129944.0
225.00	35256.0
226.00	3754.0
227.00	64600.0
228.00	8285.0
229.00	13788.0
229.90	1316.0
231.10	4270.0
231.90	572.0
233.00	1365.0
234.00	2868.0
235.10	4062.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

236.00	3436.0
237.00	5659.0
238.00	763.0
238.80	1405.0
240.00	2387.0
241.00	4206.0
242.00	7650.0
243.10	6927.0
244.00	90784.0
245.00	13189.0
246.00	19456.0
247.00	4303.0
248.00	1039.0
249.00	4628.0
250.00	813.0
251.10	1019.0
252.10	1167.0
252.90	1883.0
253.90	3464.0
255.00	507392.0
256.00	76728.0
257.00	6426.0
258.00	36088.0
259.00	5310.0
260.10	524.0
260.90	1181.0
263.10	397.0
265.00	14346.0
266.00	2131.0
267.00	331.0
269.80	553.0
271.00	734.0
272.10	1733.0
273.00	14964.0
274.00	41408.0
275.00	216000.0
276.00	29072.0
277.00	23512.0
277.90	3069.0
278.90	607.0
280.90	898.0
282.00	1516.0
282.90	2118.0
284.00	1654.0
285.00	3463.0
285.90	1079.0
288.90	1748.0
290.10	766.0
291.10	492.0
292.00	1174.0
293.00	5819.0
294.10	1441.0
294.90	1823.0
296.00	68880.0
297.00	12425.0
297.90	1266.0
301.00	746.0
302.10	1224.0
303.00	7492.0
304.00	1627.0
305.00	376.0
307.90	645.0
308.90	808.0
309.90	634.0
312.00	356.0
313.00	823.0
313.90	3793.0
314.90	7185.0
316.00	4674.0
316.90	844.0
321.00	1528.0
322.00	1497.0
323.10	22800.0
324.00	3604.0
327.00	3770.0
328.00	2115.0
328.90	507.0
331.90	1681.0
333.10	1746.0
334.00	12225.0
334.90	2389.0
336.00	522.0
339.00	324.0
340.20	504.0
341.00	3172.0
342.00	645.0
345.90	4740.0
347.00	858.0
352.10	5536.0
353.00	5367.0
354.10	5770.0
355.00	2376.0
363.20	441.0
365.00	32792.0
365.90	3954.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

366.90	446.0
370.00	470.0
371.10	1318.0
372.10	12043.0
373.00	2148.0
383.10	3038.0
384.00	685.0
390.00	707.0
391.10	392.0
392.00	416.0
400.90	450.0
401.90	4039.0
403.00	5325.0
403.80	1726.0
415.00	446.0
421.10	5924.0
422.10	4189.0
423.10	29920.0
424.00	7983.0
425.10	1662.0
436.00	314.0
437.40	504.0
438.90	434.0
439.70	357.0
441.10	90592.0
442.10	568448.0
443.10	105936.0
444.10	7840.0
444.90	577.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

m/z	Abundance
35.90	411.0
37.00	954.0
38.00	2268.0
39.10	11720.0
40.00	1230.0
41.00	619.0
43.00	316.0
44.00	1374.0
45.10	766.0
49.00	836.0
50.00	44272.0
51.00	167168.0
52.10	8110.0
53.00	923.0
55.00	843.0
56.00	4117.0
57.00	11657.0
58.00	444.0
59.20	510.0
60.90	1793.0
62.00	2756.0
63.10	7971.0
63.90	1216.0
65.00	3995.0
65.90	806.0
68.10	3626.0
69.00	173824.0
69.90	954.0
71.10	330.0
71.80	338.0
73.00	3632.0
74.00	19384.0
75.10	27472.0
76.10	10211.0
77.10	197952.0
78.10	13499.0
79.00	13963.0
80.00	10487.0
81.00	15464.0
81.90	3098.0
83.10	3122.0
84.90	3076.0
86.10	4243.0
87.10	2612.0
87.90	814.0
88.90	305.0
90.90	3773.0
92.00	3342.0
93.00	23928.0
93.90	1356.0
95.00	536.0
96.10	890.0
96.80	443.0
98.00	21792.0
99.00	15326.0
99.90	808.0
101.00	8406.0
103.00	2172.0
104.00	3701.0
105.00	5144.0
105.90	1344.0
107.00	65000.0
108.00	10225.0
108.90	1516.0
110.00	102296.0
111.00	17072.0
112.10	1919.0
112.90	611.0
115.00	313.0
115.90	4207.0
117.00	71384.0
118.00	4604.0
119.10	844.0
120.00	757.0
120.90	538.0
121.90	4131.0
123.00	6677.0
123.90	4567.0
125.00	2915.0
126.00	349.0
127.00	182272.0
128.00	13077.0
129.00	73544.0
130.00	5663.0
130.90	1565.0
132.10	1380.0
133.10	676.0
134.00	2123.0
135.00	6330.0
136.00	2566.0
136.90	4138.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

138.10	767.0
139.00	491.0
139.50	490.0
140.00	1344.0
141.00	12566.0
141.90	4435.0
142.90	3227.0
143.90	669.0
144.90	615.0
146.10	2615.0
147.00	6661.0
148.00	13688.0
149.00	2324.0
150.30	625.0
151.10	1087.0
151.90	970.0
153.00	3293.0
154.00	3385.0
155.10	5928.0
156.00	9569.0
157.10	2113.0
158.00	2639.0
159.00	1705.0
159.90	3730.0
161.00	4993.0
161.80	1116.0
163.00	423.0
163.90	750.0
164.90	4258.0
166.00	4333.0
167.00	28864.0
168.00	14571.0
169.00	2856.0
169.70	637.0
170.20	386.0
171.00	1198.0
172.00	2364.0
173.00	3580.0
174.10	4861.0
175.10	9771.0
175.90	2707.0
177.00	5318.0
178.10	1380.0
179.00	18464.0
180.00	15277.0
181.00	5380.0
182.10	891.0
183.90	2092.0
185.00	9758.0
186.00	63288.0
187.00	20208.0
188.00	2884.0
188.90	5824.0
190.10	1423.0
191.00	1541.0
192.00	5891.0
193.00	6964.0
194.00	1298.0
195.20	951.0
196.00	14753.0
196.90	2538.0
198.00	455488.0
199.00	29000.0
200.10	3617.0
201.50	1424.0
202.90	4191.0
204.00	19464.0
205.00	30728.0
206.10	129264.0
207.00	21296.0
208.00	7283.0
209.00	2875.0
209.90	2138.0
211.10	5800.0
211.80	1287.0
212.90	387.0
215.00	1022.0
216.00	3320.0
216.90	41416.0
218.00	4505.0
219.00	392.0
219.90	548.0
221.00	13147.0
221.90	6706.0
222.90	8470.0
224.00	86552.0
225.00	16584.0
226.10	2573.0
227.00	40840.0
228.00	6558.0
228.90	6887.0
230.10	1291.0
231.10	2262.0
234.00	2665.0
235.00	2577.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

235.90	2159.0
237.00	3651.0
237.80	342.0
238.80	967.0
239.90	1227.0
240.90	2093.0
242.00	3771.0
243.10	4086.0
244.00	61848.0
245.00	8165.0
246.00	12843.0
247.00	2647.0
247.90	593.0
248.90	1926.0
250.00	1140.0
250.80	730.0
251.10	698.0
252.20	821.0
252.90	1518.0
254.00	2704.0
255.00	323456.0
256.00	45960.0
257.00	4127.0
258.00	20000.0
259.00	2505.0
261.10	550.0
264.10	409.0
265.00	8820.0
266.00	1203.0
267.00	463.0
270.90	544.0
272.00	1065.0
273.00	9236.0
274.00	25872.0
275.00	149248.0
276.00	17928.0
277.00	14153.0
278.00	1853.0
279.10	581.0
280.90	1042.0
282.00	524.0
283.10	2075.0
284.00	1128.0
285.10	2681.0
286.00	542.0
288.90	729.0
289.90	459.0
290.90	502.0
291.90	609.0
293.00	3392.0
293.90	358.0
295.00	1614.0
296.00	45568.0
297.10	4678.0
302.00	733.0
303.10	3391.0
304.00	734.0
308.00	899.0
309.00	556.0
310.00	599.0
312.70	467.0
313.00	447.0
314.00	2240.0
315.00	6937.0
316.00	2843.0
316.90	467.0
321.00	1588.0
321.90	1054.0
323.10	14780.0
324.10	2740.0
326.90	3722.0
328.00	2356.0
329.10	319.0
331.90	1516.0
332.90	908.0
334.00	11612.0
335.00	2978.0
341.00	1766.0
342.10	635.0
346.00	3093.0
346.80	371.0
352.10	5113.0
353.10	2153.0
354.10	4031.0
355.10	1190.0
356.20	345.0
359.00	679.0
365.00	26016.0
365.90	3510.0
370.00	793.0
370.90	1264.0
372.10	8360.0
373.10	2010.0
373.90	421.0
383.00	2203.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

383.90	543.0
390.10	1091.0
391.20	496.0
401.00	511.0
402.10	2348.0
403.00	4430.0
404.20	973.0
415.00	455.0
421.10	4186.0
422.10	2785.0
423.10	20584.0
424.10	5576.0
424.90	1165.0
438.40	411.0
438.80	532.0
441.10	69552.0
442.10	437696.0
443.10	84128.0
444.00	7025.0
445.10	989.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

m/z	Abundance
36.00	319.0
40.10	759.0
43.10	519.0
44.00	895.0
48.90	300.0
50.20	386.0
55.10	583.0
63.90	450.0
66.00	625.0
73.10	1665.0
90.90	414.0
96.10	400.0
132.90	359.0
147.00	376.0
166.90	368.0
177.00	372.0
191.00	343.0
193.10	331.0
207.10	2271.0
208.20	1123.0
208.80	372.0
267.00	324.0
280.90	422.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36