

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021424\
 Data File : BM044089.D
 Acq On : 14 Feb 2024 19:02
 Operator : MA/JU
 Sample : P1453-04
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 TP-3

Quant Time: Feb 15 01:03:45 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM020824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 09 02:42:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.904	152	491659	20.000	ng	0.00
21) Naphthalene-d8	10.704	136	2111187	20.000	ng	-0.01
39) Acenaphthene-d10	14.545	164	1243682	20.000	ng	-0.01
64) Phenanthrene-d10	17.298	188	2460147	20.000	ng	0.00
76) Chrysene-d12	21.497	240	1549899	20.000	ng	0.00
86) Perylene-d12	23.891	264	1562424	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.481	112	5213976	154.241	ng	0.00
7) Phenol-d6	7.075	99	6114077	142.153	ng	0.00
23) Nitrobenzene-d5	9.075	82	4206550	105.730	ng	0.00
42) 2,4,6-Tribromophenol	16.039	330	1800472	133.708	ng	0.00
45) 2-Fluorobiphenyl	13.174	172	8355818	94.821	ng	0.00
79) Terphenyl-d14	19.939	244	9627434	108.892	ng	0.00
Target Compounds						
32) Benzoic acid	9.957	122	90051	3.701	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021424\
Data File : BM044089.D
Acq On : 14 Feb 2024 19:02
Operator : MA/JU
Sample : P1453-04
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
TP-3

Quant Time: Feb 15 01:03:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM020824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 09 02:42:01 2024
Response via : Initial Calibration

