

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM021521\
 Data File : BM028734.D
 Acq On : 15 Feb 2021 10:17
 Operator : CG/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4008

Quant Time: Feb 15 11:09:10 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM021221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 15 11:08:44 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.89	152	376	0.40	ng/ul	0.00
4) Naphthalene-d8	10.70	136	1519	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.54	164	815	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.28	188	1633	0.40	ng/ul	0.00
17) Chrysene-d12	21.42	240	1462	0.40	ng/ul	0.00
23) Perylene-d12	23.66	264	1424	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.21	96	178	0.40	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.29	152	922	0.39	ng/ul	0.00
18) Fluoranthene-d10	19.29	212	1735	0.41	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.24	88	176	0.401	ng/ul	89
5) Naphthalene	10.75	128	1678	0.399	ng/ul	99
7) 2-Methylnaphthalene	12.36	142	1148	0.399	ng/ul	99
8) 1-Methylnaphthalene	12.58	142	1096	0.399	ng/ul	100
10) Acenaphthylene	14.26	152	1400	0.387	ng/ul	98
11) Acenaphthene	14.60	153	1147	0.404	ng/ul	97
12) Fluorene	15.59	166	1290	0.399	ng/ul	99
14) Pentachlorophenol	16.94	266	130	0.290	ng/ul	98
15) Phenanthrene	17.32	178	2006	0.402	ng/ul	100
16) Anthracene	17.41	178	1802	0.391	ng/ul	98
19) Fluoranthene	19.32	202	2193	0.402	ng/ul	100
20) Pyrene	19.68	202	2233	0.400	ng/ul	98
21) Benzo(a)anthracene	21.41	228	2024	0.389	ng/ul	100
22) Chrysene	21.46	228	2162	0.402	ng/ul	100
24) Benzo(b)fluoranthene	22.98	252	2246	0.400	ng/ul	98
25) Benzo(k)fluoranthene	23.03	252	2235	0.404	ng/ul	99
26) Benzo(a)pyrene	23.56	252	1943	0.402	ng/ul	99
27) Indeno(1,2,3-cd)pyrene	25.91	276	2536	0.397	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.92	278	2105	0.394	ng/ul	99
29) Benzo(g,h,i)perylene	26.60	276	2209	0.411	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM021521\
Data File : BM028734.D
Acq On : 15 Feb 2021 10:17
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4008

Quant Time: Feb 15 11:09:10 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM021221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 15 11:08:44 2021
Response via : Initial Calibration

