

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021524\
 Data File : BM044095.D
 Acq On : 15 Feb 2024 11:05
 Operator : MA/JU
 Sample : SSTD0.405
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4010

Quant Time: Feb 15 12:09:48 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 15 12:08:54 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.905	152	5023	0.400	ng/ul	0.00
4) Naphthalene-d8	10.704	136	15359	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.547	164	8097	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.300	188	19454	0.400	ng/ul	0.00
17) Chrysene-d12	21.500	240	17032	0.400	ng/ul	0.00
23) Perylene-d12	23.894	264	20349	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.356	96	3188	0.464	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.299	152	9091	0.432	ng/ul	0.00
18) Fluoranthene-d10	19.331	212	21148	0.388	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.390	88	3007	0.427	ng/ul	100
5) Naphthalene	10.754	128	19344	0.424	ng/ul	100
7) 2-Methylnaphthalene	12.371	142	11931	0.435	ng/ul	100
8) 1-Methylnaphthalene	12.591	142	12083	0.442	ng/ul	100
10) Acenaphthylene	14.270	152	18757	0.414	ng/ul	100
11) Acenaphthene	14.607	153	13610	0.429	ng/ul	100
12) Fluorene	15.597	166	15457	0.440	ng/ul	100
14) Pentachlorophenol	16.946	266	2811	0.534	ng/ul	100
15) Phenanthrene	17.338	178	25615	0.417	ng/ul	100
16) Anthracene	17.436	178	25505	0.432	ng/ul	100
19) Fluoranthene	19.363	202	32358	0.402	ng/ul	100
20) Pyrene	19.726	202	34014	0.409	ng/ul	100
21) Benzo(a)anthracene	21.485	228	32414	0.433	ng/ul	100
22) Chrysene	21.538	228	35672	0.451	ng/ul	100
24) Benzo(b)fluoranthene	23.163	252	35931	0.440	ng/ul	100
25) Benzo(k)fluoranthene	23.212	252	38859	0.453	ng/ul	100
26) Benzo(a)pyrene	23.785	252	33161	0.444	ng/ul	100
27) Indeno(1,2,3-cd)pyrene	26.370	276	41506	0.433	ng/ul#	100
28) Dibenzo(a,h)anthracene	26.410	278	32337	0.425	ng/ul	100
29) Benzo(g,h,i)perylene	27.128	276	36933	0.445	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021524\
Data File : BM044095.D
Acq On : 15 Feb 2024 11:05
Operator : MA/JU
Sample : SST0.405
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4010

Quant Time: Feb 15 12:09:48 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 15 12:08:54 2024
Response via : Initial Calibration

