

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021524\
 Data File : BM044128.D
 Acq On : 16 Feb 2024 08:29
 Operator : MA/JU
 Sample : P1305-03
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0FP4

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/17/2024
 Supervised By :mohammad ahmed 02/19/2024

Quant Time: Feb 16 08:59:56 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 15 13:40:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.900	152	10070	0.400	ng/u1	0.00
4) Naphthalene-d8	10.699	136	34068	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.543	164	18093	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.296	188	37648m	0.400	ng/u1	0.00
17) Chrysene-d12	21.497	240	24186m	0.400	ng/u1	0.00
23) Perylene-d12	23.888	264	27217	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	86843	6.090	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.294	152	18604	0.390	ng/u1	0.00
18) Fluoranthene-d10	19.326	212	33312	0.462	ng/u1	0.00
Target Compounds						Qvalue
14) Pentachlorophenol	16.937	266	406	0.030	ng/u1	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021524\
Data File : BM044128.D
Acq On : 16 Feb 2024 08:29
Operator : MA/JU
Sample : P1305-03
Misc :
ALS Vial : 38 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
C0FP4

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/17/2024
Supervised By :mohammad ahmed 02/19/2024


Quant Time: Feb 16 08:59:56 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 15 13:40:24 2024
Response via : Initial Calibration





#14
 Pentachlorophenol
 Concen: 0.030 ng/ul
 RT: 16.937 min Scan# 3045
 Delta R.T. -0.008 min
 Lab File: BM044128.D
 Acq: 16 Feb 2024 08:29

Instrument :
 BNA_M
 ClientSampleId :
 C0FP4



Tgt Ion:266 Resp: 400
 Ion Ratio Lower Upper
 266 100
 264 62.6 49.4 74.0
 268 68.2 52.1 78.1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/17/2024
 Supervised By :mohammad ahmed 02/19/2024

