

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021524\
 Data File : BM044157.D
 Acq On : 17 Feb 2024 03:16
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 69 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4189

Quant Time: Feb 17 04:20:41 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 15 13:40:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.900	152	7143	0.400	ng/ul	0.00
4) Naphthalene-d8	10.699	136	22459	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.543	164	11606	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.293	188	23920	0.400	ng/ul	0.00
17) Chrysene-d12	21.496	240	13407	0.400	ng/ul	0.00
23) Perylene-d12	23.887	264	12173	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.356	96	4782	0.473	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.294	152	14640	0.466	ng/ul	0.00
18) Fluoranthene-d10	19.327	212	24863	0.622	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.390	88	4725	0.467	ng/ul#	89
5) Naphthalene	10.748	128	31190	0.467	ng/ul	98
7) 2-Methylnaphthalene	12.365	142	19433	0.472	ng/ul	100
8) 1-Methylnaphthalene	12.585	142	19675	0.472	ng/ul	100
10) Acenaphthylene	14.261	152	30841	0.484	ng/ul	99
11) Acenaphthene	14.604	153	21709	0.469	ng/ul	99
12) Fluorene	15.594	166	23802	0.457	ng/ul	99
14) Pentachlorophenol	16.943	266	2897	0.338	ng/ul	99
15) Phenanthrene	17.335	178	35491	0.468	ng/ul	99
16) Anthracene	17.432	178	34829	0.464	ng/ul	100
19) Fluoranthene	19.360	202	37914	0.620	ng/ul	99
20) Pyrene	19.722	202	38475	0.599	ng/ul	97
21) Benzo(a)anthracene	21.481	228	27988	0.456	ng/ul	100
22) Chrysene	21.534	228	30689	0.464	ng/ul	100
24) Benzo(b)fluoranthene	23.159	252	24396	0.486	ng/ul	99
25) Benzo(k)fluoranthene	23.209	252	26452	0.485	ng/ul	98
26) Benzo(a)pyrene	23.782	252	21735	0.465	ng/ul	99
27) Indeno(1,2,3-cd)pyrene	26.365	276	24523	0.412	ng/ul#	100
28) Dibenzo(a,h)anthracene	26.402	278	19391	0.414	ng/ul	99
29) Benzo(g,h,i)perylene	27.123	276	21410	0.407	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021524\
Data File : BM044157.D
Acq On : 17 Feb 2024 03:16
Operator : MA/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 69 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4189

Quant Time: Feb 17 04:20:41 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 15 13:40:24 2024
Response via : Initial Calibration

