



Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM021524\  
Data File : BM044092.D  
Acq On : 15 Feb 2024 09:16  
Operator : MA/JU  
Sample : DFTPP002  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP002

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\SFAM-EPA-SIM-BM021524.M  
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
Last Update : Thu Feb 15 13:40:24 2024



AutoFind: Scans 2491, 2492, 2493; Background Corrected with Scan 2483

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 38.0         | 115583     | PASS                |
| 68             | 69              | 0.00            | 2               | 1.7          | 1956       | PASS                |
| 69             | 198             | 0.00            | 100             | 38.1         | 115845     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.5          | 614        | PASS                |
| 127            | 198             | 10              | 80              | 48.3         | 146880     | PASS                |
| 197            | 198             | 0.00            | 2               | 0.2          | 669        | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 304107     | PASS                |
| 199            | 198             | 5               | 9               | 6.7          | 20262      | PASS                |
| 275            | 198             | 10              | 60              | 22.1         | 67195      | PASS                |
| 365            | 198             | 1               | 100             | 2.7          | 8224       | PASS                |
| 441            | 198             | 0.01            | 100             | 10.8         | 32779      | PASS                |
| 442            | 198             | 50              | 100             | 67.9         | 206528     | PASS                |
| 443            | 442             | 15              | 24              | 19.4         | 40061      | PASS                |

| m/z    | Abundance |
|--------|-----------|
| 36.00  | 164.0     |
| 36.90  | 191.0     |
| 37.90  | 206.0     |
| 39.00  | 635.0     |
| 40.00  | 3019.0    |
| 41.00  | 727.0     |
| 42.10  | 481.0     |
| 43.00  | 996.0     |
| 44.00  | 9885.0    |
| 44.90  | 604.0     |
| 46.90  | 253.0     |
| 49.00  | 803.0     |
| 49.90  | 166.0     |
| 51.00  | 553.0     |
| 52.00  | 320.0     |
| 55.10  | 549.0     |
| 56.00  | 279.0     |
| 57.00  | 641.0     |
| 58.00  | 157.0     |
| 59.90  | 204.0     |
| 62.90  | 300.0     |
| 65.00  | 246.0     |
| 69.10  | 267.0     |
| 69.90  | 177.0     |
| 73.00  | 1239.0    |
| 73.90  | 195.0     |
| 75.00  | 262.0     |
| 76.90  | 485.0     |
| 78.10  | 596.0     |
| 79.10  | 166.0     |
| 81.00  | 440.0     |
| 81.90  | 193.0     |
| 83.90  | 771.0     |
| 85.10  | 171.0     |
| 85.80  | 449.0     |
| 91.00  | 362.0     |
| 92.80  | 163.0     |
| 95.10  | 160.0     |
| 95.80  | 228.0     |
| 96.10  | 230.0     |
| 97.10  | 209.0     |
| 103.00 | 222.0     |
| 105.00 | 201.0     |
| 115.10 | 172.0     |
| 118.80 | 158.0     |
| 132.80 | 438.0     |
| 134.90 | 270.0     |
| 146.90 | 327.0     |
| 191.00 | 277.0     |
| 193.00 | 334.0     |
| 207.00 | 2452.0    |
| 208.00 | 553.0     |
| 209.10 | 500.0     |
| 220.90 | 186.0     |
| 266.90 | 239.0     |
| 280.90 | 537.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP002

| m/z    | Abundance |
|--------|-----------|
| 36.00  | 272.0     |
| 37.00  | 686.0     |
| 38.10  | 1641.0    |
| 39.10  | 10722.0   |
| 40.00  | 4233.0    |
| 41.00  | 857.0     |
| 42.00  | 514.0     |
| 43.00  | 871.0     |
| 44.00  | 10535.0   |
| 45.00  | 831.0     |
| 47.00  | 385.0     |
| 47.90  | 233.0     |
| 49.00  | 1677.0    |
| 50.00  | 31512.0   |
| 51.00  | 136576.0  |
| 52.00  | 7196.0    |
| 53.00  | 360.0     |
| 54.00  | 157.0     |
| 55.10  | 1131.0    |
| 56.10  | 4137.0    |
| 57.00  | 9804.0    |
| 58.00  | 519.0     |
| 59.00  | 344.0     |
| 60.00  | 274.0     |
| 61.00  | 1661.0    |
| 62.00  | 1841.0    |
| 63.00  | 5349.0    |
| 64.00  | 818.0     |
| 65.00  | 2718.0    |
| 65.90  | 257.0     |
| 67.10  | 349.0     |
| 68.00  | 2290.0    |
| 69.00  | 138624.0  |
| 70.00  | 1047.0    |
| 71.00  | 366.0     |
| 73.00  | 2302.0    |
| 74.00  | 13500.0   |
| 75.00  | 22480.0   |
| 76.00  | 9014.0    |
| 77.00  | 160320.0  |
| 78.00  | 11928.0   |
| 79.00  | 8698.0    |
| 80.00  | 6668.0    |
| 81.00  | 11541.0   |
| 81.90  | 2771.0    |
| 83.00  | 2739.0    |
| 84.00  | 1003.0    |
| 85.00  | 1740.0    |
| 86.00  | 3255.0    |
| 87.00  | 1211.0    |
| 88.00  | 538.0     |
| 89.00  | 255.0     |
| 91.00  | 2221.0    |
| 92.00  | 3039.0    |
| 93.00  | 16165.0   |
| 93.90  | 989.0     |
| 95.00  | 494.0     |
| 96.00  | 869.0     |
| 97.20  | 427.0     |
| 98.00  | 11606.0   |
| 99.00  | 11845.0   |
| 100.00 | 1027.0    |
| 101.00 | 6393.0    |
| 102.00 | 384.0     |
| 103.00 | 1876.0    |
| 104.00 | 2858.0    |
| 104.90 | 3133.0    |
| 106.00 | 1512.0    |
| 107.00 | 44992.0   |
| 108.00 | 6974.0    |
| 109.00 | 1275.0    |
| 110.00 | 87800.0   |
| 111.00 | 11997.0   |
| 112.00 | 1793.0    |
| 113.00 | 467.0     |
| 114.90 | 219.0     |
| 116.00 | 2514.0    |
| 117.00 | 33944.0   |
| 118.00 | 2628.0    |
| 118.90 | 633.0     |
| 120.00 | 586.0     |
| 120.90 | 392.0     |
| 122.00 | 2872.0    |
| 123.00 | 4529.0    |
| 123.90 | 2070.0    |
| 125.00 | 2015.0    |
| 126.10 | 563.0     |
| 127.00 | 163200.0  |
| 128.00 | 12099.0   |
| 129.00 | 59568.0   |
| 130.00 | 5002.0    |
| 131.00 | 1235.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP002

|        |          |
|--------|----------|
| 131.90 | 368.0    |
| 133.00 | 587.0    |
| 134.00 | 1508.0   |
| 135.00 | 4468.0   |
| 136.00 | 1925.0   |
| 137.00 | 3718.0   |
| 137.90 | 574.0    |
| 138.90 | 178.0    |
| 139.90 | 816.0    |
| 140.90 | 6549.0   |
| 142.00 | 2711.0   |
| 143.00 | 1672.0   |
| 144.10 | 646.0    |
| 145.10 | 414.0    |
| 146.00 | 1096.0   |
| 146.90 | 3895.0   |
| 148.00 | 8772.0   |
| 149.00 | 3010.0   |
| 150.00 | 619.0    |
| 151.00 | 1096.0   |
| 151.80 | 607.0    |
| 152.90 | 2090.0   |
| 154.00 | 1718.0   |
| 155.00 | 3894.0   |
| 156.00 | 6359.0   |
| 157.00 | 1066.0   |
| 157.90 | 1465.0   |
| 158.90 | 1071.0   |
| 160.00 | 2006.0   |
| 161.00 | 3415.0   |
| 161.90 | 855.0    |
| 163.00 | 299.0    |
| 163.90 | 393.0    |
| 165.00 | 2757.0   |
| 166.00 | 2103.0   |
| 167.00 | 14844.0  |
| 168.00 | 12678.0  |
| 168.90 | 1908.0   |
| 170.00 | 518.0    |
| 170.80 | 466.0    |
| 171.90 | 1182.0   |
| 172.90 | 1309.0   |
| 174.00 | 3085.0   |
| 175.00 | 6017.0   |
| 176.00 | 1164.0   |
| 176.90 | 2459.0   |
| 178.00 | 923.0    |
| 178.90 | 10285.0  |
| 180.00 | 7760.0   |
| 180.90 | 3485.0   |
| 181.90 | 531.0    |
| 183.10 | 308.0    |
| 183.90 | 948.0    |
| 184.90 | 4680.0   |
| 186.00 | 40504.0  |
| 187.00 | 11759.0  |
| 188.00 | 1371.0   |
| 188.90 | 2288.0   |
| 190.00 | 521.0    |
| 190.90 | 1676.0   |
| 192.00 | 3599.0   |
| 193.00 | 3677.0   |
| 194.00 | 618.0    |
| 195.10 | 419.0    |
| 195.90 | 7325.0   |
| 196.80 | 1347.0   |
| 197.90 | 313152.0 |
| 198.90 | 22768.0  |
| 199.90 | 1596.0   |
| 201.40 | 1015.0   |
| 203.00 | 1876.0   |
| 204.00 | 9917.0   |
| 205.00 | 18040.0  |
| 206.00 | 74728.0  |
| 207.00 | 12158.0  |
| 208.00 | 3399.0   |
| 209.00 | 1693.0   |
| 210.00 | 1208.0   |
| 211.00 | 3113.0   |
| 211.90 | 574.0    |
| 212.90 | 189.0    |
| 214.90 | 493.0    |
| 216.00 | 1654.0   |
| 216.90 | 18936.0  |
| 217.90 | 1948.0   |
| 218.90 | 321.0    |
| 221.00 | 10227.0  |
| 221.80 | 2521.0   |
| 223.00 | 4467.0   |
| 224.00 | 40128.0  |
| 225.00 | 11019.0  |
| 226.00 | 1178.0   |
| 227.00 | 16608.0  |
| 228.00 | 2169.0   |
| 228.90 | 3281.0   |
| 230.00 | 575.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP002

|        |          |
|--------|----------|
| 231.00 | 1753.0   |
| 231.90 | 222.0    |
| 232.90 | 266.0    |
| 234.00 | 1171.0   |
| 234.90 | 1102.0   |
| 235.80 | 701.0    |
| 237.00 | 1915.0   |
| 238.00 | 335.0    |
| 238.90 | 564.0    |
| 239.90 | 382.0    |
| 241.00 | 1087.0   |
| 242.00 | 2149.0   |
| 243.00 | 2266.0   |
| 244.00 | 31552.0  |
| 245.00 | 4177.0   |
| 246.00 | 5246.0   |
| 247.00 | 1184.0   |
| 247.90 | 173.0    |
| 248.90 | 1298.0   |
| 249.90 | 200.0    |
| 250.90 | 365.0    |
| 251.80 | 204.0    |
| 253.00 | 745.0    |
| 254.00 | 811.0    |
| 255.00 | 152384.0 |
| 256.00 | 23504.0  |
| 257.00 | 2135.0   |
| 258.00 | 8612.0   |
| 258.90 | 1194.0   |
| 260.00 | 383.0    |
| 261.00 | 191.0    |
| 263.00 | 160.0    |
| 264.00 | 249.0    |
| 264.90 | 3260.0   |
| 265.90 | 481.0    |
| 266.90 | 220.0    |
| 270.90 | 376.0    |
| 272.00 | 438.0    |
| 273.00 | 4837.0   |
| 274.00 | 11399.0  |
| 275.00 | 65272.0  |
| 276.00 | 9328.0   |
| 277.00 | 5089.0   |
| 277.90 | 1056.0   |
| 278.90 | 236.0    |
| 280.90 | 889.0    |
| 282.00 | 450.0    |
| 283.00 | 723.0    |
| 284.00 | 492.0    |
| 285.00 | 1006.0   |
| 285.90 | 202.0    |
| 289.00 | 223.0    |
| 289.90 | 370.0    |
| 290.80 | 193.0    |
| 291.90 | 243.0    |
| 293.00 | 1424.0   |
| 294.00 | 542.0    |
| 294.90 | 228.0    |
| 296.00 | 16992.0  |
| 296.90 | 2706.0   |
| 300.90 | 309.0    |
| 302.10 | 365.0    |
| 302.90 | 1878.0   |
| 303.90 | 623.0    |
| 305.10 | 154.0    |
| 307.90 | 222.0    |
| 308.90 | 194.0    |
| 309.90 | 260.0    |
| 314.00 | 734.0    |
| 315.00 | 1579.0   |
| 315.90 | 1096.0   |
| 317.00 | 174.0    |
| 320.90 | 536.0    |
| 322.00 | 240.0    |
| 323.00 | 6262.0   |
| 324.00 | 769.0    |
| 324.90 | 188.0    |
| 326.90 | 908.0    |
| 328.00 | 424.0    |
| 332.00 | 477.0    |
| 333.00 | 403.0    |
| 334.00 | 3781.0   |
| 335.10 | 818.0    |
| 340.90 | 693.0    |
| 342.00 | 209.0    |
| 345.90 | 1265.0   |
| 346.80 | 264.0    |
| 352.00 | 1693.0   |
| 352.90 | 1139.0   |
| 354.00 | 1660.0   |
| 355.00 | 420.0    |
| 365.00 | 6738.0   |
| 365.90 | 1080.0   |
| 370.90 | 367.0    |
| 372.00 | 2468.0   |
| 373.00 | 726.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP002

|        |          |
|--------|----------|
| 383.00 | 668.0    |
| 383.90 | 256.0    |
| 390.00 | 485.0    |
| 391.00 | 270.0    |
| 391.90 | 180.0    |
| 402.00 | 1163.0   |
| 402.90 | 1552.0   |
| 404.00 | 519.0    |
| 405.00 | 182.0    |
| 420.90 | 1378.0   |
| 422.00 | 1386.0   |
| 423.00 | 9370.0   |
| 424.00 | 2816.0   |
| 424.90 | 541.0    |
| 438.10 | 156.0    |
| 441.00 | 26896.0  |
| 442.00 | 161792.0 |
| 443.00 | 31888.0  |
| 444.00 | 3310.0   |
| 445.00 | 165.0    |

**Instrument :**  
BNA\_M  
**ClientSampleId :**  
DFTPP002

| m/z    | Abundance |
|--------|-----------|
| 35.90  | 339.0     |
| 37.00  | 562.0     |
| 38.00  | 1295.0    |
| 39.10  | 9800.0    |
| 40.00  | 4113.0    |
| 41.00  | 903.0     |
| 41.90  | 316.0     |
| 43.00  | 1146.0    |
| 44.00  | 10065.0   |
| 44.90  | 825.0     |
| 46.90  | 368.0     |
| 47.90  | 200.0     |
| 49.00  | 1724.0    |
| 50.00  | 33944.0   |
| 51.00  | 135872.0  |
| 52.00  | 6347.0    |
| 52.90  | 353.0     |
| 54.10  | 227.0     |
| 55.00  | 1299.0    |
| 56.00  | 4026.0    |
| 57.00  | 9799.0    |
| 58.00  | 496.0     |
| 59.10  | 227.0     |
| 60.00  | 288.0     |
| 61.00  | 1644.0    |
| 62.00  | 1918.0    |
| 63.00  | 6003.0    |
| 64.00  | 753.0     |
| 65.00  | 2467.0    |
| 65.90  | 181.0     |
| 67.20  | 241.0     |
| 68.00  | 2179.0    |
| 69.00  | 133824.0  |
| 70.00  | 776.0     |
| 70.90  | 325.0     |
| 73.00  | 1909.0    |
| 74.00  | 13119.0   |
| 75.00  | 21872.0   |
| 76.10  | 6964.0    |
| 77.00  | 161792.0  |
| 78.00  | 10985.0   |
| 79.00  | 8707.0    |
| 80.00  | 7137.0    |
| 81.00  | 11271.0   |
| 82.00  | 2550.0    |
| 83.00  | 2919.0    |
| 84.00  | 1138.0    |
| 85.00  | 1947.0    |
| 86.00  | 3693.0    |
| 87.00  | 1322.0    |
| 88.00  | 719.0     |
| 89.00  | 406.0     |
| 91.00  | 2585.0    |
| 92.00  | 2995.0    |
| 93.00  | 16440.0   |
| 93.90  | 1235.0    |
| 95.00  | 480.0     |
| 96.00  | 1415.0    |
| 97.00  | 365.0     |
| 98.00  | 11143.0   |
| 99.00  | 11891.0   |
| 99.90  | 1020.0    |
| 101.00 | 6081.0    |
| 102.00 | 440.0     |
| 102.90 | 1794.0    |
| 104.00 | 3858.0    |
| 105.00 | 3865.0    |
| 106.00 | 1278.0    |
| 107.00 | 47984.0   |
| 108.00 | 7293.0    |
| 108.90 | 667.0     |
| 110.00 | 89576.0   |
| 111.00 | 14077.0   |
| 111.90 | 1681.0    |
| 113.00 | 521.0     |
| 113.90 | 171.0     |
| 114.90 | 260.0     |
| 115.90 | 2721.0    |
| 117.00 | 34664.0   |
| 118.00 | 2832.0    |
| 118.90 | 465.0     |
| 119.90 | 501.0     |
| 120.90 | 285.0     |
| 122.00 | 2537.0    |
| 123.00 | 5231.0    |
| 124.00 | 1965.0    |
| 125.00 | 2175.0    |
| 126.10 | 800.0     |
| 127.00 | 174784.0  |
| 128.00 | 13485.0   |
| 129.00 | 63464.0   |
| 130.00 | 5248.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP002

|        |          |
|--------|----------|
| 131.00 | 1056.0   |
| 132.10 | 386.0    |
| 132.80 | 424.0    |
| 133.90 | 1713.0   |
| 134.90 | 5237.0   |
| 135.90 | 2178.0   |
| 137.00 | 2854.0   |
| 137.90 | 523.0    |
| 138.80 | 256.0    |
| 139.90 | 732.0    |
| 141.00 | 7116.0   |
| 142.00 | 2736.0   |
| 142.90 | 1927.0   |
| 143.90 | 570.0    |
| 145.00 | 882.0    |
| 145.90 | 1322.0   |
| 147.00 | 4121.0   |
| 148.00 | 9026.0   |
| 148.90 | 1994.0   |
| 150.00 | 792.0    |
| 151.00 | 872.0    |
| 152.00 | 434.0    |
| 152.90 | 2238.0   |
| 153.90 | 1909.0   |
| 155.00 | 4706.0   |
| 156.00 | 6374.0   |
| 157.00 | 1542.0   |
| 158.00 | 1402.0   |
| 158.90 | 1368.0   |
| 159.90 | 2323.0   |
| 161.00 | 3516.0   |
| 161.90 | 959.0    |
| 162.90 | 303.0    |
| 163.90 | 454.0    |
| 164.90 | 3357.0   |
| 166.00 | 2507.0   |
| 167.00 | 17416.0  |
| 168.00 | 11978.0  |
| 169.00 | 1832.0   |
| 169.90 | 324.0    |
| 170.40 | 274.0    |
| 171.00 | 608.0    |
| 171.90 | 1336.0   |
| 173.00 | 2257.0   |
| 174.00 | 3451.0   |
| 175.00 | 6358.0   |
| 176.00 | 1648.0   |
| 176.90 | 2799.0   |
| 177.90 | 1013.0   |
| 179.00 | 11788.0  |
| 180.00 | 9367.0   |
| 181.00 | 4505.0   |
| 182.00 | 656.0    |
| 183.10 | 477.0    |
| 183.90 | 1068.0   |
| 185.00 | 5716.0   |
| 186.00 | 46720.0  |
| 187.00 | 13136.0  |
| 187.90 | 1664.0   |
| 189.00 | 2290.0   |
| 189.90 | 354.0    |
| 191.00 | 1222.0   |
| 192.00 | 3617.0   |
| 193.00 | 4127.0   |
| 194.00 | 1123.0   |
| 194.90 | 785.0    |
| 196.00 | 8223.0   |
| 197.00 | 1027.0   |
| 197.90 | 364736.0 |
| 198.90 | 23736.0  |
| 199.90 | 1386.0   |
| 201.40 | 1559.0   |
| 202.90 | 2226.0   |
| 204.00 | 11902.0  |
| 205.00 | 20736.0  |
| 206.00 | 86408.0  |
| 207.00 | 14036.0  |
| 208.00 | 3222.0   |
| 209.00 | 1501.0   |
| 210.20 | 1600.0   |
| 211.00 | 3304.0   |
| 211.90 | 645.0    |
| 213.00 | 194.0    |
| 214.90 | 648.0    |
| 216.00 | 1660.0   |
| 217.00 | 22224.0  |
| 218.00 | 2946.0   |
| 219.00 | 224.0    |
| 221.00 | 12056.0  |
| 221.80 | 2945.0   |
| 223.00 | 4532.0   |
| 224.00 | 47488.0  |
| 225.00 | 12112.0  |
| 226.00 | 1402.0   |
| 227.00 | 18184.0  |
| 228.00 | 2632.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP002

|        |          |
|--------|----------|
| 228.90 | 3556.0   |
| 230.00 | 508.0    |
| 231.00 | 1887.0   |
| 232.00 | 354.0    |
| 232.90 | 382.0    |
| 233.90 | 1550.0   |
| 234.90 | 1171.0   |
| 235.90 | 956.0    |
| 236.90 | 1858.0   |
| 237.90 | 352.0    |
| 238.90 | 843.0    |
| 239.90 | 580.0    |
| 240.90 | 1392.0   |
| 242.00 | 2328.0   |
| 243.00 | 2189.0   |
| 244.00 | 38672.0  |
| 245.00 | 4800.0   |
| 246.00 | 6242.0   |
| 246.90 | 1232.0   |
| 247.80 | 414.0    |
| 248.90 | 988.0    |
| 249.90 | 259.0    |
| 250.90 | 457.0    |
| 251.90 | 373.0    |
| 253.00 | 850.0    |
| 254.00 | 845.0    |
| 255.00 | 188736.0 |
| 256.00 | 27968.0  |
| 257.00 | 2326.0   |
| 258.00 | 10118.0  |
| 259.00 | 1585.0   |
| 260.00 | 365.0    |
| 260.90 | 255.0    |
| 264.00 | 438.0    |
| 265.00 | 3629.0   |
| 265.90 | 788.0    |
| 267.00 | 233.0    |
| 269.90 | 258.0    |
| 270.90 | 694.0    |
| 271.90 | 382.0    |
| 273.00 | 5886.0   |
| 274.00 | 14700.0  |
| 275.00 | 80792.0  |
| 276.00 | 10520.0  |
| 277.00 | 6734.0   |
| 278.00 | 934.0    |
| 278.90 | 391.0    |
| 280.90 | 521.0    |
| 281.90 | 273.0    |
| 283.00 | 799.0    |
| 283.90 | 460.0    |
| 285.00 | 1082.0   |
| 289.00 | 336.0    |
| 290.00 | 329.0    |
| 290.80 | 218.0    |
| 292.00 | 424.0    |
| 293.00 | 1797.0   |
| 294.00 | 619.0    |
| 295.00 | 497.0    |
| 296.00 | 21744.0  |
| 297.00 | 2800.0   |
| 298.00 | 400.0    |
| 300.90 | 402.0    |
| 301.90 | 193.0    |
| 303.00 | 2253.0   |
| 303.90 | 741.0    |
| 308.00 | 394.0    |
| 308.90 | 168.0    |
| 310.00 | 370.0    |
| 314.00 | 967.0    |
| 315.00 | 2454.0   |
| 315.90 | 1475.0   |
| 317.00 | 215.0    |
| 320.80 | 667.0    |
| 322.00 | 268.0    |
| 323.00 | 7182.0   |
| 324.00 | 1050.0   |
| 325.00 | 190.0    |
| 325.80 | 164.0    |
| 327.00 | 1419.0   |
| 327.90 | 699.0    |
| 329.10 | 210.0    |
| 331.90 | 543.0    |
| 332.90 | 687.0    |
| 334.00 | 4850.0   |
| 335.00 | 1213.0   |
| 335.90 | 267.0    |
| 341.00 | 801.0    |
| 342.00 | 218.0    |
| 345.90 | 1505.0   |
| 346.80 | 401.0    |
| 352.00 | 2110.0   |
| 353.00 | 1312.0   |
| 354.00 | 2631.0   |
| 355.00 | 457.0    |
| 358.90 | 217.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP002

|        |          |
|--------|----------|
| 365.00 | 10183.0  |
| 366.00 | 1680.0   |
| 369.90 | 307.0    |
| 370.90 | 572.0    |
| 372.00 | 3832.0   |
| 373.00 | 1164.0   |
| 382.90 | 1282.0   |
| 383.90 | 397.0    |
| 390.00 | 372.0    |
| 391.00 | 357.0    |
| 391.80 | 236.0    |
| 401.00 | 217.0    |
| 402.00 | 1459.0   |
| 403.00 | 2081.0   |
| 404.00 | 845.0    |
| 405.00 | 188.0    |
| 421.00 | 1977.0   |
| 422.00 | 1721.0   |
| 423.00 | 12674.0  |
| 424.10 | 3946.0   |
| 425.00 | 595.0    |
| 441.00 | 38704.0  |
| 442.00 | 249792.0 |
| 443.00 | 50168.0  |
| 444.00 | 3914.0   |
| 445.10 | 220.0    |

**Instrument :**  
BNA\_M  
**ClientSampleId :**  
DFTPP002

| m/z    | Abundance |
|--------|-----------|
| 36.00  | 191.0     |
| 37.00  | 339.0     |
| 38.00  | 912.0     |
| 39.00  | 5521.0    |
| 40.00  | 3721.0    |
| 40.90  | 807.0     |
| 42.00  | 272.0     |
| 43.00  | 1215.0    |
| 44.00  | 10655.0   |
| 44.90  | 689.0     |
| 47.00  | 257.0     |
| 48.00  | 195.0     |
| 49.00  | 1389.0    |
| 50.00  | 20984.0   |
| 51.00  | 75960.0   |
| 52.00  | 4501.0    |
| 52.90  | 361.0     |
| 54.10  | 211.0     |
| 55.00  | 1100.0    |
| 56.00  | 2288.0    |
| 57.00  | 5785.0    |
| 57.90  | 354.0     |
| 59.00  | 323.0     |
| 59.90  | 294.0     |
| 61.00  | 1103.0    |
| 62.00  | 1250.0    |
| 63.00  | 3222.0    |
| 63.90  | 546.0     |
| 65.00  | 1502.0    |
| 65.90  | 231.0     |
| 67.00  | 432.0     |
| 68.00  | 1400.0    |
| 69.00  | 75888.0   |
| 70.00  | 551.0     |
| 70.90  | 299.0     |
| 73.00  | 1589.0    |
| 74.00  | 7593.0    |
| 75.00  | 12329.0   |
| 76.00  | 3905.0    |
| 77.00  | 90856.0   |
| 78.00  | 7113.0    |
| 79.00  | 5440.0    |
| 80.00  | 4270.0    |
| 81.00  | 6249.0    |
| 82.00  | 1245.0    |
| 83.00  | 1405.0    |
| 83.90  | 887.0     |
| 85.00  | 1250.0    |
| 85.90  | 2185.0    |
| 87.00  | 1008.0    |
| 87.90  | 348.0     |
| 89.00  | 365.0     |
| 91.00  | 1528.0    |
| 92.00  | 1429.0    |
| 93.00  | 9919.0    |
| 94.00  | 832.0     |
| 94.80  | 252.0     |
| 95.90  | 741.0     |
| 96.90  | 232.0     |
| 98.00  | 7574.0    |
| 99.00  | 7003.0    |
| 99.90  | 684.0     |
| 100.90 | 3134.0    |
| 102.00 | 289.0     |
| 102.90 | 1148.0    |
| 104.00 | 2122.0    |
| 105.00 | 2469.0    |
| 106.10 | 790.0     |
| 107.00 | 28752.0   |
| 108.00 | 4121.0    |
| 108.90 | 657.0     |
| 110.00 | 51760.0   |
| 111.00 | 7654.0    |
| 111.90 | 688.0     |
| 113.00 | 433.0     |
| 114.90 | 297.0     |
| 116.10 | 1276.0    |
| 117.00 | 19368.0   |
| 117.90 | 1761.0    |
| 119.10 | 451.0     |
| 119.90 | 302.0     |
| 121.00 | 263.0     |
| 122.00 | 1669.0    |
| 123.00 | 2897.0    |
| 123.90 | 1527.0    |
| 125.00 | 1121.0    |
| 126.00 | 283.0     |
| 127.00 | 102656.0  |
| 128.00 | 7234.0    |
| 129.00 | 36816.0   |
| 129.90 | 3206.0    |
| 130.90 | 677.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP002

|        |          |
|--------|----------|
| 131.80 | 340.0    |
| 132.70 | 292.0    |
| 133.10 | 232.0    |
| 133.80 | 1112.0   |
| 135.00 | 3778.0   |
| 135.90 | 1294.0   |
| 137.00 | 1990.0   |
| 138.10 | 420.0    |
| 140.00 | 476.0    |
| 140.90 | 4285.0   |
| 141.90 | 1466.0   |
| 143.10 | 925.0    |
| 143.90 | 363.0    |
| 144.90 | 286.0    |
| 146.00 | 838.0    |
| 146.90 | 2843.0   |
| 148.00 | 5292.0   |
| 148.90 | 1436.0   |
| 149.90 | 349.0    |
| 151.00 | 661.0    |
| 151.90 | 404.0    |
| 153.00 | 1438.0   |
| 154.00 | 1341.0   |
| 155.00 | 2143.0   |
| 156.00 | 4316.0   |
| 156.90 | 730.0    |
| 157.90 | 846.0    |
| 159.00 | 638.0    |
| 159.90 | 1509.0   |
| 161.00 | 2396.0   |
| 162.00 | 565.0    |
| 162.90 | 494.0    |
| 163.80 | 183.0    |
| 164.90 | 2028.0   |
| 166.00 | 1036.0   |
| 167.00 | 10769.0  |
| 168.00 | 7738.0   |
| 169.00 | 937.0    |
| 169.80 | 183.0    |
| 170.40 | 212.0    |
| 171.00 | 329.0    |
| 172.00 | 888.0    |
| 173.00 | 1202.0   |
| 174.00 | 2004.0   |
| 175.00 | 4292.0   |
| 175.90 | 1289.0   |
| 176.90 | 1915.0   |
| 177.80 | 613.0    |
| 178.90 | 7067.0   |
| 180.00 | 5981.0   |
| 180.90 | 2527.0   |
| 182.00 | 403.0    |
| 183.00 | 185.0    |
| 183.90 | 530.0    |
| 185.00 | 3440.0   |
| 186.00 | 27288.0  |
| 187.00 | 8512.0   |
| 188.00 | 894.0    |
| 188.90 | 1898.0   |
| 189.90 | 456.0    |
| 191.00 | 1159.0   |
| 191.90 | 1981.0   |
| 193.00 | 3245.0   |
| 194.00 | 642.0    |
| 194.90 | 313.0    |
| 195.90 | 5503.0   |
| 197.00 | 981.0    |
| 197.90 | 234432.0 |
| 198.90 | 14281.0  |
| 200.00 | 1055.0   |
| 201.40 | 725.0    |
| 202.90 | 1366.0   |
| 204.00 | 7625.0   |
| 205.00 | 13700.0  |
| 206.00 | 57376.0  |
| 207.00 | 10062.0  |
| 208.00 | 2436.0   |
| 209.00 | 957.0    |
| 209.90 | 668.0    |
| 210.30 | 772.0    |
| 211.00 | 2345.0   |
| 211.90 | 390.0    |
| 212.90 | 213.0    |
| 214.90 | 437.0    |
| 215.90 | 1258.0   |
| 216.90 | 13718.0  |
| 218.00 | 1894.0   |
| 221.00 | 8354.0   |
| 221.70 | 2159.0   |
| 223.00 | 3207.0   |
| 224.00 | 31928.0  |
| 225.00 | 7254.0   |
| 226.00 | 1089.0   |
| 227.00 | 12117.0  |
| 228.00 | 1827.0   |
| 228.90 | 2835.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP002

|        |          |
|--------|----------|
| 230.10 | 359.0    |
| 230.90 | 1150.0   |
| 232.00 | 262.0    |
| 232.90 | 226.0    |
| 233.90 | 948.0    |
| 234.90 | 886.0    |
| 235.90 | 672.0    |
| 237.00 | 1545.0   |
| 238.00 | 230.0    |
| 238.90 | 548.0    |
| 240.00 | 472.0    |
| 241.00 | 596.0    |
| 242.00 | 2003.0   |
| 243.00 | 1648.0   |
| 244.00 | 26816.0  |
| 245.00 | 3484.0   |
| 245.90 | 4722.0   |
| 246.90 | 671.0    |
| 247.80 | 193.0    |
| 248.10 | 152.0    |
| 248.90 | 799.0    |
| 250.10 | 151.0    |
| 251.00 | 423.0    |
| 251.90 | 345.0    |
| 252.90 | 461.0    |
| 254.00 | 758.0    |
| 255.00 | 128496.0 |
| 256.00 | 18512.0  |
| 257.00 | 1682.0   |
| 258.00 | 6880.0   |
| 259.00 | 842.0    |
| 264.00 | 218.0    |
| 265.00 | 2937.0   |
| 265.80 | 497.0    |
| 266.90 | 219.0    |
| 270.90 | 352.0    |
| 271.90 | 413.0    |
| 272.90 | 3896.0   |
| 274.00 | 10900.0  |
| 275.00 | 55520.0  |
| 276.00 | 7234.0   |
| 276.90 | 4704.0   |
| 278.00 | 754.0    |
| 279.10 | 209.0    |
| 280.90 | 485.0    |
| 281.90 | 215.0    |
| 283.00 | 746.0    |
| 284.00 | 317.0    |
| 285.00 | 709.0    |
| 286.00 | 165.0    |
| 289.00 | 383.0    |
| 291.90 | 313.0    |
| 292.90 | 1185.0   |
| 294.10 | 440.0    |
| 294.90 | 561.0    |
| 296.00 | 16364.0  |
| 297.00 | 1865.0   |
| 301.10 | 190.0    |
| 301.80 | 297.0    |
| 303.00 | 2303.0   |
| 304.00 | 516.0    |
| 308.00 | 197.0    |
| 309.00 | 201.0    |
| 313.10 | 158.0    |
| 313.90 | 877.0    |
| 314.90 | 1820.0   |
| 316.00 | 747.0    |
| 316.90 | 217.0    |
| 321.00 | 690.0    |
| 321.90 | 198.0    |
| 323.00 | 5426.0   |
| 324.00 | 875.0    |
| 325.10 | 202.0    |
| 325.90 | 216.0    |
| 326.90 | 993.0    |
| 328.00 | 725.0    |
| 332.00 | 262.0    |
| 332.80 | 489.0    |
| 334.00 | 3363.0   |
| 334.90 | 652.0    |
| 340.90 | 861.0    |
| 341.90 | 255.0    |
| 345.90 | 1011.0   |
| 347.00 | 239.0    |
| 352.00 | 1578.0   |
| 353.00 | 983.0    |
| 354.00 | 1875.0   |
| 354.90 | 459.0    |
| 359.00 | 150.0    |
| 364.90 | 7752.0   |
| 366.00 | 736.0    |
| 369.90 | 173.0    |
| 370.90 | 363.0    |
| 372.00 | 2595.0   |
| 373.10 | 797.0    |
| 383.00 | 807.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP002

|        |          |
|--------|----------|
| 384.00 | 425.0    |
| 384.90 | 160.0    |
| 390.00 | 489.0    |
| 391.00 | 352.0    |
| 391.90 | 191.0    |
| 400.80 | 196.0    |
| 402.00 | 1250.0   |
| 403.00 | 1347.0   |
| 403.90 | 821.0    |
| 421.00 | 1646.0   |
| 421.90 | 1293.0   |
| 423.00 | 9529.0   |
| 424.00 | 3128.0   |
| 425.10 | 395.0    |
| 441.00 | 32736.0  |
| 442.00 | 208000.0 |
| 443.00 | 38128.0  |
| 444.00 | 2890.0   |

**Instrument :**  
BNA\_M  
**ClientSampleId :**  
DFTPP002