

Data Path : Z:\HPCHEM1\BNA_M\Data\BM021615\
 Data File : BM000535.D
 Acq On : 16 Feb 2015 17:01
 Operator : TP/IZ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.414

Quant Time: Feb 17 02:26:18 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM021615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 16 14:03:19 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
3	Naphthalene	1.069	1.090	-2.0	88	0.00
4 SURR	2-Methylnaphthalene-d10	0.529	0.529	0.0	87	0.00
5	2-Methylnaphthalene	0.709	0.727	-2.5	87	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
7	Acenaphthylene	1.571	1.546	1.6	87	0.00
8 C	Acenaphthene	1.229	1.226	0.2	89	0.00
9	Fluorene	1.457	1.431	1.8	85	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	83	0.00
11	Pentachlorophenol	0.041	0.030	26.8#	76	0.00
12	Phenanthrene	1.245	1.250	-0.4	83	0.00
13	Anthracene	1.105	1.099	0.5	83	0.00
14 SURR	Fluoranthene-d10	0.896	0.872	2.7	83	0.00
15 C	Fluoranthene	1.415	1.413	0.1	83	0.01
16 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
17	Pyrene	1.662	1.683	-1.3	81	0.00
18	Benzo(a)anthracene	1.265	1.262	0.2	82	0.00
19	Chrysene	1.472	1.457	1.0	80	0.00
20 I	Perylene-d12	1.000	1.000	0.0	83	0.01
21	Benzo(b)fluoranthene	1.526	1.519	0.5	85	0.01
22	Benzo(k)fluoranthene	1.590	1.574	1.0	80	0.00
23 C	Benzo(a)pyrene	1.424	1.397	1.9	82	0.01
24	Indeno(1,2,3-cd)pyrene	1.489	1.523	-2.3	85	0.00
25	Dibenzo(a,h)anthracene	1.150	1.192	-3.7	86	0.01
26	Benzo(g,h,i)perylene	1.337	1.345	-0.6	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0