


Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA\_M\Data\BM021618\  
 Data File : BM014258.D  
 Acq On : 16 Feb 2018 08:29  
 Operator : SJ/JU  
 Sample : DFTPP93  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP93

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA\_M\METHODS\SOM-EPA-BM021418.M  
 Title : SVOA CALIBRATION  
 Last Update : Sat Feb 17 06:25:12 2018



AutoFind: Scans 2449, 2450, 2451; Background Corrected with Scan 2444

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 61.8      | 117386  | PASS             |
| 68          | 69           | 0.00         | 2            | 1.7       | 1793    | PASS             |
| 69          | 198          | 0.00         | 100          | 56.2      | 106682  | PASS             |
| 70          | 69           | 0.00         | 2            | 0.4       | 468     | PASS             |
| 127         | 198          | 10           | 80           | 53.1      | 100701  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.8       | 1577    | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 189802  | PASS             |
| 199         | 198          | 5            | 9            | 6.7       | 12710   | PASS             |
| 275         | 198          | 10           | 60           | 29.1      | 55314   | PASS             |
| 365         | 198          | 1            | 100          | 5.5       | 10417   | PASS             |
| 441         | 443          | 0.01         | 100          | 87.4      | 21725   | PASS             |
| 442         | 198          | 50           | 100          | 70.5      | 133850  | PASS             |
| 443         | 442          | 15           | 24           | 18.6      | 24856   | PASS             |

| m/z    | Abundance |
|--------|-----------|
| 37.10  | 1088.0    |
| 38.10  | 1763.0    |
| 39.10  | 7689.0    |
| 40.00  | 647.0     |
| 43.90  | 1100.0    |
| 44.90  | 425.0     |
| 49.00  | 1525.0    |
| 50.00  | 44416.0   |
| 51.10  | 137152.0  |
| 52.10  | 6293.0    |
| 55.00  | 558.0     |
| 56.00  | 2981.0    |
| 57.00  | 7976.0    |
| 60.90  | 1692.0    |
| 61.90  | 3077.0    |
| 63.10  | 4972.0    |
| 64.00  | 770.0     |
| 65.00  | 2318.0    |
| 68.00  | 2738.0    |
| 69.00  | 124144.0  |
| 69.90  | 1143.0    |
| 73.10  | 1951.0    |
| 74.00  | 19592.0   |
| 75.00  | 23232.0   |
| 76.00  | 9331.0    |
| 77.00  | 143360.0  |
| 78.10  | 9745.0    |
| 79.00  | 11537.0   |
| 80.00  | 9229.0    |
| 81.00  | 11456.0   |
| 82.00  | 2860.0    |
| 83.00  | 2484.0    |
| 83.90  | 376.0     |
| 85.00  | 2026.0    |
| 86.00  | 3557.0    |
| 87.00  | 884.0     |
| 88.00  | 503.0     |
| 89.20  | 304.0     |
| 91.00  | 2882.0    |
| 91.90  | 2685.0    |
| 93.00  | 25264.0   |
| 94.10  | 2205.0    |
| 94.90  | 309.0     |
| 96.00  | 840.0     |
| 98.00  | 22184.0   |
| 99.00  | 13590.0   |
| 99.90  | 552.0     |
| 101.00 | 5297.0    |
| 102.90 | 2263.0    |
| 104.00 | 3016.0    |
| 104.90 | 3468.0    |
| 105.90 | 1521.0    |
| 107.00 | 46064.0   |
| 108.00 | 6018.0    |
| 108.90 | 869.0     |
| 110.00 | 58632.0   |
| 111.00 | 10103.0   |
| 111.90 | 1489.0    |
| 113.20 | 521.0     |
| 116.00 | 2417.0    |
| 117.00 | 70536.0   |
| 118.00 | 4821.0    |
| 119.00 | 640.0     |
| 120.10 | 371.0     |
| 121.90 | 3319.0    |
| 122.90 | 4071.0    |
| 123.90 | 2401.0    |
| 124.90 | 1656.0    |
| 125.90 | 1247.0    |
| 127.00 | 110392.0  |
| 128.00 | 9583.0    |
| 129.00 | 53432.0   |
| 129.90 | 4556.0    |
| 130.90 | 1239.0    |
| 132.00 | 319.0     |
| 133.10 | 391.0     |
| 134.00 | 1115.0    |
| 135.00 | 4628.0    |
| 136.10 | 2625.0    |
| 137.00 | 2808.0    |
| 137.80 | 645.0     |
| 138.00 | 642.0     |
| 138.80 | 433.0     |
| 140.00 | 1554.0    |
| 141.00 | 9351.0    |
| 142.00 | 2895.0    |
| 143.00 | 1405.0    |
| 146.10 | 1666.0    |
| 147.10 | 4624.0    |
| 148.00 | 16856.0   |
| 149.00 | 2611.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP93

|        |          |
|--------|----------|
| 150.00 | 415.0    |
| 151.00 | 1229.0   |
| 152.90 | 2301.0   |
| 154.00 | 1256.0   |
| 154.90 | 3320.0   |
| 156.00 | 4589.0   |
| 157.00 | 1011.0   |
| 158.00 | 1479.0   |
| 159.10 | 935.0    |
| 160.00 | 2681.0   |
| 160.90 | 2789.0   |
| 162.00 | 1366.0   |
| 162.90 | 306.0    |
| 164.90 | 3127.0   |
| 165.90 | 3373.0   |
| 167.00 | 17176.0  |
| 168.00 | 11504.0  |
| 169.00 | 2016.0   |
| 170.00 | 566.0    |
| 171.00 | 657.0    |
| 171.90 | 1818.0   |
| 173.00 | 2622.0   |
| 174.00 | 2948.0   |
| 175.00 | 7170.0   |
| 176.00 | 1284.0   |
| 176.90 | 2998.0   |
| 178.00 | 993.0    |
| 178.90 | 11680.0  |
| 180.00 | 6731.0   |
| 181.00 | 3621.0   |
| 181.90 | 475.0    |
| 182.90 | 333.0    |
| 184.00 | 1323.0   |
| 185.00 | 5278.0   |
| 186.00 | 37840.0  |
| 187.00 | 10902.0  |
| 187.90 | 1120.0   |
| 189.00 | 3821.0   |
| 190.90 | 1099.0   |
| 191.90 | 3332.0   |
| 193.00 | 4737.0   |
| 194.10 | 1125.0   |
| 195.00 | 576.0    |
| 195.90 | 4606.0   |
| 197.10 | 1650.0   |
| 198.00 | 191424.0 |
| 198.90 | 14476.0  |
| 200.00 | 1560.0   |
| 201.40 | 1675.0   |
| 203.00 | 2504.0   |
| 204.00 | 10883.0  |
| 205.00 | 17208.0  |
| 206.10 | 67008.0  |
| 207.00 | 9121.0   |
| 207.90 | 3769.0   |
| 209.00 | 1381.0   |
| 210.00 | 1262.0   |
| 211.00 | 2985.0   |
| 211.90 | 501.0    |
| 214.90 | 1070.0   |
| 216.00 | 2379.0   |
| 217.00 | 20192.0  |
| 217.90 | 3484.0   |
| 219.90 | 426.0    |
| 220.90 | 5414.0   |
| 222.00 | 2114.0   |
| 223.00 | 3609.0   |
| 224.10 | 36968.0  |
| 225.00 | 8662.0   |
| 225.90 | 968.0    |
| 227.00 | 20464.0  |
| 228.10 | 2247.0   |
| 229.00 | 3892.0   |
| 229.90 | 669.0    |
| 230.90 | 959.0    |
| 234.00 | 1807.0   |
| 235.10 | 1719.0   |
| 235.80 | 658.0    |
| 237.10 | 1205.0   |
| 239.20 | 664.0    |
| 240.00 | 573.0    |
| 240.90 | 1125.0   |
| 242.00 | 2341.0   |
| 243.10 | 1992.0   |
| 244.00 | 27496.0  |
| 245.00 | 3470.0   |
| 246.00 | 8374.0   |
| 247.00 | 1196.0   |
| 248.90 | 465.0    |
| 249.90 | 539.0    |
| 252.10 | 438.0    |
| 252.90 | 768.0    |
| 254.00 | 1126.0   |
| 255.00 | 135040.0 |
| 256.00 | 21096.0  |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP93

|        |          |
|--------|----------|
| 256.90 | 1798.0   |
| 257.90 | 12723.0  |
| 259.00 | 1680.0   |
| 265.00 | 6039.0   |
| 266.00 | 923.0    |
| 270.90 | 462.0    |
| 272.10 | 510.0    |
| 273.10 | 4084.0   |
| 274.00 | 8883.0   |
| 275.00 | 51280.0  |
| 276.10 | 6874.0   |
| 277.00 | 7855.0   |
| 278.00 | 1250.0   |
| 282.80 | 561.0    |
| 283.20 | 595.0    |
| 284.90 | 1145.0   |
| 291.90 | 357.0    |
| 292.90 | 1344.0   |
| 293.80 | 541.0    |
| 295.10 | 559.0    |
| 296.00 | 24440.0  |
| 297.00 | 2595.0   |
| 302.10 | 707.0    |
| 303.00 | 2634.0   |
| 304.10 | 862.0    |
| 308.00 | 384.0    |
| 313.90 | 469.0    |
| 315.10 | 1710.0   |
| 315.90 | 980.0    |
| 320.80 | 416.0    |
| 321.20 | 373.0    |
| 322.20 | 640.0    |
| 323.10 | 4308.0   |
| 324.00 | 1001.0   |
| 327.10 | 699.0    |
| 327.90 | 449.0    |
| 333.10 | 589.0    |
| 334.10 | 3523.0   |
| 335.00 | 702.0    |
| 341.10 | 516.0    |
| 346.00 | 1510.0   |
| 352.00 | 1055.0   |
| 353.00 | 1652.0   |
| 354.20 | 1080.0   |
| 365.00 | 9083.0   |
| 365.90 | 1362.0   |
| 370.10 | 522.0    |
| 371.10 | 484.0    |
| 372.00 | 2021.0   |
| 383.00 | 874.0    |
| 389.90 | 484.0    |
| 402.10 | 940.0    |
| 402.90 | 596.0    |
| 404.10 | 501.0    |
| 421.10 | 539.0    |
| 422.00 | 1307.0   |
| 423.00 | 5300.0   |
| 424.10 | 1545.0   |
| 439.80 | 409.0    |
| 441.00 | 17632.0  |
| 442.10 | 106000.0 |
| 443.10 | 20536.0  |
| 444.10 | 1594.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP93

| m/z    | Abundance |
|--------|-----------|
| 37.10  | 767.0     |
| 38.10  | 2059.0    |
| 39.00  | 9830.0    |
| 40.00  | 643.0     |
| 44.00  | 684.0     |
| 49.00  | 1475.0    |
| 50.10  | 47248.0   |
| 51.00  | 136768.0  |
| 52.10  | 5442.0    |
| 55.20  | 793.0     |
| 56.00  | 4628.0    |
| 56.90  | 9173.0    |
| 61.10  | 1485.0    |
| 62.10  | 2477.0    |
| 63.00  | 5090.0    |
| 64.10  | 663.0     |
| 65.00  | 2343.0    |
| 66.80  | 434.0     |
| 67.80  | 1358.0    |
| 68.20  | 1784.0    |
| 69.00  | 124576.0  |
| 70.10  | 1089.0    |
| 72.90  | 2252.0    |
| 74.00  | 18240.0   |
| 75.00  | 22760.0   |
| 76.10  | 7746.0    |
| 77.00  | 148480.0  |
| 78.10  | 11351.0   |
| 79.00  | 12431.0   |
| 80.00  | 10320.0   |
| 81.00  | 10528.0   |
| 82.00  | 2863.0    |
| 83.10  | 1795.0    |
| 84.90  | 1390.0    |
| 86.00  | 4104.0    |
| 86.90  | 1929.0    |
| 87.90  | 810.0     |
| 91.00  | 3076.0    |
| 92.00  | 3728.0    |
| 93.00  | 26200.0   |
| 93.90  | 1941.0    |
| 95.00  | 405.0     |
| 96.10  | 919.0     |
| 97.00  | 547.0     |
| 98.00  | 22864.0   |
| 99.00  | 11942.0   |
| 99.90  | 975.0     |
| 101.00 | 5842.0    |
| 102.10 | 551.0     |
| 103.00 | 1883.0    |
| 103.90 | 4586.0    |
| 104.90 | 4828.0    |
| 106.10 | 1145.0    |
| 107.00 | 49208.0   |
| 108.00 | 6202.0    |
| 108.90 | 893.0     |
| 110.00 | 67000.0   |
| 111.00 | 11163.0   |
| 112.00 | 2011.0    |
| 112.90 | 1036.0    |
| 115.90 | 2193.0    |
| 117.00 | 76696.0   |
| 118.00 | 4162.0    |
| 119.00 | 718.0     |
| 119.90 | 839.0     |
| 122.00 | 4216.0    |
| 123.00 | 6105.0    |
| 124.00 | 3083.0    |
| 125.10 | 1227.0    |
| 126.10 | 872.0     |
| 127.00 | 119056.0  |
| 128.00 | 12483.0   |
| 129.00 | 61696.0   |
| 130.00 | 5333.0    |
| 131.10 | 1092.0    |
| 132.20 | 574.0     |
| 133.20 | 475.0     |
| 133.90 | 1649.0    |
| 135.00 | 6227.0    |
| 135.90 | 3015.0    |
| 136.90 | 3130.0    |
| 137.70 | 674.0     |
| 140.10 | 1526.0    |
| 141.00 | 10776.0   |
| 142.00 | 3640.0    |
| 143.00 | 2541.0    |
| 144.10 | 699.0     |
| 145.00 | 326.0     |
| 145.70 | 1218.0    |
| 146.10 | 1249.0    |
| 147.00 | 4647.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP93

|        |          |
|--------|----------|
| 148.00 | 16117.0  |
| 148.90 | 2040.0   |
| 151.10 | 977.0    |
| 151.80 | 531.0    |
| 153.00 | 4006.0   |
| 154.00 | 1603.0   |
| 154.90 | 4785.0   |
| 156.00 | 7251.0   |
| 157.00 | 1080.0   |
| 157.20 | 1101.0   |
| 158.00 | 1988.0   |
| 159.00 | 1178.0   |
| 160.10 | 2022.0   |
| 161.00 | 5290.0   |
| 161.90 | 1746.0   |
| 162.90 | 519.0    |
| 165.00 | 3887.0   |
| 166.00 | 3083.0   |
| 167.00 | 26064.0  |
| 168.00 | 11692.0  |
| 169.00 | 1601.0   |
| 170.10 | 535.0    |
| 170.80 | 1213.0   |
| 171.80 | 1445.0   |
| 173.00 | 2611.0   |
| 174.10 | 3382.0   |
| 175.00 | 7315.0   |
| 176.00 | 2191.0   |
| 177.00 | 2524.0   |
| 177.90 | 1883.0   |
| 179.00 | 12971.0  |
| 179.90 | 8289.0   |
| 181.10 | 3944.0   |
| 182.00 | 663.0    |
| 183.90 | 1365.0   |
| 185.00 | 5857.0   |
| 186.00 | 43760.0  |
| 187.00 | 12655.0  |
| 188.00 | 1723.0   |
| 188.90 | 3737.0   |
| 189.80 | 908.0    |
| 191.00 | 2122.0   |
| 192.00 | 4429.0   |
| 193.00 | 5619.0   |
| 193.90 | 996.0    |
| 195.00 | 490.0    |
| 196.00 | 5934.0   |
| 197.00 | 2119.0   |
| 197.90 | 225984.0 |
| 198.90 | 15699.0  |
| 200.10 | 1950.0   |
| 200.80 | 593.0    |
| 201.60 | 304.0    |
| 202.10 | 412.0    |
| 203.00 | 2822.0   |
| 204.00 | 12307.0  |
| 205.00 | 20080.0  |
| 206.00 | 76592.0  |
| 207.10 | 10771.0  |
| 208.00 | 3362.0   |
| 209.00 | 860.0    |
| 209.90 | 1924.0   |
| 211.00 | 3565.0   |
| 211.80 | 747.0    |
| 214.90 | 1410.0   |
| 215.90 | 2097.0   |
| 217.00 | 27336.0  |
| 218.10 | 2431.0   |
| 221.00 | 5617.0   |
| 221.90 | 3681.0   |
| 223.00 | 5344.0   |
| 224.00 | 47672.0  |
| 225.00 | 11809.0  |
| 225.90 | 1587.0   |
| 227.00 | 25304.0  |
| 228.00 | 3618.0   |
| 228.90 | 6054.0   |
| 229.80 | 690.0    |
| 231.10 | 1720.0   |
| 231.80 | 433.0    |
| 232.80 | 480.0    |
| 234.00 | 1973.0   |
| 234.90 | 2248.0   |
| 236.10 | 1358.0   |
| 237.00 | 1874.0   |
| 238.00 | 463.0    |
| 239.00 | 1552.0   |
| 240.10 | 699.0    |
| 241.00 | 1754.0   |
| 242.00 | 3209.0   |
| 243.00 | 3415.0   |
| 244.00 | 30680.0  |
| 245.10 | 3751.0   |
| 246.00 | 9514.0   |
| 247.00 | 1589.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP93

|        |          |
|--------|----------|
| 247.90 | 595.0    |
| 248.90 | 1128.0   |
| 249.80 | 569.0    |
| 253.00 | 1297.0   |
| 254.20 | 1441.0   |
| 255.00 | 173952.0 |
| 256.00 | 24320.0  |
| 256.90 | 2837.0   |
| 258.00 | 13969.0  |
| 259.00 | 2072.0   |
| 261.10 | 422.0    |
| 265.00 | 7149.0   |
| 265.90 | 964.0    |
| 270.80 | 651.0    |
| 272.00 | 862.0    |
| 273.00 | 5989.0   |
| 274.00 | 16584.0  |
| 275.00 | 66808.0  |
| 276.10 | 9222.0   |
| 277.00 | 9520.0   |
| 278.00 | 1993.0   |
| 282.00 | 373.0    |
| 283.00 | 997.0    |
| 284.10 | 492.0    |
| 285.00 | 1843.0   |
| 288.00 | 346.0    |
| 288.90 | 404.0    |
| 290.10 | 390.0    |
| 290.90 | 441.0    |
| 292.10 | 663.0    |
| 293.00 | 1964.0   |
| 294.00 | 679.0    |
| 295.20 | 699.0    |
| 296.00 | 33984.0  |
| 297.00 | 4146.0   |
| 297.90 | 735.0    |
| 303.00 | 2402.0   |
| 303.90 | 644.0    |
| 309.90 | 759.0    |
| 314.00 | 1293.0   |
| 315.00 | 3517.0   |
| 316.00 | 1451.0   |
| 321.00 | 692.0    |
| 322.20 | 349.0    |
| 323.10 | 6823.0   |
| 324.10 | 1673.0   |
| 326.90 | 1782.0   |
| 328.00 | 1447.0   |
| 332.10 | 937.0    |
| 332.80 | 313.0    |
| 333.30 | 886.0    |
| 334.00 | 5041.0   |
| 335.10 | 1670.0   |
| 341.00 | 1242.0   |
| 346.10 | 1865.0   |
| 352.10 | 2234.0   |
| 353.10 | 1089.0   |
| 354.10 | 2497.0   |
| 354.90 | 355.0    |
| 365.00 | 12177.0  |
| 366.10 | 1380.0   |
| 370.00 | 365.0    |
| 371.00 | 365.0    |
| 372.10 | 3145.0   |
| 373.00 | 1042.0   |
| 383.00 | 896.0    |
| 389.90 | 436.0    |
| 391.10 | 308.0    |
| 402.10 | 1199.0   |
| 403.10 | 1176.0   |
| 404.00 | 344.0    |
| 421.00 | 1642.0   |
| 422.00 | 1204.0   |
| 423.00 | 8601.0   |
| 424.10 | 2324.0   |
| 440.10 | 724.0    |
| 441.00 | 24896.0  |
| 442.10 | 158272.0 |
| 443.10 | 29808.0  |
| 444.00 | 2801.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP93

| m/z    | Abundance |
|--------|-----------|
| 38.20  | 1295.0    |
| 39.00  | 5267.0    |
| 39.90  | 530.0     |
| 43.90  | 412.0     |
| 49.00  | 700.0     |
| 50.10  | 23240.0   |
| 51.00  | 78240.0   |
| 52.00  | 3708.0    |
| 52.90  | 429.0     |
| 53.30  | 386.0     |
| 55.10  | 781.0     |
| 55.90  | 2403.0    |
| 57.00  | 4575.0    |
| 61.10  | 987.0     |
| 62.10  | 899.0     |
| 63.10  | 3751.0    |
| 63.90  | 342.0     |
| 65.00  | 1849.0    |
| 67.90  | 1284.0    |
| 69.00  | 71328.0   |
| 70.10  | 315.0     |
| 73.10  | 1404.0    |
| 74.00  | 10315.0   |
| 75.00  | 13155.0   |
| 76.00  | 4456.0    |
| 77.00  | 82008.0   |
| 78.10  | 4855.0    |
| 79.00  | 8187.0    |
| 80.00  | 4888.0    |
| 80.90  | 7383.0    |
| 81.90  | 1495.0    |
| 82.90  | 842.0     |
| 84.90  | 1168.0    |
| 85.90  | 2415.0    |
| 86.90  | 1129.0    |
| 90.90  | 1282.0    |
| 92.00  | 976.0     |
| 93.00  | 16049.0   |
| 93.80  | 878.0     |
| 96.00  | 345.0     |
| 97.00  | 500.0     |
| 98.00  | 12995.0   |
| 99.00  | 8999.0    |
| 100.00 | 779.0     |
| 101.00 | 2564.0    |
| 102.80 | 747.0     |
| 104.00 | 3192.0    |
| 104.90 | 2540.0    |
| 106.00 | 704.0     |
| 107.00 | 26848.0   |
| 108.10 | 3636.0    |
| 109.10 | 1001.0    |
| 110.00 | 40192.0   |
| 111.00 | 6648.0    |
| 112.00 | 1110.0    |
| 112.90 | 617.0     |
| 116.00 | 1698.0    |
| 117.00 | 44608.0   |
| 118.00 | 4133.0    |
| 119.00 | 469.0     |
| 120.20 | 681.0     |
| 121.90 | 2569.0    |
| 123.10 | 3322.0    |
| 123.90 | 2006.0    |
| 124.90 | 1106.0    |
| 126.10 | 687.0     |
| 127.00 | 72656.0   |
| 128.10 | 4751.0    |
| 129.00 | 37576.0   |
| 129.90 | 3243.0    |
| 130.90 | 624.0     |
| 132.10 | 381.0     |
| 133.90 | 1145.0    |
| 135.00 | 3578.0    |
| 136.00 | 1231.0    |
| 137.00 | 1488.0    |
| 137.70 | 458.0     |
| 140.10 | 459.0     |
| 141.00 | 5512.0    |
| 141.90 | 1586.0    |
| 142.90 | 1383.0    |
| 144.10 | 592.0     |
| 145.20 | 328.0     |
| 146.00 | 1337.0    |
| 146.90 | 4033.0    |
| 148.00 | 9593.0    |
| 149.00 | 1878.0    |
| 150.90 | 676.0     |
| 151.80 | 396.0     |
| 152.80 | 1477.0    |
| 154.00 | 1612.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP93



|        |          |
|--------|----------|
| 154.90 | 3024.0   |
| 156.00 | 3558.0   |
| 157.70 | 919.0    |
| 158.10 | 605.0    |
| 159.00 | 839.0    |
| 160.00 | 2505.0   |
| 160.90 | 2179.0   |
| 161.90 | 614.0    |
| 165.10 | 2989.0   |
| 165.80 | 1667.0   |
| 167.00 | 14123.0  |
| 168.00 | 8093.0   |
| 169.00 | 1058.0   |
| 170.00 | 315.0    |
| 171.00 | 509.0    |
| 172.00 | 983.0    |
| 173.00 | 1827.0   |
| 174.10 | 2610.0   |
| 175.00 | 4938.0   |
| 176.00 | 1043.0   |
| 177.10 | 2173.0   |
| 177.90 | 773.0    |
| 179.00 | 7915.0   |
| 180.00 | 5047.0   |
| 180.90 | 1890.0   |
| 181.70 | 404.0    |
| 182.10 | 389.0    |
| 183.90 | 949.0    |
| 185.00 | 5103.0   |
| 186.00 | 30680.0  |
| 187.00 | 10061.0  |
| 188.00 | 1412.0   |
| 189.00 | 3592.0   |
| 191.10 | 1216.0   |
| 192.10 | 3056.0   |
| 193.00 | 3896.0   |
| 193.80 | 645.0    |
| 194.90 | 333.0    |
| 196.00 | 4594.0   |
| 197.10 | 964.0    |
| 198.00 | 152000.0 |
| 199.00 | 7955.0   |
| 200.00 | 1163.0   |
| 201.70 | 471.0    |
| 203.00 | 1233.0   |
| 204.00 | 7271.0   |
| 205.00 | 13494.0  |
| 206.00 | 53528.0  |
| 207.00 | 7179.0   |
| 208.10 | 3354.0   |
| 208.90 | 1175.0   |
| 209.90 | 1346.0   |
| 211.10 | 3216.0   |
| 211.70 | 382.0    |
| 215.00 | 1178.0   |
| 216.00 | 1831.0   |
| 216.90 | 18696.0  |
| 218.10 | 2543.0   |
| 219.00 | 389.0    |
| 219.90 | 397.0    |
| 221.00 | 4586.0   |
| 221.90 | 1923.0   |
| 223.00 | 3657.0   |
| 224.00 | 34208.0  |
| 225.10 | 9276.0   |
| 226.00 | 471.0    |
| 227.00 | 17128.0  |
| 228.00 | 2676.0   |
| 228.90 | 3118.0   |
| 229.80 | 409.0    |
| 231.10 | 1825.0   |
| 233.00 | 459.0    |
| 234.00 | 780.0    |
| 234.90 | 1556.0   |
| 236.00 | 968.0    |
| 237.00 | 1055.0   |
| 237.90 | 337.0    |
| 239.00 | 1099.0   |
| 239.90 | 1341.0   |
| 241.10 | 1057.0   |
| 241.90 | 2197.0   |
| 243.00 | 2490.0   |
| 244.10 | 23952.0  |
| 245.10 | 3075.0   |
| 246.00 | 7581.0   |
| 246.80 | 1448.0   |
| 248.00 | 346.0    |
| 248.90 | 873.0    |
| 251.90 | 607.0    |
| 253.10 | 809.0    |
| 254.10 | 1170.0   |
| 255.00 | 121104.0 |
| 256.00 | 17768.0  |
| 257.20 | 1450.0   |
| 258.00 | 9824.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP93

|        |          |
|--------|----------|
| 259.00 | 1345.0   |
| 264.90 | 4273.0   |
| 272.10 | 677.0    |
| 272.90 | 2811.0   |
| 274.00 | 11730.0  |
| 275.00 | 47856.0  |
| 276.00 | 7615.0   |
| 277.00 | 6679.0   |
| 278.00 | 2529.0   |
| 281.00 | 369.0    |
| 282.80 | 349.0    |
| 283.30 | 499.0    |
| 285.00 | 841.0    |
| 292.00 | 340.0    |
| 292.90 | 1254.0   |
| 294.00 | 440.0    |
| 296.00 | 22272.0  |
| 297.10 | 3317.0   |
| 298.00 | 354.0    |
| 301.10 | 310.0    |
| 302.10 | 432.0    |
| 302.90 | 1681.0   |
| 304.10 | 768.0    |
| 310.00 | 852.0    |
| 312.90 | 388.0    |
| 313.90 | 885.0    |
| 315.00 | 1642.0   |
| 316.10 | 1142.0   |
| 321.00 | 587.0    |
| 323.00 | 4983.0   |
| 324.00 | 822.0    |
| 326.80 | 1238.0   |
| 328.00 | 756.0    |
| 332.00 | 423.0    |
| 332.80 | 480.0    |
| 334.00 | 4193.0   |
| 335.10 | 1007.0   |
| 341.10 | 585.0    |
| 346.00 | 1746.0   |
| 347.10 | 415.0    |
| 352.20 | 1593.0   |
| 353.00 | 1114.0   |
| 354.00 | 1417.0   |
| 365.00 | 9993.0   |
| 365.80 | 1361.0   |
| 370.80 | 317.0    |
| 372.10 | 1911.0   |
| 373.00 | 375.0    |
| 383.00 | 646.0    |
| 389.90 | 307.0    |
| 402.00 | 1374.0   |
| 403.00 | 1174.0   |
| 404.00 | 422.0    |
| 421.10 | 1075.0   |
| 422.20 | 1319.0   |
| 423.10 | 6800.0   |
| 423.90 | 1838.0   |
| 438.70 | 302.0    |
| 440.00 | 595.0    |
| 441.10 | 22648.0  |
| 442.00 | 137280.0 |
| 443.10 | 24224.0  |
| 444.00 | 2345.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP93

Scan 2444 (17.362 min): BM014258.D  
DFTPP93

| m/z    | Abundance |
|--------|-----------|
| 43.90  | 534.0     |
| 206.90 | 550.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP93