

Data Path : Z:\HPCHEM1\BNA_M\Data\BM021715\
 Data File : BM000537.D
 Acq On : 17 Feb 2015 12:13
 Operator : TP/IZ
 Sample : SSTD0.415
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 18 03:50:53 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-SIM-BM021215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 12 16:45:55 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	-0.02
2 I	Naphthalene-d8	1.000	1.000	0.0	100	-0.02
3	Naphthalene	0.938	0.932	0.6	98	-0.02
4 SURR	2-Methylnaphthalene-d10	0.510	0.492	3.5	96	-0.02
5	2-Methylnaphthalene	0.634	0.613	3.3	96	-0.02
6 I	Acenaphthene-d10	1.000	1.000	0.0	94	-0.02
7	Acenaphthylene	1.505	1.458	3.1	94	-0.02
8 C	Acenaphthene	1.155	1.133	1.9	92	-0.03
9	Fluorene	1.396	1.328	4.9	91	-0.02
10 I	Phenanthrene-d10	1.000	1.000	0.0	89	-0.02
11	Pentachlorophenol	0.051	0.032	37.3#	61	-0.03
12	Phenanthrene	1.086	1.064	2.0	89	-0.02
13	Anthracene	0.977	0.951	2.7	88	-0.02
14 SURR	Fluoranthene-d10	0.841	0.828	1.5	86	-0.01
15 C	Fluoranthene	1.211	1.169	3.5	84	-0.02
16 I	Chrysene-d12	1.000	1.000	0.0	82	-0.02
17	Pyrene	1.439	1.424	1.0	85	-0.02
18	Benzo(a)anthracene	1.128	1.099	2.6	81	-0.02
19	Chrysene	1.300	1.282	1.4	79	-0.02
20 I	Perylene-d12	1.000	1.000	0.0	79	-0.02
21	Benzo(b)fluoranthene	1.352	1.382	-2.2	80	-0.02
22	Benzo(k)fluoranthene	1.345	1.291	4.0	78	-0.02
23 C	Benzo(a)pyrene	1.264	1.246	1.4	79	-0.03
24	Indeno(1,2,3-cd)pyrene	1.410	1.397	0.9	78	-0.03
25	Dibenzo(a,h)anthracene	1.114	1.119	-0.4	80	-0.04
26	Benzo(g,h,i)perylene	1.244	1.232	1.0	78	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0