

Data Path : Z:\HPCHEM1\BNA_M\Data\BM021915\
 Data File : BM000549.D
 Acq On : 19 Feb 2015 09:26
 Operator : TP/IZ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.420

Quant Time: Feb 19 13:35:22 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM021815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 19 01:10:38 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
3	Naphthalene	1.002	0.914	8.8	103	0.00
4 SURR	2-Methylnaphthalene-d10	0.534	0.480	10.1	101	0.00
5	2-Methylnaphthalene	0.714	0.628	12.0	100	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	96	-0.02
7	Acenaphthylene	1.659	1.470	11.4	97	0.00
8 C	Acenaphthene	1.273	1.142	10.3	96	0.00
9	Fluorene	1.576	1.410	10.5	97	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
11	Pentachlorophenol	0.113	0.085	24.8#	84	0.00
12	Phenanthrene	1.130	1.039	8.1	94	0.00
13	Anthracene	1.063	0.941	11.5	93	-0.02
14 SURR	Fluoranthene-d10	0.852	0.752	11.7	90	-0.01
15 C	Fluoranthene	1.293	1.147	11.3	91	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	85	0.00
17	Pyrene	1.816	1.712	5.7	90	0.00
18	Benzo(a)anthracene	1.255	1.109	11.6	85	0.00
19	Chrysene	1.363	1.310	3.9	87	0.00
20 I	Perylene-d12	1.000	1.000	0.0	88	0.00
21	Benzo(b)fluoranthene	1.496	1.335	10.8	86	0.00
22	Benzo(k)fluoranthene	1.476	1.313	11.0	83	0.00
23 C	Benzo(a)pyrene	1.337	1.221	8.7	88	0.00
24	Indeno(1,2,3-cd)pyrene	1.254	1.227	2.2	95	-0.01
25	Dibenzo(a,h)anthracene	0.994	0.964	3.0	96	0.00
26	Benzo(g,h,i)perylene	1.085	1.079	0.6	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0