

Data Path : Z:\HPCHEM1\BNA_M\Data\BM021915\
 Data File : BM000552.D
 Acq On : 19 Feb 2015 15:00
 Operator : TP/IZ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.421

Quant Time: Feb 19 17:01:59 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM021815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 19 10:41:02 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	115	0.00
3	Naphthalene	1.002	0.921	8.1	116	0.00
4 SURR	2-Methylnaphthalene-d10	0.534	0.490	8.2	115	-0.01
5	2-Methylnaphthalene	0.714	0.649	9.1	116	-0.01
6 I	Acenaphthene-d10	1.000	1.000	0.0	104	-0.02
7	Acenaphthylene	1.659	1.574	5.1	113	-0.01
8 C	Acenaphthene	1.273	1.191	6.4	109	0.00
9	Fluorene	1.576	1.463	7.2	109	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
11	Pentachlorophenol	0.113	0.090	20.4#	102	0.00
12	Phenanthrene	1.130	1.057	6.5	111	-0.02
13	Anthracene	1.063	0.937	11.9	107	-0.02
14 SURR	Fluoranthene-d10	0.852	0.749	12.1	103	-0.01
15 C	Fluoranthene	1.293	1.134	12.3	103	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	96	-0.01
17	Pyrene	1.816	1.733	4.6	103	0.00
18	Benzo(a)anthracene	1.255	1.157	7.8	100	0.00
19	Chrysene	1.363	1.260	7.6	95	0.00
20 I	Perylene-d12	1.000	1.000	0.0	94	0.00
21	Benzo(b)fluoranthene	1.496	1.440	3.7	99	0.00
22	Benzo(k)fluoranthene	1.476	1.318	10.7	89	0.00
23 C	Benzo(a)pyrene	1.337	1.230	8.0	95	0.00
24	Indeno(1,2,3-cd)pyrene	1.254	1.216	3.0	101	-0.01
25	Dibenzo(a,h)anthracene	0.994	0.954	4.0	101	0.00
26	Benzo(g,h,i)perylene	1.085	1.051	3.1	101	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0