

Data Path : Z:\HPCHEM1\BNA M\DATA\BM021916\
 Data File : BM004298.D
 Acq On : 19 Feb 2016 23:04
 Operator : SJ/UM
 Sample : H1497-18
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 0311-S06-0312

Quant Time: Feb 19 23:56:54 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM021216.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Feb 19 23:51:45 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	25466	20.00	ng/ul	0.00
18) Naphthalene-d8	10.42	136	131843	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.29	164	90277	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.04	188	214234	20.00	ng/ul	0.00
78) Chrysene-d12	21.26	240	248515	20.00	ng/ul	0.00
86) Perylene-d12	23.48	264	226161	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.09	96	2079	4.11	ng/uL	0.00
5) Phenol-d5	6.82	99	57811	28.26	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.97	67	32110	29.07	ng/ul	0.00
9) 2-Chlorophenol-d4	7.17	132	50500	27.84	ng/ul	0.00
13) 4-Methylphenol-d8	8.36	113	55243	31.71	ng/ul	0.00
19) Nitrobenzene-d5	8.79	128	27786	27.41	ng/ul	0.00
22) 2-Nitrophenol-d4	9.51	143	32745	28.90	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.06	165	56553	27.67	ng/ul	0.00
29) 4-Chloroaniline-d4	10.57	131	84735	34.83	ng/ul	0.00
44) Dimethylphthalate-d6	13.70	166	233242	31.05	ng/ul	0.00
47) Acenaphthylene-d8	13.98	160	270743	29.66	ng/ul	0.00
52) 4-Nitrophenol-d4	14.54	143	34314	28.23	ng/ul	0.00
58) Fluorene-d10	15.29	176	201720	31.86	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.43	200	31886	24.44	ng/ul	0.00
71) Anthracene-d10	17.14	188	320734	30.62	ng/ul	0.00
79) Pyrene-d10	19.45	212	389726	29.54	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.34	264	386166	32.00	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
45) Dimethylphthalate	13.75	163	73850	9.41	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM021916\
Data File : BM004298.D
Acq On : 19 Feb 2016 23:04
Operator : SJ/UM
Sample : H1497-18
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
0311-S06-0312

Quant Time: Feb 19 23:56:54 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM021216.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Feb 19 23:51:45 2016
Response via : Initial Calibration

