


Data Path : Z:\HPCHEM1\BNA M\Data\BM021918\
 Data File : BM014296.D
 Acq On : 19 Feb 2018 15:54
 Operator : SJ/JU
 Sample : DFTPP94
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP94

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM021418.M
 Title : SVOA CALIBRATION
 Last Update : Thu Feb 22 03:35:18 2018



AutoFind: Scans 2448, 2449, 2450; Background Corrected with Scan 2443

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	57.4	136658	PASS
68	69	0.00	2	2.0	2525	PASS
69	198	0.00	100	54.0	128501	PASS
70	69	0.00	2	0.6	717	PASS
127	198	10	80	53.8	128072	PASS
197	198	0.00	2	0.8	1940	PASS
198	198	100	100	100.0	238037	PASS
199	198	5	9	6.8	16192	PASS
275	198	10	60	27.3	64869	PASS
365	198	1	100	4.8	11502	PASS
441	443	0.01	100	80.5	24823	PASS
442	198	50	100	66.4	158125	PASS
443	442	15	24	19.5	30845	PASS

m/z	Abundance
37.00	865.0
38.10	2253.0
39.10	10178.0
40.00	696.0
44.00	365.0
44.80	555.0
49.00	827.0
50.00	46032.0
51.10	141504.0
52.10	6786.0
53.10	390.0
54.90	429.0
56.00	4612.0
57.00	6885.0
57.90	333.0
60.90	1627.0
62.00	2054.0
63.00	5127.0
64.10	703.0
65.00	1225.0
68.00	2414.0
69.00	129744.0
69.90	397.0
73.00	1606.0
74.00	19064.0
75.00	23248.0
76.10	8480.0
77.10	147584.0
78.10	10062.0
79.00	13641.0
80.00	8667.0
81.00	10393.0
82.10	2007.0
82.90	2324.0
85.10	1435.0
86.00	4287.0
87.00	1632.0
88.10	364.0
90.80	2875.0
92.00	2918.0
93.00	25408.0
94.00	1958.0
96.00	710.0
98.00	24400.0
99.00	12212.0
100.10	695.0
101.00	5827.0
103.00	1961.0
104.00	4548.0
105.00	3924.0
106.00	1886.0
107.00	45128.0
108.00	6073.0
109.10	1297.0
110.00	64448.0
111.00	11061.0
112.00	1633.0
112.90	796.0
115.90	2840.0
117.00	69632.0
118.00	5122.0
119.10	740.0
122.00	3550.0
123.00	6315.0
124.00	2416.0
125.00	1978.0
125.90	480.0
127.00	117944.0
128.10	9045.0
129.00	54680.0
130.00	5297.0
130.90	972.0
131.90	418.0
134.10	2159.0
135.00	4181.0
136.00	2074.0
137.00	3530.0
140.00	416.0
141.00	7959.0
141.90	2781.0
143.00	2169.0
143.80	533.0
144.80	348.0
146.00	1506.0
147.10	3821.0
148.00	10871.0
148.90	1728.0
149.90	481.0
151.00	626.0
151.60	534.0
153.00	2008.0

Instrument :
BNA_M
ClientSampleId :
DFTPP94

153.90	1765.0
154.90	4718.0
156.10	5870.0
157.10	1090.0
158.10	1461.0
159.10	844.0
160.00	3944.0
161.10	3078.0
162.10	1331.0
163.00	748.0
163.70	317.0
165.00	3352.0
165.90	2286.0
167.00	20160.0
168.00	10454.0
169.00	2533.0
170.10	706.0
171.00	863.0
172.00	1693.0
172.90	1943.0
174.10	4543.0
175.00	5513.0
175.90	1303.0
177.00	3287.0
178.10	966.0
178.90	9605.0
180.00	7293.0
181.00	2801.0
182.10	371.0
183.80	634.0
185.00	5720.0
186.00	36416.0
187.00	10787.0
187.90	1048.0
188.20	1077.0
189.10	3168.0
190.00	752.0
191.00	1687.0
191.90	4262.0
193.00	4442.0
193.80	1273.0
195.20	671.0
196.00	6704.0
196.80	965.0
197.10	1225.0
198.00	196032.0
198.90	14818.0
199.90	1330.0
201.40	835.0
202.90	3141.0
204.10	9468.0
205.00	17568.0
206.00	62464.0
207.10	10795.0
208.00	2601.0
209.10	840.0
210.00	910.0
211.00	2371.0
211.70	370.0
212.90	393.0
215.10	1062.0
215.90	1980.0
217.00	18728.0
217.90	2473.0
218.90	354.0
220.90	5520.0
222.00	2387.0
223.10	4002.0
224.00	36024.0
225.10	9822.0
226.10	1099.0
227.00	20496.0
228.00	2956.0
229.00	3968.0
230.00	424.0
231.00	1216.0
232.90	377.0
233.80	1259.0
234.90	1405.0
236.00	1043.0
237.00	1538.0
239.00	691.0
240.10	523.0
241.00	930.0
242.10	2012.0
242.90	1710.0
244.00	25888.0
245.00	3597.0
246.00	8196.0
247.00	1695.0
248.80	950.0
251.00	467.0
252.90	1274.0
253.80	973.0
255.00	129104.0

Instrument :
BNA_M
ClientSampleId :
DFTPP94

256.00	18264.0
257.00	1175.0
258.00	10350.0
258.90	2132.0
260.20	327.0
265.00	5053.0
266.00	680.0
270.00	486.0
270.90	686.0
271.90	527.0
273.10	3482.0
274.10	9157.0
275.00	47832.0
276.00	7347.0
277.00	6860.0
278.00	962.0
281.00	504.0
282.10	368.0
283.00	978.0
284.10	739.0
285.00	1144.0
285.90	583.0
289.10	371.0
290.20	488.0
292.10	377.0
293.10	1327.0
293.90	400.0
295.00	629.0
296.00	25576.0
297.00	3828.0
302.00	451.0
303.10	2663.0
304.10	625.0
307.80	419.0
308.80	403.0
314.00	580.0
314.90	2920.0
316.00	1213.0
320.90	787.0
323.10	4691.0
324.00	1087.0
326.90	1437.0
331.90	737.0
333.20	813.0
334.00	3535.0
334.90	677.0
341.00	456.0
345.90	1195.0
352.00	871.0
353.00	1633.0
354.10	1296.0
365.00	7868.0
366.00	807.0
372.10	1951.0
372.90	586.0
383.00	895.0
390.10	332.0
391.10	443.0
401.90	749.0
403.20	1585.0
403.90	373.0
421.20	1172.0
422.10	811.0
423.10	4345.0
424.00	1749.0
441.10	14262.0
442.10	91272.0
443.10	20888.0
444.10	1607.0

Instrument :
BNA_M
ClientSampleId :
DFTPP94

m/z	Abundance
37.00	946.0
38.10	3066.0
39.10	12472.0
40.00	544.0
41.00	313.0
44.10	554.0
49.10	1200.0
50.10	48016.0
51.10	163200.0
52.00	9062.0
53.00	332.0
55.10	689.0
56.00	5142.0
57.00	8951.0
61.00	1331.0
62.10	2789.0
63.10	6486.0
64.00	870.0
65.00	3184.0
68.00	3375.0
69.00	161024.0
70.00	1220.0
72.90	2053.0
74.00	19944.0
75.00	29184.0
76.10	10026.0
77.00	180928.0
78.10	11625.0
79.00	15600.0
80.00	10492.0
81.00	15089.0
82.10	2823.0
83.00	2475.0
83.80	308.0
84.90	2724.0
86.00	4138.0
86.90	2189.0
88.10	354.0
89.20	303.0
91.10	3220.0
92.00	3632.0
93.00	31752.0
94.00	2145.0
95.80	1046.0
97.10	683.0
98.00	27368.0
99.00	15226.0
99.90	1050.0
101.00	6971.0
101.90	320.0
103.00	2673.0
104.00	5726.0
105.00	5182.0
106.00	1355.0
107.00	54896.0
107.90	9287.0
109.10	1337.0
110.00	80640.0
111.00	14377.0
112.00	1699.0
113.00	396.0
114.90	363.0
116.00	3595.0
117.00	86224.0
118.00	5501.0
119.20	672.0
120.00	944.0
120.90	822.0
122.00	4757.0
123.00	5349.0
124.00	2258.0
125.00	2072.0
126.20	806.0
127.00	155520.0
128.00	10968.0
129.00	73768.0
129.90	5415.0
131.00	1590.0
132.00	637.0
132.90	615.0
134.00	2887.0
135.00	7179.0
136.00	1484.0
136.90	2979.0
138.00	427.0
138.90	853.0
139.80	666.0
141.00	12952.0
141.90	2142.0
143.20	1445.0
143.90	591.0

Instrument :
BNA_M
ClientSampleId :
DFTPP94

144.90	621.0
145.90	2317.0
146.90	5450.0
148.00	18176.0
149.00	3027.0
150.00	1148.0
150.80	994.0
151.20	1176.0
152.10	507.0
153.00	3215.0
154.00	1974.0
155.00	6448.0
156.00	7771.0
156.90	1472.0
158.10	1679.0
158.90	1365.0
160.10	4390.0
161.00	5093.0
162.10	1156.0
163.00	512.0
163.90	498.0
165.00	4819.0
166.00	2974.0
167.00	24600.0
168.00	13725.0
168.90	1685.0
169.90	1225.0
170.90	1446.0
171.90	2326.0
173.00	2513.0
174.00	4680.0
175.00	8232.0
176.00	1853.0
177.00	2701.0
177.90	2227.0
178.90	17296.0
180.00	10214.0
180.90	4733.0
182.10	660.0
182.80	342.0
184.00	1334.0
185.00	8789.0
186.00	52160.0
187.00	15328.0
188.00	1626.0
189.00	3676.0
189.80	649.0
191.00	1897.0
192.00	5356.0
193.00	6250.0
193.90	771.0
195.10	515.0
196.00	10164.0
197.00	3125.0
198.00	295744.0
199.00	19752.0
199.90	1460.0
201.30	1430.0
202.10	673.0
203.00	3448.0
204.00	14568.0
205.00	24944.0
206.00	92944.0
207.10	12358.0
208.00	4904.0
209.10	1108.0
210.10	2460.0
211.00	3788.0
211.80	580.0
213.10	764.0
215.00	1551.0
215.90	2580.0
217.00	31560.0
217.90	3713.0
219.00	498.0
219.80	358.0
221.00	8843.0
222.00	3328.0
223.00	7308.0
224.00	56168.0
225.10	13733.0
225.90	1896.0
227.00	32408.0
228.00	4557.0
229.00	6629.0
229.80	323.0
231.00	1736.0
232.90	506.0
234.10	1342.0
235.00	1450.0
235.80	1226.0
237.00	1802.0
239.00	1434.0
240.00	1204.0
240.90	1300.0

Instrument :
BNA_M
ClientSampleId :
DFTPP94

242.00	3351.0
243.00	3664.0
244.00	36552.0
245.00	6245.0
246.00	11824.0
247.00	2049.0
247.90	313.0
248.90	1582.0
251.90	571.0
252.90	1940.0
254.00	1467.0
255.00	214784.0
256.00	29864.0
257.10	1961.0
258.00	19664.0
258.90	2607.0
259.90	522.0
264.10	399.0
265.00	7328.0
265.90	795.0
269.90	361.0
270.90	302.0
271.90	579.0
273.00	5241.0
274.00	16029.0
275.00	80208.0
276.00	10434.0
277.00	8965.0
278.10	1591.0
278.90	455.0
280.90	614.0
282.20	310.0
283.10	602.0
284.00	1110.0
285.10	1174.0
289.00	701.0
292.00	582.0
293.00	2733.0
294.00	414.0
294.80	630.0
296.00	36280.0
297.00	5644.0
302.20	600.0
303.00	2619.0
304.00	765.0
308.20	701.0
309.10	452.0
309.90	550.0
314.10	1327.0
315.00	3972.0
316.10	1925.0
317.00	560.0
321.10	800.0
321.90	611.0
323.10	7236.0
324.10	1478.0
327.00	1827.0
328.10	751.0
332.00	743.0
332.80	613.0
334.00	5788.0
335.00	2215.0
338.90	303.0
341.20	762.0
346.00	1946.0
347.00	440.0
350.90	332.0
352.00	2947.0
353.10	2080.0
354.00	3662.0
355.00	388.0
365.00	13926.0
366.00	1792.0
370.80	341.0
372.10	3600.0
373.30	591.0
383.00	758.0
390.20	376.0
401.90	908.0
403.00	1949.0
404.10	688.0
420.70	1281.0
421.20	1135.0
421.80	1328.0
423.10	11593.0
424.00	2494.0
439.40	431.0
439.70	440.0
441.10	29040.0
442.10	189184.0
443.10	37504.0
444.00	3620.0

Instrument :
BNA_M
ClientSampleId :
DFTPP94

m/z	Abundance
37.10	647.0
38.20	1435.0
39.00	6919.0
40.00	765.0
43.90	672.0
44.90	381.0
49.00	1060.0
50.00	29768.0
51.00	105272.0
52.00	4849.0
55.10	348.0
56.00	2826.0
57.00	5779.0
60.90	1475.0
62.00	1688.0
63.10	4167.0
65.00	1264.0
66.00	325.0
68.00	1788.0
69.00	94736.0
70.10	534.0
73.00	1757.0
74.00	15270.0
75.00	19432.0
76.10	7870.0
77.10	116888.0
78.00	8129.0
79.00	10726.0
80.00	7563.0
81.00	9031.0
82.00	2451.0
83.00	1292.0
84.90	1532.0
86.00	2238.0
86.90	988.0
91.00	2387.0
92.00	2708.0
93.00	18352.0
94.00	1278.0
96.00	608.0
97.20	686.0
98.00	15948.0
99.00	10228.0
99.80	641.0
101.00	4525.0
103.00	2076.0
104.10	3544.0
105.00	3408.0
106.00	1171.0
107.00	43296.0
108.00	6144.0
109.10	1313.0
110.00	57208.0
111.00	8620.0
111.90	1833.0
116.00	2148.0
117.00	64448.0
118.00	4326.0
119.00	367.0
119.80	311.0
121.90	3718.0
123.00	4469.0
124.00	2051.0
125.20	1495.0
127.00	110752.0
128.00	9714.0
129.00	52784.0
129.90	5114.0
130.90	1837.0
132.80	484.0
134.00	1679.0
135.00	4349.0
135.90	3125.0
137.10	3291.0
138.00	558.0
138.90	338.0
140.20	981.0
141.00	7069.0
142.10	1980.0
142.90	1228.0
144.10	424.0
145.10	542.0
145.90	1329.0
147.00	4354.0
148.00	12413.0
148.90	2186.0
150.10	436.0
151.00	1203.0
151.90	730.0
153.00	1611.0
154.00	2143.0

Instrument :
BNA_M
ClientSampleId :
DFTPP94

155.00	4417.0
156.00	6986.0
156.80	492.0
157.90	1255.0
158.80	832.0
160.00	2791.0
161.00	3949.0
162.00	1257.0
163.10	316.0
164.00	459.0
165.00	4255.0
166.00	1961.0
167.00	18808.0
168.00	8092.0
169.00	1390.0
169.90	629.0
171.10	865.0
171.80	1024.0
172.90	1621.0
173.90	3974.0
175.10	6885.0
175.80	1625.0
177.00	3363.0
178.00	989.0
178.90	14060.0
180.00	7325.0
180.90	4115.0
182.10	389.0
182.50	490.0
183.90	711.0
185.00	4490.0
186.00	39552.0
187.00	11654.0
188.10	1230.0
189.10	3354.0
191.00	1633.0
192.00	5614.0
192.90	4982.0
194.00	637.0
194.90	324.0
196.00	5825.0
197.00	1470.0
198.00	222336.0
199.00	14008.0
200.20	1013.0
201.00	689.0
201.70	1261.0
203.00	3868.0
204.00	10896.0
205.00	17936.0
206.10	72072.0
207.10	10503.0
208.00	3169.0
209.00	1402.0
210.10	1272.0
211.00	2900.0
213.00	445.0
215.00	1152.0
216.00	2513.0
217.00	25544.0
217.90	3316.0
218.90	483.0
219.90	521.0
221.00	7190.0
222.00	2520.0
223.00	6819.0
224.00	42048.0
225.10	10125.0
226.00	2129.0
227.00	24776.0
228.00	2892.0
229.00	3463.0
230.00	784.0
231.00	1656.0
233.20	316.0
234.10	1471.0
234.90	1551.0
236.00	1646.0
236.80	1822.0
239.00	954.0
239.80	314.0
241.00	1033.0
241.90	2368.0
243.00	3041.0
244.10	28616.0
245.10	4760.0
246.00	8237.0
246.90	1509.0
248.20	462.0
249.10	857.0
251.10	340.0
252.90	924.0
254.10	1237.0
255.00	166592.0
256.00	23536.0

Instrument :
BNA_M
ClientSampleId :
DFTPP94

257.00	1269.0
258.00	12628.0
259.00	1941.0
259.70	353.0
264.10	384.0
265.00	6797.0
266.10	813.0
267.00	333.0
271.90	347.0
273.00	4451.0
274.10	12933.0
275.00	66568.0
276.00	8551.0
276.90	8474.0
277.80	669.0
281.10	426.0
282.90	1230.0
283.10	1286.0
284.00	394.0
285.10	896.0
289.00	353.0
292.90	2419.0
294.00	455.0
294.90	911.0
296.00	31568.0
297.00	4636.0
302.00	409.0
303.00	3610.0
304.10	884.0
308.20	544.0
309.80	584.0
314.10	1569.0
315.00	3180.0
316.10	1374.0
320.80	752.0
321.90	585.0
322.20	598.0
323.00	7498.0
324.10	953.0
326.90	1928.0
327.90	580.0
332.00	464.0
333.10	1216.0
334.10	5173.0
335.10	925.0
341.20	655.0
345.90	1652.0
352.10	2263.0
353.00	855.0
354.10	2282.0
355.00	429.0
365.00	12713.0
366.00	1574.0
370.80	997.0
372.10	3735.0
373.20	689.0
383.00	986.0
389.80	415.0
402.10	1390.0
403.00	1783.0
404.10	821.0
421.10	2337.0
422.10	1599.0
423.10	10490.0
424.00	1701.0
440.20	503.0
441.10	31168.0
442.10	193920.0
443.10	34144.0
444.10	3012.0

Instrument :
BNA_M
ClientSampleId :
DFTPP94

m/z	Abundance
44.00	359.0
73.00	649.0
207.00	1071.0
207.80	379.0

Instrument :
BNA_M
ClientSampleId :
DFTPP94