

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM021920\
 Data File : BM024945.D
 Acq On : 19 Feb 2020 15:52
 Operator : CG/JU
 Sample : PB126805BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK05

Quant Time: Feb 19 23:31:16 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM020620MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 19 03:36:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.39	152	307528	20.00	ng/ul	0.00
18) Naphthalene-d8	10.14	136	1136550	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.03	164	671557	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.79	188	1398507	20.00	ng/ul	0.00
78) Chrysene-d12	21.00	240	1403175	20.00	ng/ul	0.00
86) Perylene-d12	23.10	264	1545647	20.00	ng/ul	-0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.03	96	44392	6.94	ng/uL	0.00
5) Phenol-d5	6.59	99	608067	29.76	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.74	67	380302	32.96	ng/ul	-0.01
9) 2-Chlorophenol-d4	6.93	132	523623	28.61	ng/ul	-0.01
13) 4-Methylphenol-d8	8.10	113	469856	27.26	ng/ul	-0.01
19) Nitrobenzene-d5	8.52	128	237624	28.91	ng/ul	-0.01
22) 2-Nitrophenol-d4	9.24	143	269466	27.75	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.77	165	483904	24.70	ng/ul	0.00
29) 4-Chloroaniline-d4	10.28	131	604970	33.01	ng/ul	0.00
44) Dimethylphthalate-d6	13.46	166	1410994	27.83	ng/ul	0.00
47) Acenaphthylene-d8	13.72	160	1705458	28.81	ng/ul	0.00
52) 4-Nitrophenol-d4	14.26	143	168968	20.51	ng/ul	0.00
58) Fluorene-d10	15.04	176	1238161	27.58	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.17	200	187786	20.60	ng/ul	0.00
71) Anthracene-d10	16.89	188	1951016	30.56	ng/ul	-0.01
79) Pyrene-d10	19.20	212	2188404	31.18	ng/ul	0.00
90) Benzo(a)pyrene-d12	22.97	264	2404897	30.98	ng/ul	-0.01

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM021920\
Data File : BM024945.D
Acq On : 19 Feb 2020 15:52
Operator : CG/JU
Sample : PB126805BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK05

Quant Time: Feb 19 23:31:16 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM020620MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Feb 19 03:36:08 2020
Response via : Initial Calibration

