

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021924\
 Data File : BM044192.D
 Acq On : 20 Feb 2024 00:22
 Operator : MA/JU
 Sample : P1455-01
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AA8

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/20/2024
 Supervised By :mohammad ahmed 02/21/2024

Quant Time: Feb 20 00:52:55 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 15 13:40:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.896	152	5932	0.400	ng/ul	0.00
4) Naphthalene-d8	10.694	136	17732	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.538	164	8611	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.292	188	17855m	0.400	ng/ul	0.00
17) Chrysene-d12	21.494	240	12526m	0.400	ng/ul	0.00
23) Perylene-d12	23.879	264	20361	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.348	96	37486	4.462	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.288	152	9256	0.373	ng/ul	-0.01
18) Fluoranthene-d10	19.322	212	18083	0.484	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.386	88	6124	0.729	ng/ul#	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

